A-A-53195A
12 February 1990
SUPERSEDING
A-A-53195
18 September 1987

COMMERCIAL ITEM DESCRIPTION

ANALYZING KIT, MEAT FAT CONTENT

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-A-36926 which was cancelled.

This commercial item description covers the requirements for a meat fat analyzing kit used to determine the percentage of fat in ground beef in a simple, fast and accurate manner.

Salient characteristics.

The analyzing kit shall be contained in a carrying case having overall dimensions in inches of 6-1/4 width by 12-1/2 length by 17-1/2 height with a tolerance of ± 1 inch. The case shall be double wall construction, hinged along one 17-1/2 side to open into halves or compartments, and shall have a carrying handle and catches.

The analyzing kit shall contain the following components:

1. A heat unit for rendering fat, rated for 50/60 Hz, AC and 500 watts operation.

2. A three conductor power cord having a minimum overall length of 6 feet. When required for use on 115 volt supply lines (see ordering data) the power cord shall include a plug suitable for use in receptacles conforming to ANSI C73.11. When required for use on 230 volt supply lines (see ordering data) the power cord is to be supplied without a plug, but wires at plug end to be stripped for user adaptation to overseas systems.

3. A 15 minute interval timer, with bell, for controlling power to the heating unit. The end of the 15 minute interval shall be indicated by the sounding of the bell.

AMSC N/A FSC 6695

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.

4. A bracket to support and position a sample screen, funnel and test tube beneath the heating unit to collect the rendered fat.

5. A calibrated scale with sliding pointer to measure the level of fat collected in the test tube and determine the percentage of fat for the sample test:

6. Two (2) funnels.

7. Three (3) test tubes.

8. Two (2) sample screens.

9. Two (2) cleaning brushes.

Workmanship. The analyzing kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One analyzing kit, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice. The unit of product for tests purposes shall be one analyzing kit.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6695-00-926-4502 ANALYZING KIT, MEAT FAT CONTENT, 110v, 60 Hz

6695-01-143-2054 ANALYZING KIT, MEAT FAT CONTENT, 230v, 50 Hz

Ordering data. (Applicable voltage required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6695-0093

Location: CID/ANALYZING/835/12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.