A-A-53200
21 September 1987

COMMERCIAL ITEM DESCRIPTION

CENTRIFUGE, LABORATORY (Hand Held)

The General Services Administration has authorized the use of this commercial item description in preference to military specification MIL-C-37099

This Commercial Item Description covers a lightweight laboratory centrifuge with bench-type clamp and a two-place head.

Salient characteristics:

The centrifuge shall consist of the following: housing with a bench clamp, output shaft, driving mechanism, crankshaft, removable hand crank, tube head, and tube shields with rubber cushions. The centrifuge shall accommodate two centrifuge tubes having a nominal length of 118 millimeters (mm.) and an outside diameter of 16.8 $\pm$ 0.5 mm.

The bench clamp shall be an integral part of the housing. The bench clamp shall include a non-detachable thumbscrew for clamping the centrifuge to a working surface, plank or tree limb having a thickness up to or including 1-5/8 inches. The thumbscrew head shall be large enough to permit tightening and loosening to be done without using tools. The contact end of the thumbscrew shall include a button to facilitate clamping on wood surfaces. The thumbscrew shall be suitably protected against corrosion.

The output shaft shall be vertically mounted and shall have sufficient length above the housing enclosure to accommodate the tube head and permit the tubeshields to clear the housing when they are vertically oriented in the standstill position.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

The driving mechanism shall be enclosed within the housing. The driving mechanism shall be gear driven and shall include a machined work gear for the rotation of the output shaft. The gear ratio shall be 15:1 or 16:1 (i.e. 15 or 16 revolutions of the output shaft for each turn of the crankshaft). The gears shall be machined and permanently lubricated. The gears, output shaft, crankshaft, and any necessary bearings shall be fitted within the housing to provide a smooth and easy rotation of the components.

The detachable handcrank shall engage the crankshaft in such a manner as to impart only a clockwise (right hand) rotation. The handcrank shall have a throw of 2-5/16 to 3-3/16 inches, and shall consist of a metal lever arm with a handle grip that rotates freely. The handle grip shall have a smooth finish. The crankshaft end of the lever shall engage the crankshaft without slipping during the manual rotation, and then disengage the rotating crankshaft when the manual rotation is stopped.

The tube-head shall be a demountable two-place type. The tube-head shall be for attaching to the output shaft and being secured by suitable hardware (e.g. set screw or thumbscrew). The frame of the tube-head shall include two freely rotating trunnion rings that shall accommodate the tube shields. The trunnion rings shall be positioned to represent a nominal 2-5/16 inch radial distance from the vertical shaft (center-to-center). The frame of the tube-head shall be fabricated of aluminum or steel that is suitably protected against corrosion.

The tube shields shall be fabricated of corrosion-resisting steel or aluminum, and shall be supplied with rubber cushions.

The exposed surfaces of the housing, tube head frame, and lever arm shall be finished in a suitable non-reflecting color.

The combined weight of the centrifuge unit (including the handcrank, tube head and tube shields) shall not exceed 4 pounds. The overall dimensions of the housing with projecting shafts and thumbscrew in the closed position (i.e. without handcrank and tube head) shall not exceed 10-1/2 inches in vertical height, 4-1/2 inches width across the crankshaft, and 3-13/16 inches deep.

Workmanship. The centrifuge shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One centrifuge, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for test shall be conducted in accordance with commercial practice.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for one or more tests.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Metric Products. Product manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6640-00-926-6913 CENTRIFUGE, LABORATORY

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS
PHS
FDA-MPQAS

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

Project No. 6640-1422

Location: NCR Disk #2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-LS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:16Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER | 2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION

b ADDRESS *(Street, City, State, ZIP Code)*

4 TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional

b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional

c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional

8 DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:16Z
Check the source to verify that this is the current version before use.