AMENDMENT

TO

INTERIM FEDERAL SPECIFICATION

CART, CLINICAL CHART HOLDER

This amendment was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended thatFederal agencies use it is procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this amendment as a valid exception to Interim Federal Specification GG-C-00129a (DSA-DM).

Paragraph 3.3.8 Push handle. Delete third sentence and substitute: Push handle shall be not less than 10 inches long, excluding bumpers, and shall be attached to the cart at a height of 33 ± 1/4 inches from the floor.

Paragraph 5.1.3.1 Unit package. Line 3, following "unit", insert: "As an alternate unit shall be packed with casters mounted to the cart."

Paragraph 5.2.1.1 Line 6, following "breakage", insert: "When unit is packed with the casters mounted, the casters shall not make contact with the bottom of the container."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1256

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.