GG-C-0080(DSA-DM)
31 March 1965

INTERIM FEDERAL SPECIFICATION

CABLE ASSEMBLY, SPECIAL PURPOSE, ELECTRICAL

This interim Federal specification was developed by the Defense Medical Supply Center, Third Avenue and 29th Street, Brooklyn, New York, 11232, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers special purpose electrical cable assemblies for use with medical diagnostic equipment.

1.2 Classification. Cable sets shall be of the following types, as specified (see 6.1):

Type I. Bayonet connector.
Type II. Positive contact connector.

2. APPLICABLE SPECIFICATIONS, STANDARDS, DRAWING AND OTHER PUBLICATIONS

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

2.1 Specifications and standards. The following specifications and standards, of the issue in effect on date of invitation for bids or request for proposal form a part of this specification to the extent specified herein:

Federal Specifications:

PPP-B-636 Box, Fiberboard.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards, from established distribution points in their agencies.)

Military Specifications:

MIL-P-12902 Packaging of Surgical, Dental and Veterinary Instruments, Diagnostic Equipment, and Related Items.

Military standards:

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

Drawing

Defense Medical Supply Center
20751 Cable Assembly, Special Purpose, Electrical.

(Copies of military specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

UNDERWRITERS' LABORATORIES, INC.

Underwriters' Laboratories Standard:

UL 62 Standard for Flexible Cord and Fixture Wire.

(Underwriters' Laboratories Standards are available from the Underwriters' Laboratories, Inc., Walt Whitman Road, Melville, L.I., N.Y. 11749, 207 East Ohio Street, Chicago, Ill. 60611, and 1655 Scott Boulevard, Santa Clara, Calif. 95050)

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 The contractor shall submit to the contracting agency, prior to delivery, proof that the insulated conductor he will supply under this specification conforms to the requirements of Underwriters' Laboratories Inc. Standard UL 62. The label or listing of the Underwriters' Laboratories, Inc., shall be acceptable as evidence that the equipment conforms to the above.

3.1.2 In lieu of the label or listing, the contractor may submit, prior to delivery, independent proof satisfactory to the contracting agency that his equipment conforms to the published standards, including methods of tests, of the Underwriters' Laboratories, Inc.

3.1.3 Compliance with Underwriters' Laboratories, Inc. requirements regarding fire and casualty hazards does not absolve the contractor from complete compliance with the requirements of this specification in order to secure the acceptance of his commodity.

3.2 Material.

3.2.1 Insulation. Insulation shall be rubber, neoprene, or polyvinyl chloride (PVC), meeting the physical and electrical requirements of UL 62 Standard for Flexible Cord and Fixture Wire.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

3.2.2 Conductor. Conductor shall be stranded, soft-annealed copper, coated or uncoated.

3.2.3 Female connector. Female connector body shall be brass chrome plated, or as an alternate, monel.

3.2.4 Pin jacks (phone tip plugs). Pin jack shall be electrically conductive brass, silver or nickel plated, with insulated handles.

3.3 Style, design and construction. The style, design and construction for type I and type II cords shall be in accordance with drawing 20751. The assembled two-conductor cords shall be terminated at one end in a female connector. The opposite end shall be terminated in a red and black handled pin jack. The portion of the cord to which the jacks are attached shall be separated for no less than 3 inches nor more than seven inches. To prevent further separation a permanently molded, or permanent mechanical means shall be used to prevent further separation when tested in accordance with 4.4.1.

3.3.1 Cord. The two-conductor insulated cord shall meet the requirements of table VI of UL 62 Standard for Flexible Cord and Fixture Wire, for type SP cords, except that stranding of the copper conductors shall be a minimum of 41 No. 34 Awg Strands (18 Awg), or as an alternate, a minimum of 65 No. 34 Awg Strands (16 Awg).

3.3.2 Female connector. Female connectors shall be in accordance with drawing 20751.

3.3.3 Pin jacks. Pin jacks (phone tip plugs), shall be electrically conductive brass, silver or nickel-plated with insulated handle. The uninsulated plug tip shall be 0.080 inch in diameter and a minimum of 1/2-inch in length; commercial tolerances shall apply (see drawing).

3.3.4 Assembled cord. Assembled cord shall be 6 feet ± 2 inches in length. One conductor shall be soldered, or crimped to the inner contact of the female connector, and terminated in a red handled pin jack. The other conductor shall be soldered or crimped to the outer contact sleeve of the female connector and terminated in a black handled pin jack. The jacks shall be tested in accordance with 4.4.1 for security of termination. The pin jacks shall not separate from the conductors, and there shall be no separation, displacement or breaking of insulation at the stops, when tested as specified in 4.4.1. The female connector shall not separate from the conductor, when tested as specified in 4.4.2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

3.4 Dielectric strength. The completed assembly, when tested in accordance with 4.4.3, shall show no evidence of breakdown.

3.5 Identification marking. The cable shall be legibly marked with the contractor's name or registered trademark.

3.6 Workmanship. Workmanship shall be first class throughout. The cable assemblies shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION AND TESTS PROCEDURES

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, except that inspection level shall be S-3, and acceptance number shall be zero for all sample sizes.

4.2.3 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:48Z
Check the source to verify that this is the current version before use.

4.3 Examination. The cables shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Security of pin jacks and stop. To determine whether the completed cord has been prepared properly to prevent further separation of the two separate conductors at the molded or mechanical stop, and to determine whether the pin jacks are securely attached to the copper conductor, the pin jacks (phone tip plugs) are to be wired in the intended manner. The cable, with the insulated handles removed, shall be supported at one pin jack, with the other pin jack projecting downward. A 20-pound weight shall be supported by this jack for one minute. The pin jacks shall be examined for separation from the conductors, the stops for separation, displacement, and breaking of the insulation.

4.4.2 Security of female connector. The cord shall be supported at the female connector; the pin jacks shall project downward. A 20-pound weight shall be supported by this connector for one minute. The connector shall be examined for separation from the conductors.

4.4.3 Dielectric test. The cord set shall be subjected to the application of a 60-cycle sinusoidal potential of 625 volts for a period of one minute. The jacks are to be separated from each other, the proper connections made to the transformer terminals and the test voltage applied. The test potential is to be supplied from a suitable 500-volt-ampere or larger capacity testing transformer, the output voltage of which can be regulated; and the wave form of the voltage should approximate a sine wave as closely as possible. The applied potential is to be increased gradually from zero until the required test voltage is reached, and is to be held at that voltage for one minute. This test shall be performed both before proceeding with, and upon the completion of the tests in 4.4.1 and 4.4.2.

5. PREPARATION FOR DELIVERY

5.1 Packaging and packing.

5.1.1 Unit of issue. One (1) cable assembly as specified, constitutes one unit of issue.

5.1.2 Level of protection. The required level of protection for packaging and packing shall be as specified in the procurement document for the required number of units of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

5.2 Procedures. Packaging and packing for the specified level of protection shall be in accordance with the requirements of MIL-P-12902, dated 20 July 1953, and Amendment-1, dated 25 August 1961, for the applicable code number as indicated in table I. Commercial colors shall be acceptable on paperboard cartons when furnished as specified. Method I closure of PPP-B-636 shall be utilized for level B fiberboard packs. In addition, the following revisions shall be effected:

In 3.4.8.2 of MIL-P-12902, at end of paragraph, add 'As an alternate, exterior container shall be a carton of appropriate size constructed in accordance with PPP-B-636, type CF, class weather-resistant. Strapping and closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4, of PPP-B-636.'

In 3.4.8.2.1 of MIL-P-12902, at end of paragraph, add 'Waterproof barrier bag shall not be required for fiberboard boxes.'

5.2.1 Quantities. The number of units of issue to be contained in each unit or intermediate package and exterior container shall be as indicated in table I.

Table I.

Type	Level of Protection	Procedure Code No.	Packaging and Packing Quantities		
			Unit Package	Intermediate Package	Exterior container
	B				
I	(II Domestic)	22	1 Unit	12 Units	96 Units
II		22	1 Unit	12 Units	72 Units
	A				
I	(III Overseas)	38	1 Unit	12 Units	96 Units
II		38	1 Unit	12 Units	72 Units

5.3 Marking.

5.3.1 Interior packages. Interior packages shall be marked in accordance with MIL-STD-129. When labels are utilized on cartons, waterproofing shall be required only when applicable cartons are fabricated from water-resistant material. Contract or purchase order number and name of contractor shall be shown. In addition, and as applicable, marking as required by 3.6.1.1 and 3.6.1.2 of MIL-P-12902 shall also be shown.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:48Z
Check the source to verify that this is the current version before use.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Type required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Military Custodians:
Army. - MD
Navy. - MS
Air Force. - 03

Preparing activity:
Defense Supply Agency. - DM

Review activities:
Army. - MD
Navy. - MS
Air Force. - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity (as indicated on reverse hereof).

SPECIFICATION

CABLE ASSEMBLY, SPECIAL PURPOSE, ELECTRICAL GG-C-0080(DSA-DM)

ORGANIZATION (Of submitter)	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS (Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)

SUBMITTED BY (Printed or typed name and activity)	DATE

DD FORM 1 APR 63 1426

JWD PPSO 12503

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:48Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

FOLD UNDER

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

OFFICIAL BUSINESS

DEFENSE SUPPLY AGENCY

DEFENSE MEDICAL SUPPLY CENTER

ATTN: DIRECTORATE OF TECHNICAL OPERATIONS (CODE TM5)

3RD AVENUE AND 29TH STREET

BROOKLYN, N.Y., 11232

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.