GG-P-00106b(DSA-DM)
8 November 1966
Interim Revision of
Fed. Spec. GG-P-106a
17 December 1956

INTERIM FEDERAL SPECIFICATION

PAPER, ARTICULATING, DENTAL

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal Specification GG-P-106a.

1. Scope and Classification

1.1 Scope. This specification covers articulating paper suitable for dental use.

1.2 Classification. Dental articulating paper shall be of the following types and classes as specified (see 6.1).

Type I. - Sheets
  Class A. - 0.004 inch thick
  Class B. - 0.014 inch thick

Type II. - Rolls
  Class C. - 0.0025 inch thick

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.

/FSC 6520/

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

FPP-B-621 Boxes, Wood, Nailed and Lock-Corner.
FPP-B-636 Box, Fiberboard.
FPP-B-676 Box, Setup.

Military

MIL-L-10547 Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

(Copies of Military specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Articulating paper shall be of suitable paper impregnated with a vegetable wax substance mixed with Prussian blue dye. The material, vegetable wax substance and dye shall be nontoxic. Dyeing shall be uniform throughout.

3.2 Properties. Articulating paper shall be suitable for use by the dental profession to ascertain degree of occlusal contact between opposing teeth. It shall have a soft cushion, rough surface and shall be supersensitive. Paper shall give a clear mark on porcelain, plastic, natural teeth and metals, wet or dry, with the slightest pressure.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

3.3 Dimensions. Dimensions of articulating paper shall be in accordance with table I.

Table I. Dimensions

Type	Class (Thickness)	Width	Length
I (Sheets)	A (0.004 to 0.005 inch)	3/4 to 7/8 inch	3-7/8 to 4-5/16 inches
I (Sheets)	B (0.0145 ± 0.0010 inch)	3/4 to 7/8 inch	3-7/8 to 4-5/16 inches
II (Rolls)	C (0.0025 to 0.003 inch)	5/8 inch	25 ft. ± 2 inches

3.4 Style and design.

3.4.1 Type I. Type I articulating paper shall be furnished in 12 individual sheets, placed within a paper cover and stapled together at 1/4 inch ± 1/32 inch from bound end of sheet. Unbound end of sheet may be free of dye and wax impregnation for a distance of not more than 3/4 inch in length. Sheets shall be furnished with or without perforations. If perforated, line of perforations shall extend not more than 3/4 inch from bound end.

3.4.2 Type II. Unless otherwise specified, type II articulating paper shall be furnished in rolls. Each roll shall be completely interleaved with tissue paper. The maximum outside diameter of the roll shall be 1.56 inches.

3.5 Workmanship. Workmanship shall be first class throughout. Articulating paper shall be free from defects which may detract from its appearance or impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105.

4.3 Examination. Articulating paper shall be examined in accordance with MIL-STD-105 and the following classification of defects.*

Table II. Classification of Defects

Categories	Defects
Critical:	None defined.
Major:	
101	Type and class not as specified.
102	Texture not as specified.
103	Dimensions not as specified.
104	Does not give clear mark on porcelain, plastic, natural teeth and metals.
105	Stapling missing, defective, or situated at distance other than specified from bound end of sheet (Type I).
106	Roll not properly wound; does not unwind uniformly and evenly (Type II).
107	Unbound end of sheet free of dye and wax impregnation for a distance greater than that specified (Type I).
Minor:	
201	Edges of paper ragged, not cleancut.
202	Number of sheets within paper cover not as specified (Type I).
203	Line of perforations (if present) extend more than specified distance from bound end of sheet (Type I).
204	Packaging and marking not as specified.

* Examination shall not be restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted as necessary to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. The unit of issue shall be as specified in Table III.

5.1.2 Packaging quantities. The number of units of issue indicated in table III shall be packaged in each unit, intermediate and exterior container, as applicable for the required level of protection specified in procurement document.

Table III. Packaging Quantities

Class	Unit of issue	Unit package (unit)	Intermediate package (units)	Exterior container (units)
A	Pkg (144 Sheets)	1	12	144
B	Pkg (72 Sheets)	1	12	144
C	Roll (1)	1	48	1,152

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a carton of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors will be acceptable. Closure shall be adequate to prevent spilling of contents under normal handling. Each carton shall be overwrapped with a moisture resistant transparent plastic sheeting. Wrapper shall be sealed.

5.1.3.2 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-636 type CF, class domestic. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit packages. Each unit package shall be marked with stock number, item identification, unit, quantity, date of manufacture (month, year), and name of contractor.

5.3.2 Intermediate packages. Each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown. Date of manufacture shall be shown in lieu of date packed.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Date of manufacture shall be shown in lieu of date packed.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Type and class required (see 1.2).
(c) Selection of applicable levels of packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION: GG-P-00106b(DSA-DM)

ORGANIZATION:

CITY AND STATE:

CONTRACT NUMBER:

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 68 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.