FEDERAL SPECIFICATION

SAW, PLASTER-CAST, ENGEL

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE

1.1 This specification covers one type of plaster-cast saw (hand type), Engel design or equal, for use in cutting through plaster casts.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification.

Federal Specification:

GG–I–526—Instruments, Dental and Surgical; General Specifications.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging, and Packing Levels.

Fed. Std. No. 123—Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

Military Specification:

MIL–P–12902—Packaging of Surgical, Dental and Veterinary Instruments, Diagnoistic Equipment, and Related Items.

Military Standards:

MIL–STD–105—Sampling Procedures and Tables for Inspection by Attributes.

MIL–STD–129—Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Saw shall be fabricated from materials in accordance with Federal Specification GG–I–526, as follows:

Component	Specification Designation	
	Class	Composition
Blade	1	E
Handle	4	N or P

3.2 Style, design, and dimensions. Shall be a plaster-cast saw, of the Engel design or equal, suitable for use in cutting through plaster casts. Shall be in accordance with figure 1.

3.3 Finish. Finish shall be mirror, or mirror and satin.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.




FIG. 1

3.4 Additional requirements. Except as specified herein, the saws shall comply with requirements of Federal Specification GG–I–526.

3.5 Workmanship. Workmanship shall be first class throughout. Saws shall be free from defects which detract from their appearance or impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Supplier responsibilities for inspection. The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

inspection facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 *Certificates of quality.* When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling for lot acceptance.

4.2.1 *Inspection lot.* The inspection lot shall be determined by the cognizant government inspector and shall be based on the unit of product being one saw (end item).

4.2.2 *Sampling for examination.* A random sample of saws shall be selected from each lot offered to the Government in accordance with Military Standard MIL–STD–105. Sampling for examination of dimensions shall be in accordance with Military Standard MIL–STD–105 and as indicated in table I.

4.2.3 *For tests.* Sampling for tests shall be conducted in accordance with Military Standard MIL–STD–105 and as indicated in table I.

4.2.4 *Raw materials.* Four ounces of each lot of each metal used in the construction of the saws shall be selected for chemical analysis.

4.2.5 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 *Examination.* The saws shall be examined to determine compliance with all requirements contained in this specification.

4.4 *Tests.* Tests shall be conducted, as necessary, to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

For civil agency procurement, Federal Standard No. 102 should be referred to for definitions and applications of the various levels of packaging and packing protection for supplies and equipment.

TABLE I.—*Sampling criteria*

Component	Characteristic	Req.	Test procedure	Unit of product	Insp. level	AQL (percent defective)
Blade	Thickness of coating	GG–I–526	GG–I–526	1—saw	L3*	**
	Hardness	GG–I–526	GG–I–526	1—saw	L3*	**
	Salt-spray test	GG–I–526	GG–I–526	1—saw	L3*	**
	Boil test	GG–I–526	GG–I–526	1—saw	L3*	**
	Dimensions	3.2	.....	1—saw	L3	4.0
Handle	Thickness of coating	GG–I–526	GG–I–526	1—saw	L3*	**
	Salt-spray test	GG–I–526	GG–I–526	1—saw	L3*	**
	Boil test	GG–I–526	GG–I–526	1—saw	L3*	**
	Dimensions	3.2	.....	1—saw	L3	4.0

* A minimum of three units shall be tested regardless of lot size.

** Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

5.1 Packaging and packing.

5.1.1 *Unit of issue.* One saw, as specified, constitutes one unit of issue.

5.1.2 *Packaging quantities.* The number of units of issue indicated in table II shall be packaged in each unit, intermediate and exterior container, as applicable for the required level of protection specified in the procurement document.

TABLE II.—*Packaging quantities*

Unit package	Intermediate package	Exterior container
1 unit	6 units	96 units

5.1.2.1 *Packing variation permitted.* If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.* Packaging and packing shall conform to the requirements of Military Specification MIL–P–12902, Procedure Code No. 35.

5.1.4 *Level B.* Packaging and packing shall conform to the requirements of Military Specification MIL–P–12902, Procedure Code No. 20.

5.1.5 *Level C.* The subject commodity shall be packaged and packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2 Marking—nonmilitary.

5.2.1 Unless otherwise specified in the contract or order unit packages and shipping containers shall be marked in accordance with Federal Standard No. 123.

5.3 Marking—military.

5.3.1 *Interior packages.* Marking shall conform to the requirements of Military Specification MIL–P–12902, except that the ASMPA stock number will not be required.

5.3.2 *Exterior container.* Exterior container shall be marked in accordance with Military Standard MIL–STD–129.

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(*a*) Title, number, and date of this specification.

(*b*) Selection of applicable levels of packaging and packing (see 5.1).

(1) Level B preservation and packaging is intended to provide economical, but limited, protection and should be specified only when it is determined the saws will be held in covered storage not more than 12 months from date of initial packaging.

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Federal Government requirements.

6.3 Transportation description. Transportation description applicable to this item is:

Rail:

Surgical instruments, not otherwise indexed by name.

Carload minimum weight 10,000 pounds, subject to Rule 34, Uniform Freight Classification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

Motor:

Surgical instruments, not otherwise indexed. Truckload minimum weight, 10,000 pounds, subject to Rule 34, National Motor Freight Classification.

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied said drawings, specifications, or other data is not to regarded by implication or otherwise as in any man licensing the holder or any other person or corporati or conveying any rights or permission to manufacti use, or sell any patented invention that may in any v be related thereto.

MILITARY CUSTODIANS:

Army—M
Navy—Md
Air Force—AFCSG

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:49Z
Check the source to verify that this is the current version before use.