GG-W-101b

SEPTEMBER 25, 1957

SUPERSEDING
Int. Fed. Spec. GG-W-00101a (GSA-FSS)
April 17, 1956, and
Fed. Spec. GG-W-101
May 2, 1944

FEDERAL SPECIFICATION

WASHERS, BEDPAN AND URINAL

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope.—This specification covers bedpan and urinal washers suitable for use in hospitals.

1.2 Classification.

1.2.1 *Types and classes.*—Bedpan and urinal washers shall be of the following types and classes, as specified (see 6.3):

Type I.—Floor mounted.
Class A.—With vertical front door.
Class B.—With horizontal top cover.
Type II.—Wall hung.
Type III.—Wall recessed.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

UU-P-271 — Paper, Wrapping, Waterproof, Kraft.
WW-P-541—Plumbing Fixtures, Land Use.
PPP-B-601—Boxes, Wood, Cleated-Plywood.
PPP-B-621—Boxes, Wood, Nailed and Lock-Corner.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging and Packing Levels.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Standards:

MIL-STD-105 — Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 — Marking for Shipment and Storage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

(Copies of Military Specifications and Standards required by the contractor in connection with specific procurement functions, should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 **Material.**—Material, including plating, enameling castings, and pipe shall be suitable for the purpose and shall conform to the requirements of Federal Specification WW-P-541. The interior of the washing chamber and cover, and any other internal parts, shall be of vitreous enameled, iron or steel, or brass or bronze. External portions shall be suitably plated or finished to resist corrosion. Exposed pipes, valves, and fittings shall be steel, iron, brass or bronze, except that the trap may be cast iron, enameled inside. The outside of the cast iron traps shall be porcelain or vitreous enameled.

3.2 **Design.**—Bedpan and urinal washers shall be designed for steam flushing, and shall be complete and suitable for the type of mounting specified (see 1.2.1).

3.2.1 *Type I, floor mounted.*—Type I, floor mounted bedpan washer shall be provided with pedestal and adjustable flanges for floor support, independent of wall supports and plumbing connections, except that the adjustable flanges will not be required where waste line comes up through the floor.

3.2.2 *Type II, wall hung.*—Type II, wall-hung bedpan washer shall have suitable supporting flanges.

3.2.3 *Type III, wall recessed.*—Type III, wall recessed bedpan washers shall have a suitable metal support for mounting in the wall.

3.3 **Construction.** — Bedpan and urinal washers shall be so constructed as to be suitable for emptying, washing, and steam flushing.

3.3.1 *Door and pan rack.*—Washers, with the exception of type I, class B, shall have a front vertical door operated by a quiet functioning foot pedal. When the door is opened, the pan rack shall be brought into position; the bedpan or urinal shall be turned over and emptied by the closing of the door. The door shall be provided with a suitable check or gravity counterbalance to prevent slamming. The door or cover shall be odor tight, and seated on a gasket or ground joint, or be provided with a water seal or baffle to prevent leakage. Door shall have substantial hinges which provide for proper seating of the door, under all operating conditions. The pan rack shall accommodate various sizes, of bedpans and urinals and shall hold them firmly in position so as to prevent rattle or noise.

3.3.2 *Hopper.*—Washers shall have a round hopper, or a rectangular hopper with rounded corners, and shall have a three-inch trap and a brass or bronze clean-out plug, drain connections, and not less than 1½ inch diameter vent connection arranged to provide adequate aeration of the chamber. The washing chamber, pan rack, and other internal parts shall have rounded corners to prevent accumulation of fecal matter, and shall have a smooth free passageway to bottom of trap.

3.3.3 *Flush valve.*—A hand or foot operated flushing valve shall be fitted for flushing the bedpan or urinal. Bedpans and urinals shall be effectively washed inside and outside with water at a minimum pressure of 20 pounds per square inch (p.s.i.). Water connections with adjustable flush valve, set to maintain a steady flow of water for 25 seconds, shall be fitted. Water connection and valve shall be not less than 1 inch iron pipe size (I.P.S.). A steam connection with valve shall be provided so that the bedpan or urinal may be flushed with jets of live steam. A steam connection with valve shall be provided so as to regulate flow of steam. The steam valve shall have a handle of heat-insulating material. Steam connection shall be not less than ⅜ inch I.P.S. The washer shall operate

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

satisfactorily at steam pressures from 10 to 35 p.s.i. Flushing valve shall be of rugged type, and shall have a back flow preventer at least four inches above the highest water level, including the level of the water seal (if furnished), so arranged as to prevent syphoning the contents of the bedpan and urinal washer back into the water supply line. Flush valve and back flow preventer shall conform to the requirements of Federal Specification WW–P–541.

3.3.4 *Fittings.*—All pipes, flanges and fittings (other than trap) shall be I.P.S., and shall be complete to floor and wall, at which points necessary and suitable adjustable flanges shall be provided.

3.4 **Identification marking.**—Each washer shall be permanently and legibly marked with a metal nameplate permanently affixed thereto. Marking shall include name of the manufacturer or registered trade-mark, manufacturer's type, and the maximum steam pressure which should be connected to the steam valve. Decalcomanias shall not be used. In addition, all valves shall be suitably labeled to indicate their function.

3.5 **Service data.**—The contractor shall furnish with each unit two copies of service data containing complete instructions for installation, operation, maintenance, and a list of component parts. This information will specifically include instructions, drawings, and templates for plumbing, structural modifications, and special devices necessary for roughing-in preparations prior to installation of the unit. The service data shall include such detailed information which will permit ready installation and maintenance of the item. All information contained in the manual shall be suitably indexed to permit ready reference to the required information. Drawings and descriptive material shall include adequate information so that the replacement parts can be ordered by the manufacturer's part number, and without confusion as to the item identificatiion. Identical or interchangeable replacement parts will be identified by a single part number only. One set of the service data, as indicated above, will be placed in a sealed, waterproof envelope, and affixed securely to the outside of the shipping container. The envelope shall be secured to the shipping container by use of metal ties, nails, or similar means, and will be located at the end of the shipping container, preferably, when space permits, at the end bearing the markings required by Military Standard MIL–STD–129. In cases where complete service data is too bulky to be conveniently attached to the outside of the shipping container, the contents of the exterior envelope may be limited to information necessary for roughing-in preparations and a note shall be made on the exterior envelope, which will indicate where the remainder of the first set of service data will be located. When two or more containers are required for packing a disassembled item (components of a single stock number), set markings shall be applied in accordance with Military Standard MIL–STD–129. The envelope shall be affixed to the container designated as No. 1 of the set. The second set, and if necessary, the remainder of the first set of service data, will be attached to the unit inside the shipping container. In addition, the contractor shall mail six copies of the service data to the contracting officer for distribution.

3.6 **Workmanship.**—Workmanship shall be first class throughout. Bedpan and urinal washers shall be free from defects which detract from their appearance or which may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 **Sampling.**—Unless otherwise specified, samples shall be submitted in accordance with instructions furnished by the procuring activity (see 6.3). In addition, sampling procedures shall be conducted in accordance with the procedures set forth in Military Standard MIL–STD–105.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.

6.2 If boiling type bedpan sterilizers are required, see Federal Specification GG–S–751, Sterilizers, of the issue in effect on date of invitation for bids.

6.3 The following should be specified in the invitation for bids as applicable, in addition to requiring that contractors shall replace without cost to the Government any material found to be defective within 1 year from date of installation:

(*a*) Type, class, and mounting. (See 1.2.1).

(*b*) Sampling and inspection required. (See 4.1 and 4.2).

(*c*) Packaging and packing required. (See 5.1 and 5.2).

6.4 It is believed that this specification adequately describes the characteristics necessary to secure the desired material and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid samples should be definitely stated, the specification to apply in all other respects.

6.5 Federal specifications do not include all types, classes, grades, sizes, etc., of the commodities indicated by the titles of the specifications, or which are commercially available, but are intended to cover the types, etc., which are suitable for Federal Govern requirements.

6.6 Transportation description.—Tran tation description applicable to these i is:

Dishwashing Machines, Other Than H hold Type:

Carload minimum weight 1 pounds, subject to Rule 34, Uni Freight Classification.

Motor Volume minimum weight 1 pounds.

6.7 This specification covers the follo item appearing in the Armed Services M cal Stock List:

6530–708–5255 — Washer, Bedpan Urinal, Steam Heated, Floor Mou

Notice. — When Government drawings, fications, or other data are used for any purpose than in connection with a definitely related Go ment procurement operation, the United States ernment thereby incurs no responsibility nor obligation whatsoever; and the fact that the Go ment may have formulated, furnished, or in any supplied the said drawings, specifications, or data, is not to be regarded by implication or other as in any manner licensing the holder or any person or corporation, or conveying any rights o mission to manufacture, use or sell any pat invention that may in any way be related th

MILITARY INTERESTS:
Army—MI
Navy—Md
Air Force—AF.

Copies of this specification may be purchased for 5 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:48Z
Check the source to verify that this is the current version before use.