MILITARY SPECIFICATION

TEST KIT, SERUM TRANSAMINASE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers Serum Transaminase Test Kit for 100 determinations (see 6.2)

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-C-186	- Containers, Packaging and Packing for Drugs, Chemicals and Pharmaceuticals.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document, or by letter.

FSC 6505

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:28Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U.S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION (FDA).

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

U.S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia (USP).

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

The National Formulary (NF).

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. Shall be a colorimetric test kit for the determination of serum glutamic-oxalacetic transaminase. The components of the kit shall be sufficient for the performance of 100 tests. In addition, the test kit shall be in compliance with all applicable requirements of the Bureau of Medical Devices and Diagnostics.

3.1.1 Methodology. The kit shall use a modified Babson, Shapiro, Williams and Phillips methodology.

3.2 Components of the kit. The components of the kit shall consist of the following;

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:28Z
Check the source to verify that this is the current version before use.

A. Buffered substrate solution. The buffered substrate solution shall be a solution, buffered to between pH 7.3 and 7.5, and shall contain alpha-ketoglutaric acid, L-aspartic acid, sodium edetate, dibasic, potassium phosphate, monobasic potassium phosphate, and polyvinylpyrrolidone.

B. Control buffer solution. The control buffer solution shall be buffered to between pH 7.3 and 7.5, and shall contain dibasic potassium phosphate, monobasic potassium phosphate and sodium edetate.

C. Color developer tablets. The color developer tablets shall contain 6-benzamido-4 methoxy-m-toluidine diazonium chloride.

D. Ethoxylated tridecyl alcohol.

E. Instructions for use.

3.5 Expiration dating period. The expiration dating period shall be not less than 36 months. All reagent components of the kit shall bear the same expiration date as the kit.

3.5.1 Delivery. Not more than 3 months of the expiration dating period shall have elapsed at the time of delivery to the Government.

3.6 Workmanship. The material and its containers shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period specified by the contractor for such item or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the item to the Government.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purposes of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of product produced from one formulation, subjected to the same compounding and manufacturing operation, and filled into final containers.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:28Z
Check the source to verify that this is the currentversion before use.

4.3 Sampling. Sampling shall be conducted in accordance with MIL-STD-105.

TABLE I. Sampling.

For visual examination	Inspection level	AQL (percent defective)	Unit of product
Major	S-2	1.0	Unit of issue
Minor	II	2.5	Unit of issue

4.4 Examination. The finished kit shall be examined to determine compliance with all requirements of this specification. Nonconformance will be permitted to the extent indicated in 4.3.

4.4.1 Classification of defects. Examination will be conducted in accordance with the following classification of defects:

TABLE II. Classification of defects.

Categories	Defects 1/
Major	
101	Any component of the kit missing (see 3.2).
Minor	
201	Exteriors of the immediate containers are not free from contents or foreign matter.
202	Label not free from defacing marks or tears.

1/ Inspection is not restricted to the classified possible defects listed above.

4.5 Packaging inspection. The inspection of the intermediate package, packing an marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and marking requirements of MIL-STD-129.

4.6 Labeling inspection. Sampling shall be in accordance with MIL-STD-105 and th unit of product shall be the unit of issue. The inspection lot shall be the number units of issue of each batch, lot control number, etc., offered for inspection at on time. The inspection level for labeling shall be S-2 and the acceptance number shall be zero.

4.7 Tests and examinations. Tests and examinations shall be performed to determi compliance with all requirements of this document and all referenced documents.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:28Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Packaging requirements. The requirements for packaging shall be in accordance th all applicable requirements of PPP-C-186, and as specified herein.

5.2 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.2.1 Unit of issue. Each test kit consisting of one bottle of buffered substrate, 0 ml; control buffered solution, 25 ml; ethoxylated tridecyl alcohol, 5 ml; and 12 blets in aluminum foil, as specified, constitutes one unit of issue. As an alternate, tablets shall be supplied in a vial.

5.2.2 Level A.

5.2.2.1 Immediate container. The immediate container for each reagent shall be a mmercially available immediate container considered acceptable by the Food and Drug lministration (FDA) for the product contained therein and shall be of such com- atibility that neither the contents nor the container are altered in any way by each :her.

5.2.2.2 Unit package. Unit package shall be a box of appropriate size. Carton esign shall include a die-cut polyurethane or polystyrene fitting so designed as to old each immediate container in an individual cell. Closure shall be adequate to revent spilling of contents under normal handling.

5.2.2.3 Intermediate package. Six units of issue shall be packaged in a box of ppropriate size and design constructed in accordance with PPP-B-636, type CF, class omestic. Closure shall conform to method I as specified in PPP-B-636.

5.2.3 Level C. Units shall be packaged in standard commercial containers of the ize and kind commonly used, which will afford the degree of protection required for hipment and use of the product for its intended purpose.

5.3 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.3.1 Level A. Twenty-four units of issue shall be packed as specified in 5.4.4 f PPP-C-186.

5.3.2 Level B. Twenty-four (24) units of issue shall be packed as specified in 5.4.3.4 of PPP-C-186.

5.3.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, to point of delivery called for in the contract or purchase order.

5.4 Marking.

5.4.1 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug and Cosmetic Act and regulations promulgated thereunder, and shall include the information required below:

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:28Z
Check the source to verify that this is the current version before use.

5.4.1.1 Immediate container: Each immediate container label shall bear the following information:

(a) label information in accordance with commercial practice.

(b) the expiration date.

5.4.1.2 Unit package. Each unit package label shall bear the following information:

(a) label information in accordance with commercial practice.

(b) the National Stock Number

NOTES: (1) The National Stock Number need not be preceded by any prefix.
(2) Strip labeling may be used.
(3) Label adhesion test shall not apply.
(4) When a package insert (instructions for use) is utilized to comply w statutory labeling requirements, the package insert shall accompany of issue so that it will not become separated from the unit of issue

5.4.2 Intermediate package. Intermediate package shall be marked as specified in 5.5.3 of PPP-C-186. Type I shelf life marking as specified in MIL-STD-129 shall be shown. Marking shall include the legend:

"STORE BETWEEN 2°-8° C (35°-46° F)".

5.4.3 Exterior container. Exterior containet shall be marked as specified in 5.5.4 of PPP-C-186. Type I shelf life marking as specified in MIL-STD-129 shall be shown. Marking shall include the legend:

"STORE BETWEEN 2° - 8°C (35° - 46°F)."

NOTE: Unrefrigerated shipping time for this item shall be not more than 6 days (144 hours).

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.

(b) National Stock Number.

(c) Selection of applicable levels of packaging and packing (see 5.2 and 5.3).

6.2 This specification covers the following item:

National Stock Number	Item Identification
6505-00-586-9335	TEST KIT, SERUM TRANSAMINASE, 100 Determinations

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DoD - MB

Preparing activity:
DLA-DM

Project No. 6505-1918

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:28Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-T-37771 TEST KIT, SERUM TRANSAMINASE

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*

TELEPHONE NO.

DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:28Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20th STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:28Z
Check the source to verify that this is the current version before use.