MILITARY SPECIFICATION

TEST KIT, PREGNANCY DETERMINATION

This specification is approved for use by Defense Personnel Support Center, Defense Logistics Agency, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers Test Kit, Pregnancy Determination (see 6.2).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-C-186 - Containers, Packaging and Packing for Drugs, Chemicals and Pharmaceuticals.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

Beneficial comments (recommendations, additions, deletions) and any pertient data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6505

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:32Z
Check the source to verify that this is the current version before use.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION (FDA)

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

U. S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia (USP).

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

The National Formulary (NF).

(Applications for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. Shall be as specified hereafter for the specified component. The kit shall consist of the following components necessary for performing 100 slide tests for pregnancy:

3.1.1 Components. Each kit shall contain the following components:

(a) One hundred disposable slides each containing two dried reagents; latex coated with Human Chorionic Gonadotropin and Anti-Human Chorionic Gonadotropic Serum (Rabbit), dried as individual spots on the disposable slide.

(b) One vial of positive urine control freeze-dried and one vial of negative urine control freeze-dried.

(c) One hundred dispenser/stirrers capable of dispensing 0.03 ml per drop.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:32Z
Check the source to verify that this is the current version before use.

(d) One dropper vial for water.

3.1.2 Suitability. Shall be suitable as an in vitro qualitative latex inhibition test to detect Human Chorionic Gonadotropin in urine.

3.2 Each lot of Test Kits, Pregnancy Determination, shall be released by the Bureau of Biologics (B of B), FDA.

3.3 Sensitivity test. Shall be capable of detecting between 1 to 2 I.U. Human Chorionic Gonadotropin (HCG) per ml when tested as specified in 4.7.1.

3.4 Reaction test. Shall be capable of detecting a positive response (no agglutination) and negative response (agglutination) when tested as specified in 4.7.2.

3.5 Ingredients. All ingredients except for Human Chorionic Gonadotropin used in the manufacture of the test kit shall be in accordance with the tests, standards, and requirements of the USP or NF including any supplements or revisions thereto; or, if not included in either of these compendia, the ingredients shall be of the highest pharmaceutical grade.

3.6 Expiration dating period. The expiration dating period shall be not less than 24 months.

3.6.1 Delivery. Not more than 3 months of the expiration dating period shall have elapsed at the time of delivery to the Government.

3.7 Workmanship. The material and its containers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:32Z
Check the source to verify that this is the current version before use.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the item to the Government.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purposes of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of product produced from one formulation, subjected to the same compounding and manufacturing operation, and filled into final containers.

4.3 Sampling. Sampling shall be conducted in accordance with MIL-STD-105.

TABLE I. Sampling.

For visual examination	Inspection level	AQL (percent defective)	Unit of product
Major	S-2	1.0	Each (unit package)
Minor	II	2.5	Each (unit package)

4.3.1 For end item testing. Noncompliance with any of the following is cause for rejection.

TABLE II. Criteria for end item testing.

Characteristic	Requirement paragraph	Inspection level 1/	Minimum sample
Components	3.1.1	S-2	5
Sensitivity test	3.3	S-2	5 } Composite
Reaction test	3.4	S-2	5

1/ Where possible, the sample shall be used for two or more tests.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:32Z
Check the source to verify that this is the current version before use.

4.4 Examination. The finished Pregnancy Test Kit shall be examined to determine compliance with all requirements of this specification. Nonconformance will be permitted to the extent indicated in 4.3 and 4.3.1.

4.4.1 Classification of defects. Examination will be conducted in accordance with the following classification of defects:

TABLE III. Classification of defects.

Categories	Defects 1/
Major	
101	A minimum of 100 disposable slides containing the dried latex coated with HCG and anti-HCG serum (rabbit) not included with each kit.
102	A minimum of 100 dispenser/stirrers not included with each kit.
103	Urine controls (positive and negative), freeze-dried in a vial not included with each kit.
104	One dropper vial for water not included with each kit.
Minor	
201	Exterior of kit not free from contents or foreign matter.
202	Labels for the components of the kit not free from defacing marks and tears.

1/ Inspection is not restricted to the classified possible defects listed above.

4.5 Packaging inspection. The inspection of the intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.6 Labeling inspection. Sampling shall be in accordance with MIL-STD-105 and the unit of product shall be the unit of issue. The inspection lot shall be the number of units of issue of each batch, lot, control number, etc., offered for inspection at one time. The inspection level for labeling shall be S-2 and the acceptance number shall be zero.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:32Z
Check the source to verify that this is the current version before use.

4.7 Tests.

4.7.1 Sensitivity test. The following method shall be employed in the detection of the sensitivity of the pregnancy kit. Prepare diluted standard solutions from normal, non-pregnant female urines containing 0, 500, 1000, 2000, 3000, 4000 and 6000 I.U. of HCG per liter. These final diluted standard solutions (0, 0.5, 1,2, 3, 4 and 6 HCG/ml) are tested in place of urine according to the procedure in the labeling. A positive test reaction (no agglutination) shall occur in the standard solutions with a final concentration equal to the sensitivity specified in 3.3 (1-2 I.U. HCG/ml). Note: Final concentrations of standard solutions 1 I.U. HCG/ml or greater shall give a positive test reaction.

4.7.2 Reaction test. The test kit shall be tested against ten (10) known positive and ten (10) known negative pregnant female urines (controls) using the method specified in the labeling. Positive responses (no agglutination) shall result when the positive control is used and negative responses (macro-agglutination, clumping) shall result when the negative control is used. A negative test shall be visible within two minutes and a positive test shall occur at end of two minutes.

5. PACKAGING

5.1 Packaging requirements. The requirements for packaging shall be in accordance with all applicable requirements of PPP-C-186 and as specified herein.

5.2 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.2.1 Unit of issue. One test kit, as specified, constitutes one unit of issue.

5.2.2 Level A.

5.2.2.1 Immediate container. The immediate container for the disposable slide and the positive and negative urine controls shall be commercially available container considered acceptable by the Food and Drug Administration (FDA) for the product contained therein and shall be of such compatibility that neither the contents nor the container are altered in any way by each other.

5.2.2.2 Unit package. Each unit shall be packaged as specified in 5.2.5 of PPP-C-186.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:32Z
Check the source to verify that this is the currentversion before use.

5.2.2.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.3 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.3.1 Level A. Twenty-three (23) units of issue shall be packed as specified in 5.4.4 of PPP-C-186.

5.3.2 Level B. Twenty-three (23) units of issue shall be packed as specified in 5.4.3.2 of PPP-C-186.

5.3.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.4 Marking.

5.4.1 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug and Cosmetic Act, and shall include the information required below:

5.4.1.1 Immediate container. Each immediate container label shall bear the following information:

(a) label information in accordance with commercial practice.
(b) the expiration date.

5.4.1.2 Unit package. Each unit package label shall bear the following information:

(a) label information in accordance with commercial practice.
(b) the National Stock Number.

NOTES: (1) The National Stock Number shall not be preceded by any prefix.
(2) Strip labeling may be used.
(3) Label adhesion test shall not apply.

5.4.2

5.4.2 Intermediate package. Intermediate package shall be marked as specified in 5.5.3 of PPP-C-186. Type I shelf life markings as specified in MIL-STD-129 shall be shown. Marking shall include the legend: "STORE AT CONTROLLED ROOM TEMPERATURE 15° - 30°C (59° - 86°F)".

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:32Z
Check the source to verify that this is the current version before use.

5.4.3 Exterior container. Exterior container shall be marked as specified in 5.5.4 of PPP-C-186. Type I shelf life marking as specified in MIL-STD-129 shall be shown. Marking shall included the legend: "STORE AT CONTROLLED ROOM TEMPERATURE 15° - 30°C (59° - 86°F)".

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title; number and date of this specification.
(b) National Stock Number.
(c) Selection of applicable levels of packaging and packing (see 5.2 and 5.3).

6.2 This specification covers the following item:

National Stock Number	Item Identification
6505-01-037-4877	TEST KIT, PREGNANCY DETERMINATION, 100 Tests

Preparing activity:
DLA-DM

Project No. 6505-1949

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:32Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-T-37805(DLA-DM) TEST KIT, PREGNANCY DETERMINATION

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO 1974— 603-680 6126 2-1

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:32Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:32Z
Check the source to verify that this is the current version before use.