GG-C-001321(DSA-DM)
13 October 1967
See Section 6 (6.3)

INTERIM FEDERAL SPECIFICATION

CASTING MACHINE, DENTAL METALS, and REPLACEMENT PARTS

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers a dental casting machine (with replacement parts) used to cast molten alloy gold from a crucible into a flask mold by means of centrifugal force.

1.2 Classification.

1.2.1 Casting machine, complete. The casting machine shall be furnished complete with all component parts as specified (see 3.2 and 6.1).

1.2.2 Replacement parts. Replacement parts for the casting machine shall be of the following designated types and sizes, as specified (see 6.1):

Type I. Flasks.
- Size 1. Inlay.
- Size 2. Saddle and bridge.
- Size 3. Partial.
- Size 4. Full denture.

Type II. Crucible formers.
- Size 1. Small.
- Size 2. Large.

Type III. Crucibles.
- Size 1. Small.
- Size 2. Large.

Type IV. Asbestos strip
- Size 1. 100 feet.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

2. APPLICABLE DOCUMENTS

2.1 The following specifications, standards, and drawing of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

MMM-A-250	Adhesive, Water Resistant, (For Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (For Carton Sealing).

STANDARDS

Federal

FED-STD-123	Marking for Domestic Shipment (Civilian Agencies).
FED-STD-151	Metals; Test Methods.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center

20705 Casting Machine, Dental Metals

(Copies of Military standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS.

3.1 Material. The casting machine's components shall be fabricated of the materials shown in table I. Materials shall be new and shall conform to the applicable specifications and requirements hereinafter specified. Material not specifically designated shall be of high commercial quality throughout and shall be entirely suitable for the purpose intended.

Table I. Materials

Material	Casting Machine Components
Die cast zinc or sand cast iron	Base housing; cover
Case hardened steel	Clutch; rotor
Machine steel	Rotor shaft
High grade spring steel	Coiled main spring
Nickel plated brass	Stop pin
Nickel plated cast aluminum bronze	Straight arm
Brass or cast iron	Counterbalance weights
Nickel plated cast aluminum bronze	Sling arm
Chromium over nickel plated aluminum bronze	Crucible mount

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table I. Materials (cont.)

Material	Casting Machine Components
18-8 Seamless or welded corrosion-resisting steel tubing	Flasks
Chromium over nickel plated brass or highly polished corrosion-resisting steel	Crucible formers
Corrosion resisting steel, #21 Birmingham Wire Gage (BWG)	Cradles
Nickel plated spring steel	Tongs
Corrosion resisting steel wire #14 Brown and Sharpe (B&S) gage	Sprue pins
Rubber	Stop pin cap
Asbestos strip	Liners for crucible, casting flasks.
Fire clay	Crucible

3.2 Description. The casting machine shall be suitable for casting molten alloy golds with a controlled centrifugal force. The casting machine shall consist essentially of a circular base housing with cover, spring, upright clutch rotor and shaft, straight arm, with integrally pivoting sling arm carrying cradle, crucible and flask assembly, and a stop pin. The arms shall rotate in a horizontal plane as the shaft is turned by the unwinding mainspring. When not in use the sling arm shall rest against a stop on the straight arm in a position nearly at right angles to straight arm, but shall be free to swing into alignment with the straight arm when the machine is in rotation. The basic machine shall be furnished complete with the following component parts in the quantities shown in table II.

Table II. Basic Machine and Component Parts

Item	Component parts	Quantity	Reference
Basic machine, complete		1	3.3.1 through 3.3.1.8
	Weights.		3.3.1.6
	Permanent.	1	
	Lock.	1	
	Extra weights.	2	

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

Table II. Basic Machine and Component Parts (cont.)

Item	Component parts	Quantity	Reference
	Flasks.		3.3.1.12
	Inlay.	2	
	Saddle and bridge.	1	
	Partial.	1	
	Full denture.	1	
	Crucible formers		3.3.1.13
	Small.	2	3.3.1.13.1
	Large.	1	3.3.1.13.2
	Crucibles.		3.3.1.14
	Small.	1	3.3.1.14.1
	Large.	1	3.3.1.14.2
	Cradles.		3.3.1.9
	Small.	1	
	Medium.	1	
	Large.	1	
	Tongs.		3.3.1.10
	Small.	1	
	Large.	1	
	Sprue pins	Package of 12	3.3.1.11
	Asbestos strip	1 roll	3.3.1.15

3.3 Style, design, and dimensions. The style, design, and dimensions of the casting machine with component parts, and replacement parts shall be in accordance with the Defense Personnel Support Center (DPSC) Drawing 20705, and as specified herein.

3.3.1 Casting machine, complete.

3.3.1.1 Base housing and cover. The base housing and cover shall house the flat main coiled spring, chuck rotor and shaft and ball bearing supports, and stop pin. The circular cover shall be secured to the body housing by four screws spaced 90 degrees apart. The body base flange shall have four 3/16 to 1/4 inch diameter countersunk holes spaced 90 degrees apart, for screw fastening the complete unit to a solid level foundation. The base housing and cover shall be finished in a smooth black enamel.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

3.3.1.2 Clutch and rotor shaft. The rotor shaft shall be 6-3/4 ± 1/8 inches long. The bottom of the shaft shall rest on a steel bearing plate, and the upper threaded end shall project through the circular opening in the center of the cover to an exposed length of 1-5/8 ± 1/8 inches. The shaft shall turn on ball bearing supports located at the bottom and at a distance within 2 inches from the top. The shaft shall be attached to the spring through the clutch body. A pin, or other device used for engaging the straight arm, shall be located 1-1/8 ± 1/8 inches below the top of the threaded end of the shaft. The end shall be threaded to a minimum length of 7/8 inches, onto which the straight arm shall be placed and secured by a nickel plated brass knurled retaining nut.

3.3.1.3 Main spring and assembly. The main spring shall be at least one inch wide and a minimum of 0.032 inch thick. The coiled spring shall rest on an annular steel plate or shelf built inside the body, positioned at a distance below the top rim of the body so that the edge of the supported spring does not protrude above the body rim. The coiled spring shall be adequately secured to the clutch body which in turn shall be secured to the rotor and shaft. Tension on spring shall be accomplished by winding the straight arm. When at rest, the spring shall coil to fit snugly into the base.

3.3.1.4 Stop pin. The stop pin shall be 5-1/16 ± 1/16 inches long, with a 3/8 ± 1/64 inch diameter. The pin shall move freely in the receptacle hole provided in the cover and base housing. The properly located pin shall prevent rotation of the wound-up machine arms when in raised position, and shall release the arms to rotate when dropped into its receptacle in the recessed position.

3.3.1.5 Straight arm. The straight arm shall be mounted onto the exposed section of the shaft so that the recess in the bottom of the arm hole shall engage or rest on the balancing pin of the shaft. The extreme (peripheral) end of the straight arm shall carry the counter balance weight assembly. Pivoted on ball bearing at the opposite end of the straight arm, so as to move in a horizontal plane, shall be a secondary or sling arm. A stop, positioned on the top side of the straight arm near the ball bearing pivot, shall serve to limit the amount of rotation or free swing of the sling arm.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.3.1.6 Counterbalance weights. Secured to the vertical extension at extreme end of the straight arm, shall be the counterbalance weight assembly, consisting of a lock weight and a permanent weight, mounted on threaded weight shaft. To balance large flask, two additional weights, weighing 10 ± 1/2 ounces each, shall be furnished and shall be easily adapted onto the counterbalance weight assembly. All counterbalance weights shall be finished in a durable black crinkle finish.

3.3.1.7 Sling arm. The sling arm shall be designed and constructed to properly carry the casting flask assembly. With the machine at rest, the sling arm, limited by the stop on the straight arm, shall assume a position almost perpendicular to the straight arm. When in motion, the sling arm shall align itself with the straight arm. The sling arm shall have a slot along its longitudinal axis, at the pivot end of the arm to facilitate installation of the flasks and crucible formers. The extreme end of the sling arm shall consist of a vertical, integrally cast, circular plate on which two cradle pins (1/4 ± 1/64 inch diameter), 3-7/8 inches apart, shall be secured. The cradle pins shall support the required flask-bearing cradles.

3.3.1.8 Crucible mount body. A crucible mount body, carrying a horizontally positioned circular baffle plate and vertical asbestos-backed stainless steel back plate, shall be free to move along the slot in the sling arm. The crucible mount body shall be loosely secured to the sling arm by means of a circular screw passing through the bottom of the slot in the sling arm trough, and screwed into the mount body. The clay crucibles shall fit into the recesses of the crucible mount, with the bottom surface of the crucible resting on the horizontal baffle plate, and front face of crucible resting against vertical back plate.

3.3.1.9 Cradles. Cradles shall be suitable for mounting on the two cradle pins and shall support and center the corresponding flask, with the crucible outlet hole. Each cradle shall be permanently and legibly marked with a coded size designation to correspond with coded markings on flasks.

3.3.1.10 Tongs. Each casting machine shall be furnished with two pair of tongs. The tongs shall be of one-piece construction, rounded at the holding end and curved at the operative end for grasping the given sized flask. Grasping ends shall meet when in closed position. The small tongs shall be 11-1/2 ± 1/2 inches long, 1 ± 1/16 inch wide, and 1/16 ± 1/64 inch thick.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

In the open position, the tips at grasping ends shall be 3-1/2 ± 1/2 inches apart. Each member of the small tongs shall have a depression along the longitudinal axis to provide the necessary rigidity required of the item. The small tongs shall be suitable for use with inlay and saddle and bridge flasks. The large tongs shall be 15 ± 1/2 inches long, 9/16 ± 1/16 inch wide, 1/8 ± 1/64 inch thick and 4 ± 1/2 inches apart at grasping tip ends. Large tongs shall be suitable for use with partial and full denture flasks.

3.3.1.11 Sprue pins. A separate envelope, labeled as to contents, containing one dozen sprue pins, shall be furnished with each machine. The sprue pins shall each be 3/4 ± 1/16 inch long.

3.3.1.12 Type I, flasks. Flasks, made of the welded tubing specified in 3.1, shall pass the test specified in 4.4.2 without cracking or showing of flaws. Ends shall be free from burs and the inside surface shall be smooth and free from pits or elevations. Flasks shall set properly on related size formers and cradles, except for the full denture flask which shall sit directly on upper trough of sling arm. Flasks shall be suitable for use in the preparation of gold restorations and appliances. Size 1 flasks shall be suitable for use in casting inlays. Size 2 flasks shall be suitable for use in casting saddles and bridges. Size 3 flasks shall be suitable for use in casting partial dentures. Size 4 flasks shall be suitable for use in casting full dentures.

3.3.1.13 Type II, crucible formers. The crucible formers shall be suitable for holding the wax patterns upright on a sprue pin and for making the proper conical depression in the investment material. The formers shall be fabricated of material having a minimum thickness of 0.040 inch (No. 18 B&S) gage.

3.3.1.13.1 Crucible former, small, size 1. The small crucible former shall be suitable for use with the inlay and saddle and bridge flask.

3.3.1.13.2 Crucible former, large, size 2. The large crucible former shall be suitable for use with the partial and full denture flasks.

3.3.1.14 Type III, crucibles. Crucibles shall be suitably shaped to fit crucible mount on the sling arm.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

3.3.1.14.1 Crucible, small, size 1. The small crucible shall be suitable for use in casting gold inlays and full crowns.

3.3.1.14.2 Crucible, large, size 2. The large crucible shall be suitable for use in casting partial and full denture appliances.

3.3.1.15 Type IV, asbestos strip, dental. A roll of 1/32 inch thick asbestos strip, 1-1/4 ± 1/16 inches wide and 100 ± 3 feet long, shall be furnished with each casting machine. The asbestos strip shall be suitable for use as a lining in the casting machine.

3.3.2 Replacement parts. Replacement parts, when specified, shall be furnished one each as described in 3.3.1.12 through 3.3.1.15.

3.4 Performance. The centrifugal casting machine shall be suitable for the preparation of various sized castings, when tested as specified in 4.4.3.

3.5 Identification marking. The machine shall have a corrosion-resistant metal nameplate permanently mounted on the base body. The nameplate shall be permanently and legibly stamped or embossed with the name or registered trademark of the manufacturer and the model number. In addition, all cradles, flasks, and crucible formers shall be permanently and legibly marked with the manufacturer's name or registered trademark and with the following coded size specified in table III.

Table III. Marking

Item	Size	Description	Required Marking	Marking Location
Flask	1	Inlay	1C	Outside, Surface Parallel To Centerline
	2	Saddle and Bridge	2C	
	3	Partial	3C	
	4	Full Denture	4C	

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

Table III. Marking (cont.)

Item	Size	Description	Required Marking	Marking Location
Cradle	1	Small	1C	Top or Bottom Flat Surface
	2	Medium	2C	
	3	Large	3C	
Crucible Former	1	Small	1C	Top or Bottom Flat Surface
	2	Large	2C	

3.6 Service data. The contractor shall enclose, with each unit, two copies each of one or more booklets which contain complete instructions for installation, operation, maintenance, and a list of component parts. Each part of the unit other than common fasteners, etc., shall be identified in the service data and parts list by the manufacturer's part number. In addition, the contractor shall mail 15 copies of the service data to the contracting officer for distribution.

3.7 Workmanship. Workmanship shall be first class throughout. The centrifugal casting machine shall be free from defects which detract from its appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection and acceptable quality levels (AQL) shall be as indicated in table IV.

Table IV. Inspection and Acceptable Quality Levels

Type defect	Inspection level	AQL (percent defective)
For visual examination		
Major	II	1.0
Minor	II	2.5
For dimensional examination	S-2	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table V except that the acceptance number shall be zero for all sample sizes.

Table V. Tests

Component	Characteristic	Reqt.	Method	Inspection level
Entire machine	Material identification	3.1	4.4.1	Not applicable
Flask	Flattening test	3.3.1.12	4.4.2	S-2
Entire machine	Performance test	3.4	4.4.3	S-2

4.3 Examination.

4.3.1 Visual examination. Casting machines shall be examined for defects including, but not limited to, those listed in table VI.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

Table VI. Classification of Defects

Categories	Defects
Major:	
101	Base housing, cover, weights, sling arm or straight arm have significant casting defects.
102	Burrs, nicks and feathers present on casting machine components.
103	Casting machine components can not be properly assembled.
104	Style and design not as specified.
105	Insufficient pressure on main coiled spring to provide quick start required for casting operation.
Minor:	
201	Enamel finish not as specified on base housing cover and counterbalance weights.
202	Identification marking missing or not as specified.
203	Plating of components shows signs of peeling or flaking.
204	Welds not ground or polished to blend into adjacent surfaces.

4.3.2 Dimensional examination. The casting machine shall be examined for defects in dimensions. Any dimension not within the tolerance specified in Drawing 20705 or in 3.3 shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical analysis of the metals. In the absence of certificates of quality for the raw material, metals specified in table I shall be analyzed in conformance with methods 111 or 112 of FED-STD-151. In case of dispute, method 111 shall be used for referee analysis.

4.4.2 Flattening test. Sample flask shall be flattened between two flat plates by a gradually applied load, normal to the axis of the tubing, until the distance between the inside faces of the plates is equal to three times the wall thickness of the tube.

4.4.3 Performance test. The centrifugal casting machine shall be suitable for casting gold inlays to full dentures, when done in accordance with standard dental prosthetic procedures. For inlays and small bridges, two, two and one-half, or three turns of the straight arm shall be sufficient. For partial and full dentures, four, four and one-half, or five turns shall be sufficient.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One complete casting machine or one each of replacements parts (one crucible, one flask, one crucible former or one roll of asbestos (100 ft.)), as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table VII shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table VII. Packaging Quantities

Item	Unit package	Intermediate package	Exterior container
Casting machine	1 unit	Not required	1 unit
Asbestos strip (size 1)	1 unit	Not required	48 units
Crucible (size 2)	1 unit	12 units	144 units
Crucible (size 1)	1 unit	12 units	144 units
Flask (size 2)	1 unit	12 units	144 units
Flask (size 4)	1 unit	Not required	48 units
Flask (size 3)	1 unit	Not required	48 units
Flask (size 1)	1 unit	12 units	144 units
Crucible former (size 2)	1 unit	12 units	144 units
Crucible former (size 1)	1 unit	12 units	144 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

5.1.3.1.1 Casting machine. Each unit shall be disassembled as far as practicable. Disassembled parts and component parts shall be wrapped, cushioned, or cartoned in such a manner as to prevent damage or marring of surfaces, and further overpackaged in a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Box shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to, the contents. Box shall be closed by taping the top and bottom flaps with 2-inch wide tape, conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.3.1.2 Asbestos and flask size 3 and 4. Unit package shall be a water-resistant envelope of appropriate size and design. Envelope shall be closed by heat sealing.

5.1.3.1.3 Crucible, flask size 1 and 2 and crucible former. Unit package shall be an envelope or chipboard box of appropriate size and design. Closure shall be adequate to prevent accidental opening.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2-inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the currentversion before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking of shipments shall be in accordance with FED-STD-123.

5.3.2 Military agencies.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement document should specify the following:

(a) Number and date of this specification.
(b) Complete casting machine or replacement part type and size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:53Z
Check the source to verify that this is the current version before use.

6.2 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Federal Government requirements.

6.3 Supercession data. This specification includes the requirements of MIL-C-002173A(DM), dated 2 March 1964.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

☆ U. S. GOVERNMENT PRINTING OFFICE - 1967: 303-221 Z492

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements .

SPECIFICATION

GG-C-001321(DSA-DM)

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 68 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:53Z
Check the source to verify that this is the current version before use.