GG-C-2736A
10 August 1987
SUPERSEDING
GG-C-002736(DM)
29 August 1986

FEDERAL SPECIFICATION

CATHETER, CARDIOVASCULAR
(Flow Directed, Thermodilution)

This specification is approved by the Commissioner, Federal Supply Service, General Services Administration for the use of all Federal Agencies.

1. SCOPE

1.1 Scope. This specification covers a sterile, cardiovascular flow directed, thermodilution four-lumen catheter suitable for use in determining hemodynamic pressures, cardiac output and thermal change, and conducting blood sampling and solution infusion. The catheter is used with a compatible cardiac output computer.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

FEDERAL SPECIFICATIONS

GG-I-526	- Instruments, Dental and Surgical; General Specification
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Set-up.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, DPSC-RST, 2800 S. 20th Street, Philadelphia, PA 19101, by using self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

Federal Standard:

Fed. Std. No. 123 - Marking for Shipment (Civil Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington, DC 20402).

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA).

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks, and the Index of Federal Specifications and Standards, from established distribution points in their agencies).

Military Specifications:

MIL-S-36586 - Sterilization, Test Strip Set, Bacterial Spore.

Military Standards:

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-147 - Palletized Unit Loads.
MIL-STD-794 - Parts and Equipment, Procedures for Packaging of.

(Copies of Military Specifications and Standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless a specific issue is identified, the issue in effect on date of invitation for bids or request for proposal shall apply.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE

Bureau of Biologics. Food and Drug Administration Code of Federal Regulation part 273. Minimum Requirements (as applicable).

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI/HIMA MD70.1 - American National Standard for Medical Materiel, Luer Taper Fittings, Performance.

(Applications for copies shall be addressed to Publication Sales, Dept. STD, American Institute of Physics, 335 East 54th St., NY, NY 10017).

UNITED STATES PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States (USP).

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 Material. All rubber and plastic components shall be of the best quality and suitable in all respects for the intended purpose. All rubber and plastic components shall be sterile, free from pyrogens, and safe (non-toxic) when tested as specified in section 4. Those rubber and plastic components coming in contact with the blood stream and solutions being administered into the blood stream shall contain not more than 5 parts per million (ppm) of barium, cadmium and lead when tested as specified in 4.5.1 or 4.5.1.1. The plastics shall meet the requirements of the Pharmacopeia of the United States (USP) for Class IV plastics.

3.1.1 Rubber components. Rubber components shall be made of natural rubber.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:54Z
Check the source to verify that this is the current version before use.

3.1.2 Plastic tubing. The plastic tubing shall be non-toxic, radiopaque polyvinyl chloride which is flexible and pliable, but sufficiently rigid for controlled manipulation during normal usage. The tubing shall be free from all liquid substances which oxidize, exude or sweat-out to the surface of the tubing.

3.1.3 Balloon. The rubber balloon shall be made from natural, non-oxidizing, low-sulfur, wax-free rubber.

3.1.4 Other materials. All other materials hereinafter specified shall be of the best quality and suitable in all respects for the purpose intended.

3.2 Style and design. Style and design shall be as hereinafter specified. The cardiovascular catheter shall be of four-lumen construction comprising one each for the distal lumen, the balloon, the thermistor and the proximal injectate post.

3.2.1 Design features.

Effective length	110 cm
Thermistor resistance, unadjusted, 37^0C	14,000 ohms $\pm$ 15%
Catheter size	7 Fr
Diameter of inflated balloon	12 to 13 mm
Diameter of deflated balloon	8 Fr
Balloon volume	1.5 cc
Distance of injectate port from tip	29 to 30 cm
Distance of thermistor from tip	3.5 to 4 cm
Guide wire diameter	0.025 $\pm$ .001
Injectate lumen volume, nominal	0.9 cc
Markings	At each 10 cm of length
Deflation time from 1.5 cc air	No more than 5 secs.

At 1.5 cc inflation , the tip of the catheter shall not protrude beyond the balloon.

The hubs for proximal injectate, distal lumen and the balloon inflation valve shall all be in compliance with ANSI/HIMA MD70.1 requirements.

The thermistor connector shall be compatible with the cardiac computer(s) designated by the catheter manufacturer on the individual container for the catheter.

Radiopacity shall be sufficient to be visible under conventional fluoroscopic in vivo illumination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

3.3 Pressure test. The balloon shall show no leakage when tested in accordance with 4.4.2.

3.4 Saline soak resistance. The balloon shall withstand 72 inflations to 1.5 cc air after exposures as shown in 4.4.3.

3.5 Thermistor circuit.

3.5.1 Response time. The thermistor 60% response time shall be no more than 1.5 secs when tested in accordance with 4.4.4.

3.5.2 Current leakage. There shall be no more than 10 microampers current leakage when tested as specified in 4.4.5.

3.5.3 Thermistor and lead wire resistance. The thermistor and lead wire resistance shall be 14,000 ± 15% ohms when tested as specified in 4.4.6.

3.6 Patency and leakage. The proximal lumen shall withstand ten 10 cc injections without leakage at the female Luer hub and without evidence of obstruction when tested as directed in 4.4.7.

3.7 Sterility, freedom from pyrogens and non-toxicity. The catheter shall be sterile, non-toxic and pryogen-free. Sterility testing shall be in accordance with 4.4.8 through 4.4.8.4. Freedom from pyrogens testing shall be in accordance with 4.4.9 and non-toxicity shall be tested in accordance with 4.4.9.

3.8 Certificates. Certificate of sterility, freedom from pyrogens, safety, and bacterial retention as specified below, shall be submitted. For destination inspection, certificates shall be submitted by the contractor directly to Directorate of Medical Materiel, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, ATTN: Quality Assurance Branch, and one copy of each shall be forwarded with each shipment to the consignee. For source inspection, certificates shall be furnished to the procuring activity via the government inspector.

3.9 Certificate of sterility. A certificate of sterility, listing all sterility lot or control numbers shall state that samples from each sterility lot or control have been tested for sterility as required in 3.4 and were found to be sterile. The certificate shall reference the method used in determining sterility.

3.10 Certificate of freedom from pyrogens. A certificate of freedom, listing all lot or control numbers, shall state that samples from each lot or control have been tested for pyrogens as specified in 3.4 and were found to comply with such requirements. The certificate shall reference the method used in determining freedom from pyrogens.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

3.11 Certificate of Safety. A certificate of safety, listing all lot or control numbers, shall state that samples from each lot or control have been tested for safety as specified in 3.4 and were found to comply with such requirements. The certificate shall reference the method used in determining safety.

3.12 Instructions. The container for the individual catheter shall provide detailed instructions and indicate French size, usable length and cardiac computer compatibility.

3.13 Workmanship. Workmanship shall be first class throughout. The catheter shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Responsibility for compliance. All items must meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of assuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling in quality conformance does not authorize submission of known defective material, either indicated or actual, or does it commit the Government to acceptance of defective material.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc. investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties), of the item.

4.1.3 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

4.1.4 Certificate of quality. Certificate of quality supplied by the manufacturer of the materials of components may be furnished in lieu of actual performance of composition testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the item identification, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination shall be one complete catheter.

TABLE I. Sampling for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	S-2	1.5
Minor	S-2	1.5

4.2.2 For tests. Sampling for tests shall be in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one complete catheter.

TABLE II. Sampling for tests

Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Metals	3.1	4.5, 4.5.1	S-1	*
Class IV Plastic	3.1.2	4.4.1	S-1	*
Pressure test	3.3	4.4.2	S-1	*
Saline soak resistance	3.4	4.4.3	S-1	*
Response time	3.5.1	4.4.4	S-1	*
Current leakage	3.5.2	4.4.5	S-1	*
Thermistor and lead wire resistance	3.5.3	4.4.6	S-2	*
Patency and leakage	3.6	4.4.7	S-2	*
Sterility	3.7	4.4.8 thru 4.4.8.4	See 4.4.8,etc.	*
Freedom from pyrogens	3.7	4.4.9	S-1	*
Non-toxicity	3.7	4.4.9	S-1	*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

* Acceptance number shall be zero for all sample sizes.

4.2.3. Sample unit usage. Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The sets shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with MIL-STD-105 and the classification of defects specified in Table III. Examination, however, shall not be restricted to the classified possible defects listed in Table III.

TABLE III. Classification of defects

Categories	Defects
Major	
101	Components missing.
102	Style and design not as specified (see 3.2.1).
103	Presence of foreign matter.
104	Lack of sterility resulting from break in package.
105	Catheter, not smooth, not free of burrs.
106	Catheter kinked.
Minor	
201	Packaging, packing and marking not as specified.
202	Instructions missing, incomplete or illegible.
203	Contractor's name or registered trademark, missing, incomplete, or illegible.

4.4 Tests.

4.4.1 Class IV plastic. The plastic shall be tested in accordance with the USP for Class IV plastics. Sampling shall be as shown in Table II.

4.4.2 Pressure test. The balloon shall be inflated with 4.5 cc of air, immersed in water and examined for leaks over a period of 10 secs.

4.4.3 Saline soak resistance. The balloon shall be placed in USP isotonic saline solution at $37^0C \pm 1.0^0C$. The catheter shall remain in the saline solution for 16 hours before beginning the inflation test. After 16 hours of soaking, the balloon shall be inflated with 1.5 cc of air for 2 to 3 secs every 20 minutes for a total of 24 inflations in 8 hours. The same 24 inflation pattern is repeated after 40 hours of soaking and again after 72 hours for a total of 72 inflations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

4.4.4 Response time. The thermistor shall exhibit a 60% response time of no more than 1.5 secs when the catheter is taken from room temperature (20^0 to 25^0C) to immersion in a 37.0^0C $\pm$ 1.0^0C USP isotonic saline bath.

4.4.5 Current leakage. There shall be no more than 10 microampers current leakage when the catheter is immersed in USP isotonic saline at room temperature for 72 hours and 117 volts AC current applied.

4.4.6 Thermistor and lead wire resistance. The thermistor and lead wire resistance shall be 14,040 $\pm$ 2,100 ohms. The resistance of the thermistor shall be measured with the thermistor bead immersed in 37.0^0C $\pm$ 0.1^0C USP isotonic saline solution. The measuring current shall not exceed 10 microampers.

4.4.7 Patency and leakage. A source of water at 45 psig shall be connected by a male Luer lock to the female Luer fitting (proximal injectate hub). Ten cc of water at that pressure shall be injected ten times to demonstrate patency and the absence of leakage at the Luer fittings. The Luer fittings shall be in accordance with ANSI/HIMA MD70.1.

4.4.8 Sterility test. The catheter shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates it from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation media for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30^0 to 35^0C for one half of the specimens and soybean-casin digest at 20^0 to 25^0C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Medium	Paper Spore Strip	Inoc. Prod. or Inoc. Simulated Prod.
Soybean-casein digest	S-3*	S-3*
Fluid thioglycollate or Alternative thioglycollate		S-3*

* Use a minimum of 20 specimens. Acceptance number is zero.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

4.4.8.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism
Steam	B. stearothermophilus
Dry heat	B. stearothermophilus
	B. subtilis
Ethylene oxide	B. subtilis
Gamma radiation	B. pumilus

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.

4.4.8.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.8 Sterility test and 4.4.8 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.8.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs 4.4.8 and 4.4.8.1.

4.4.8.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer and approval obtained before employment of the alternate method on a government contract.

4.4.8.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with USP. The inspection level shall be S-2. Acceptance number shall be zero.

4.4.9 Freedom from pyrogens and non-toxicity. Samples of the catheters shall be tested for freedom from pyrogens and non-toxicity in accordance with the USP. Sampling shall be shown in Table II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

4.5 Heavy metals extractable test. A 2-gram composite sample shall be obtained from the rubber and plastic components. The material shall be cut into small slivers, so that no thickness exceeds 3 mm, and eluted for 72 hours at 50^0C, 24 hours at 70^0C, 60 minutes at 121^0C in one menstruum of 0.9% sodium chloride. Each gram of the material shall be covered with 5 milliliters (ml) of the menstruum. The menstrua are collected, agitated after elution, and then submitted to spectographic atomic absorption, spectrophotometric or specific wet analysis for lead, barium, and cadmium. The metals are reported in parts per million. Thus 1.0 ppm in elute is equivalent to extracting 1.0 mg from 200 g of test material dispersed in 1,000 ml of elution menstruum.

4.5.1 Alternate heavy metals extractable test. Accurately weight a 5-gram sample for each of the metals (barium, cadmium or lead) into a 1-oz. Vycor crucible. Reduce to ash in a muffle furnace at 650^0C for eight hours. Let cool to room temperature and add 15 ml of 1:1 hydrochloric acid. Bring to a boil on a hot plate, allow to boil for 2 to 3 minutes, remove, let cool, and filter through Whatman #2 filter paper into a 50 ml volumetric flask. Rinse the residue on the filter with deionised water and dilute to 50 ml. Prepare standard solutions of 1, 2 and 5 ppm for each of the metals, namely barium, cadmium, and lead. Determine the absorbances of these solutions at appropriate wave lengths using a Hollow Cathode Ba, Cd, or Pb lamp for each of the respective metals and a 10X scale expansion on a spectrophotometer, such as the Perkin-Elmer Model 303, Atomic Absorption Spectrophotometer. Determine the absorbance of the sample solutions using the same conditions as with the standards.

Calculation:

$$\frac{A \times F \times \text{dilution}}{W} = \text{ppm barium, cadmium, or lead}$$

A = Absorbance

$$F = \text{Factor} = \frac{(\text{Concentration standard in ppm Ba, Cd, or Pb})^*}{\text{Absorbance}}$$

W = Sample weight in grams

* The three standards are average to give ppm/unit A.

4.5 Inspection of Packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each catheter shall be packaged in a see through rigid plastic tray having a rigid plastic lid so designed to avoid crushing of the catheter. The balloon shall be provided with a protector for protection from the atmosphere. The tray shall be further packaged in a sealed peel-open type bag capable of maintaining sterility of the contents unless the bag is damaged or opened. The bag shall be constructed of tyvek having a clear plastic backing.

5.1.1.2 Intermediate package. Five unit packages shall be packaged in a box conforming to PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C. The level C preservation for catheters shall conform to MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Four intermediate packages (20 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure and strappings shall be as specified in the applicable box specification.

5.2.2 Level B. Four intermediate packages (20 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. The level C packing for catheters shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in a exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length. Pallet shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Catheters packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets conforming to MIL-STD-147.

5.2.5.2 Level C. Catheters packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129. Marking shall also include the lot number and expiration date. In addition, each unit package shall also be marked to indicate the French size, usable length, cardiac computer compatibility and the legend:

"STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED," or similar statement.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

6.1 Intended use. The catheter shall be used to determine hemodynamic pressures and cardiac output.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) National Stock Number.

6.3 This specification covers the following National Stock Number:

National Stock No.	Item Identification
6515-01-030-2055	CATHETER, CARDIOVASCULAR, Flow-Directed, Four-Lumen, Thermodilution, With Balloon, Radiopaque, 14 Kohms Thermistor Resistance , 110 Cm Long, 7 Fr

6.4 Subject term (key word) listing.

Cardiovascular catheter
Catheter
Thermodilution catheter

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3913

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
GG-C-2736A	CATHETER, CARDIOVASCULAR (Flow Director, Thermodilution)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:54Z
Check the source to verify that this is the current version before use.