INCH-POUND

MIL-S-15103F(SH)
AMENDMENT 1
17 December 1992

MILITARY SPECIFICATION

SALINITY INDICATING EQUIPMENT

This amendment forms a part of MIL-S-15103F(SH), dated 16 March 1988 and is approved for use by the Naval Sea Systems Command, Department of the Navy and is available for use by all Departments and Agencies of the Department of Defense.

PAGES 3 and 4

2.1.1, under SPECIFICATIONS, MILITARY: Delete "MIL-W-16878 - Wire, Electrical, Insulated, General Specification For and MIL-W-16878/4 - Wire, Electrical, Polytetrafluoroethylene (PTFE) Insulated, 200°C, 600 Volts, Extruded Insulation." and substitute "MIL-C-24643 - Cable and Cord, Electrical, Low Smoke, For Shipboard Use, General Specification For and MIL-C-24643/2 - Cable, Electrical, 300 Volts, Types LSDCOP and LSTCOP."

PAGE 14

3.6, line 9: Delete "150 lb/in^2" and substitute "250 lb/in^2".
line 17: After "use of thread" insert "sealing compounds".

PAGE 15

3.6.1.5: Delete and substitute:

"3.6.1.5 Sensor cable. The sensor cable shall be made in accordance with MIL-C-24643 and MIL-C-24643/2 and shall be 6 feet long. Individual conductors shall have ring tongue terminations and shall be labelled with letter designations. The outer electrode (SBB-O) lead shall be red, the inner electrode (SB-O) lead shall be white and the temperature compensator (SB-1) lead shall be black."

3.6.2, line 2: Delete "150 lb/in^2" and substitute "250 lb/in^2".
line 4: Delete "225 lb/in^2" and substitute "375 lb/in^2".

PAGE 17

3.8.11, line 2: Delete "225 lb/in^2" and substitute "375 lb/in^2".
line 7: Delete "150 lb/in^2" and substitute "250 lb/in^2".

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T05:35Z
Check the source to verify that this is the current version before use.

4.5.15.1, line 3: Delete "225 lb/in^2" and substitute "375 lb/in^2".
line 4: Delete "150 lb/in^2" and substitute "250 lb/in^2".
line 7: Delete "225 lb/in^2" and substitute "375 lb/in^2".
line 12: After "shall be" insert "verified as specified in 3.8.11. Following the leakage test, the sensor constant shall be".

4.5.15.2.2, line 3: Delete "225 lb/in^2" and substitute "375 lb/in^2".

Preparing activity:
Navy - SH
(Project 6630-N543)

Source: https://assist.dla.mil -- Downloaded: 2026-10-19T05:35Z
Check the source to verify that this is the currenversion before use.