COMMERCIAL ITEM DESCRIPTION

FREQUENCY SELECTIVE LEVELMETER AND SPECTRUM DISPLAY (FSLSD)

The General Services Administration has authorized the use of the Commercial Item Description.

Scope. This Commercial Item Description (CID) describes a frequency selective levelmeter and frequency selective levelmeter with spectrum display. This CID is meant as a minimum requirement for the equipment, in either configuration, and only those manufacturers that meet or surpass the requirements herein are qualified per this CID. The frequency selective levelmeter and/or spectrum display shall be referenced as a FSLSD, the equipment, or the test equipment.

Part Numbering. The configuration will be referenced as follows:

A or B XXXXX
- A or B: Configuration A-frequency selective levelmeter; Configuration B-frequency selective levelmeter with spectrum display
- XXXXX: CID Number

Classification. Equipment covered by this CID may be commercially available equipment modified to the extent necessary to meet the following description. The equipment shall be type III, class 5, style F and color R in accordance with MIL-T-28800. This CID makes provisions for the bid sample testing in accordance with MIL-T-28800.

Applicable documents.

Specifications and standards. The following specifications and standards form a part of this CID to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

MILITARY

MIL-T-28800	Test Equipment for use with Electrical and Electronic Equipment, General Specification for.

(Copies of specifications, standards, publications, and other Government documents required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

Other publications. The following document forms a part of this CID to the extent specified herein. Unless otherwise specified, the issue of the documents which are DOD adopted shall be those listed in the issue of the DODISS specified in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

AMERICAN NATIONAL STANDARDS INSTITUTE

IEEE-587-1980 Guide for Surge Voltage in Low Voltage AC Power Circuits.

(Application for copies of ANSI publications should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018.)

Order of precedence. In the event of a conflict between the text of this CID and the references cited herein (except for associated detail specifications, specification sheets or MS standards) the text of this CID shall take precedence. Nothing in this CID, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

Salient characteristics. The equipment shall be capable of operation within the accuracies, limits, and specifications herein.

1. SAFETY. The equipment shall be UL (Underwriters Laboratories) listed and approved and/or shall comply with the type III safety requirements of MIL-T-28800.

a. Leakage current protection. Protection shall be provided from leakage current in excess of 3.5 milliamperes (ma) rms (AC and DC) from any accessible conductive parts of equipment (including control shafts with knobs broken or removed and recessed calibration or adjustment controls) to either pole of the power source for any position of the power switch(es).

b. Radio frequency interference (RFI) capacitors. Any capacitor used inRFI filters shall be rated for at least four (4) times the operating voltage.

c. Type III ventilation openings.

(1) Ventilation openings shall be designed to prevent accidental contact with live electrical circuits and mechanical moving parts.

(2) Top ventilation openings shall incorporate small mesh screeningor similar design to prevent the expulsion of hot electronic particles due to failure of electronic devices such as capacitors.

2. Parts, materials, and processes. Except for marking, the FSLSD shall comply with the type III parts, materials, and process requirements of MIL-T-28800. Nonstandard parts approval requests shall be in accordance with MIL-T-28800. Front panel marking shall be in accordance with MIL-T-28800.

a. Restricted material. The equipment shall comply with the restricted material requirements of MIL-T-28800, including the requirement for mercury and radioactive materials, and shall contain no combination of materials which cause deterioration of any material contained in the equipment due to the effects of outgassing.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:46Z
Check the source to verify that this is the current version before use.

3. Design and construction. Unless otherwise specified herein, the equipment shall comply with the type III design and construction requirements of MIL-T-28800. The equipment shall be of solid-state construction. Encapsulation and embedment (potting) of subassemblies may be used only for high voltage circuits where construction is susceptible to leakage or corona, indicators and associated controls shall be in close proximity to each other, all potentiometer controls shall have at least a 270 degree rotation range (4.71 radians).

a. Type of construction. Construction of the FSLSD shall be of a monolithic or a separate component design.

b. Modular design restrictions. The modular design concept for the FSLSD is acceptable with the following restrictions:

(1) The FSLSD shall consist of a single unit with one or more separate modules.

(2) Each modules shall contain a power circuit protection fuse.

(3) Each modules plug-in card shall be keyed and mated with the back plane receptacle so that modules may not be randomly placed into the mainframe.

(4) Modules shall not be removable by access through the front or rear of the mainframe with the unaided hand.

(5) The FSLSD shall come with one (1) blank space cover and one(1) extender board.

(6) A package part number shall be assigned to the FSLSD as a whole system. This system part number shall be marked on the mainframe identification plate and all plug-in(s).

(7) All module interface shall be done in the back of the mainframe.

c. Signal input and output connectors. Unless otherwise specified, all signal input and output connectors shall be Type BNC female.

d. Internal cooling. The equipment shall be capable of operating continuously for twenty four(24) hours at 55°C. If forced air is used, it shall be as specified in MIL-T-28800.

(1) Forced air cooling motor. DC brushless motors shall be utilized for the forced air cooling motor in the equipment.

(2) Air filter. Equipment utilizing a cooling scheme which exhausts air from the equipment through the cooling fan shall not require an externally accessible air filter.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:46Z
Check the source to verify that this is the current version before use.

(3) Thermal switch. A thermal switch shall be provided which will interrupt circuit power to the FSLSD when the internal temperature rises to the degree of possible damage to the electronic components. A minimum of 10°C hysteresis drop from this tripping temperature must occur before circuit power is again applied.

4. Electrical power sources. The equipment shall operate from nominal commercial, military, and shipboard power sources of 100, 110, 120, 200, 220 and 240 volts ac (±10%) at line frequencies of 50 Hz to 400 Hz.

a. Electrical power connectors. The equipment shall operate with a ground at the primary source on either side of the single phase input.

b. Power consumption. Maximum power consumption of all components combined shall not exceed 150 watts.

c. Fuses. Fuse(s) shall be provided, on the exterior rear panel for replacement without entering the enclosure. The fuse(s) shall automatically disconnect the equipment from the power source in the event of excessive current requirements.

d. Surge and transient voltage protection. The FSLSD shall contain surge and transient voltage protection internal of the enclosure. The transient voltage protection circuit shall be so designed as to prevent elec-tronic device damage within the system per type III solid state tolerance bands of MIL-T-28800, the surge voltage protection circuit shall conform to IEEE-587-1980, category A. The surge protection circuit shall drain excess power due to surges directly to the ground conductor for personnel protection. Normal operation of the FSLSD may be inhibited and false operational signals may be generated under input voltage transients greater than ±20 percent and frequency transient conditions greater than ±5 percent from within the nominal type III steady-state tolerance bands of MIL-T-28800, damage (other than fuse failure) to the FSLSD shall not occur during voltage transient conditions greater than ±20 percent and frequency transient conditions greater than ±5 percent within the maximum limits (30 percent voltage transients and 10 percent frequency transients) specified in MIL-T-28800. Normal operation of the FSLSD shall be automatically resumed when the transient has passed and any blown fuses are replaced. Blown fuses shall be externally accessible. Testing shall be in accordance with MIL-T-28800.

5. Dimensions and weight. The size and weight of individual components shall be consistent with current commercial capabilities, refer to paragraphs 5.a and 5.b of this CID.

a. Equipment dimensions The nominal dimensions per component for the FSLSD should be approximately 7 inches height x 17 inches wide x 18 inches deep.

b. Equipment weight. The weight including all accessories (excluding manual)per component, shall not exceed 50 lbs.

Source: https://assist.dla.mil -- Downloaded: 20261O-26T08:46Z
Check the source to verify that this is the current version before use.

6. Enclosure requirements. Unless otherwise specified herein, the equipment shall comply with the style F enclosure requirements of MIL-T-28800.

a. Rackmount. Rackmount capability shall be available as an option for the equipment specified herein.

b. Tilt bail handle. A tilt-bail handle shall be provided in lieu of the hinged metal handles in accordance with MIL-T-28800. Exception to this paragraph may be acceptable as this is not a prime consideration of the equipment.

c. Reusable pouch or container. An attached accessory pouch may be used for storage of all supplied accessories in lieu of front-cover storage.

7. Marking and identification.

a. Marking. The visible surface adjacent to panel facilities such as variable controls, indicators, jacks, connectors and fuse holder(s), shall be marked with a suitable word, phrase, or abbreviation thereof, indicating the purpose or use of the part.

(1) Panel marking. The front panel marking shall comply with the type III parts, materials and processes panel marking requirements of MIL-T-28800; however, marking identification processes other than the type III requirements specified in MIL-T-28800 may be used, provided the marking and identification processes are capable of passing the front panel marking test when tested in accordance with this CID.

b. Items to be marked (if applicable).

(1) Volume and weight (including accessories) shall be marked on three external surfaces of the enclosure.

(2) The type designation of the AN nomenclature of the FSLSD shall be stenciled on the exterior or main frame on at least three adjacent surfaces, on the transient case on two opposite sides. Marking shall be in block letters approximately 2 inches in height.

8. Environmental requirements.

a. Temperature requirements. Unless otherwise specified herein, the equipment shall meet its specified performance characteristics and other requirements when subjected to the class 5 temperature requirements specified by MIL-T-28800 and as listed below.

(1) Non-operating temperature. The equipment shall meet its performance and accuracy requirements in an non-operating environment of -40°C to +71°C.

(2) Operating temperature. The equipment shall meet its performance and accuracy requirements in an operating environment of 0°C to 50°C.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:46Z
Check the source to verify that this is the current version before use.

b. Humidity requirements. The equipment shall meet its required performance and accuracy requirements under humidity requirements for class 5 equipment as specified by MIL-T-28800.

c. Altitude requirements. The equipment shall meet its required performance and accuracy requirements under the altitude requirements for class 5 equipment as specified by MIL-T-28800.

d. Vibration requirements. The equipment shall meet the class 5 vibration requirements as specified by MIL-T-28800.

e. Mechanical shock requirements.

(1) Shock, functional. The equipment shall meet the class 5 shock, pulse requirement as specified by MIL-T-28800.

(2) Transit drop. The equipment shall meet transit drop requirement as specified by MIL-T-28800.

(3) Bench handling. The equipment shall meet the requirements as specified by MIL-T-28800.

(4) Shock, high impact. The equipment shall meet the requirements as specified by MIL-T-28800.

f. Electromagnetic interference (EMI). The equipment shall comply with MIL-T-28800.

g. Magnetic environment. The equipment shall comply with the magnetic environment requirements of MIL-T-28800.

9. Reliability. Reliability verification testing, shall be in accordance with the type III, class 5 environmental conditions of MIL-T-28800.

a. Mean time between failures (MTBF). The equipment mean time between failures (MTBF) shall be 2,000 hours. The lower test MTBF shall be 2,000 and the upper test shall be MTBF 4,000.

b. Pre-conditioning requirement. Each deliverable equipment shall be capable of being subjected to the environmental screening effects of burn in and thermal cycling, with thermal cycling being accomplished in the sample testing phase.

(1) Burn-in. Burn-in shall expose the equipment to extended periods of elevated temperatures, (40°C for a minimum of 100 hours). Each equipment which fails during burn-in shall be repaired and fully retested, until failure free. Failures which occur during the burn-in test shall be noted and reported but shall not count toward the establishment of MTBF. Prior to and following burn-in, a satisfactory operation test (SOT) shall be performed. A satisfactory operation check (SOC) shall be conducted at selected times during the burn-in test.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

(2) Thermal cycling. Thermal cycling shall expose the equipment to hot and cold cycles for class 5 equipment per MIL-T-28800. Each equipment which fails during thermal cycling shall be repaired and full cycling test shall be noted and reported but shall not count toward the cycling, a satisfactory operation check (SOC) shall be conducted during the last two upper and lower temperature limits.

10. Calibration and maintenance adjustments. The design of the system shall provide for readily accessible calibration adjustments and maintenance adjustments. These adjustments shall be provided by variable value components which are adjustable by the use of simple means. The calibration by substitution of selected components or parts is unacceptable unless specifically approved. The calibration adjustments, wherever possible, shall be accessible without removal of the instrument case or modules.

a. Maintainability. The equipment shall meet the type III maintainability requirements specified by MIL-T-28800.

11. Interchangeability. The equipment shall meet the type III interchangeability requirements specified by MIL-T-28800.

12. Digital interface. The equipment shall meet the digital interface requirements per MIL-T-28800.

13. Performance characteristics. Unless otherwise specified, all venier controls shall be in the calibrated position and all performance requirements shall apply after a 30 minute warm up period in a laboratory ambient operating temperature.

a. General requirements. This commercial item description describes two types of FSLSD systems.

(1) Configuration A XXXXX. Configuration A system shall consist of a frequency selective levelmeter, and shall meet all performance characteristics.

(2) Configuration B XXXXX. Configuration B system shall consist of a frequency selective levelmeter and spectrum display, and shall meet all performance characteristics.

b. Electrical and functional requirements. This section shall reflect the minimum characteristics and accuracies for the frequency selective levelmeter.

(1) Frequency.

(a) Range. Equipment shall be capable of responding to signals within a frequency range from 1 KHz to 25 MHz with an input impedance as follows:

75 ohms	1 KHz to 25 MHz
124 ohms	60 KHz to 11.5 MHz
135 ohms	10 KHz to 3.2 MHz
600 ohms	1 KHz to 650 MHz

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

(b) Display. Equipment shall have the capability of allowing frequencies to be displayed linearly on the horizontal axis of a cathode ray tube display.

(c) Tuning. Equipment shall be capable of digitally being set to a selected frequency. It shall have the option of being swept about this center frequency manually.

(d) Manual tuning. Equipment shall have the capability of being tuned manually by both an individual frequency decade control and a single knob tuning control.

(e) Decade frequency selector. The decade frequency selector shall consist of six individual decade selectors, each individually selectable from 0 to 9, and one selectable from 0 to 2 in the following ranges:

10 Hz	per step	0 to 9
100 Hz	per step	0 to 9
1 KHz	per step	0 to 9
10 KHz	per step	0 to 9
100 KHz	per step	0 to 9
1 MHz	per step	0 to 9
10 MHz	per step	0 to 2

(f) Resolution. The equipment shall be capable of resolving frequencies with an accuracy of 10 Hz.

(g) Accuracy. The equipment shall have an accuracy of ± 1 ppm ± 20 Hz from 20°C to 30°C, and ± 2 ppm from 0°C to 45°C.

(h) Stability. The equipment shall be stable to within 5 ppm per year.

(2) Level.

(a) Range. The equipment shall have a level measuring range from -129 dBm to +12 dBm over the specified frequency range referenced to 0 dBm on meter.

(b) Indicator. The equipment shall have an indicator showing what level measurement range has been selected and it shall be by means of a 3 digit LED display.

(c) Resolution. The equipment shall be capable of resolving levels in 1 dBm and 10 dBm steps and shall include a continuous vernier control with a calibrate mode.

(d) Modes. The equipment shall have a noise mode and a tone mode with the following ranges:

tone level range	-89 dBm to 10 dBm for meter zero
noise level range	-109 dBm to -10 dBm for meter zero

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:46Z
Check the source to verify that this is the current version before use.

(e) Accuracy. Accuracy of the equipment after internal calibration at each measurement frequency, shall be as follows:

75 ohms

Tone Mode 25°C ± 5°C

75 Hz Bandwidth

10 KHz to 18.6 MHz	
-69 dBm to +20 dBm	±0.15 dB
-89 dBm to -69 dBm	±0.20 dB
18.6 MHz to 25 MHz	
-69 dBm to +10 dBm	±0.25 dB
-89 dBm to -69 dBm	±0.30 dB
1 KHz to 10 KHz	
-69 dBm to +10 dBm	±0.30 dB
-89 dBm to -69 dBm	±0.70 dB

Tone Mode 0°C to 45°C ±0.10 dBm additional

Noise Mode

75 Hz Bandwidth

10 KHz to 100 KHz	
-99 dBm to -10 dBm	±0.40 dB
-109 dBm to -99 dBm	±0.90 dB
100 KHz to 18.6 MHz	
-99 dBm to -10 dBm	±0.25 dB
-109 dBm to -99 dBm	±0.35 dB
18.6 MHz to 25 MHz	
-99 dBm to -10 dBm	±0.30 dB
-109 dBm to -99 dBm	±0.50 dB

Noise Mode 0°C to 45°C ±0.10 dB additional

124, 135, 600 ohms

25°C to ± 5°C	±0.10 dB in addition to 75 ohms
0°C to ±45°C	±0.15 dB in addition to 75 ohms

(3) Input.

(a) Impedance. The equipment shall have four values of input impedances. They shall be 75 ohms in the Unbalanced state and 124, 135 and 600 ohms in the Balanced Std Unit state. All four values shall be push button selectable and shall be mounted on the front panel.

(b) Return Loss. Return loss for the equipment shall equal or better the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

Tone Mode 75 ohms

-79 dBm to +10 dBm	30 dB
-89 dBm to -79 dBm	2 dB

Noise Mode

-89 dBm to +10 dBm	30 dB
-99 dBm to -89 dBm	20 dB

Tone Mode 124, 135, and 600 ohms

-79 dBm to +10 dBm	30 dB except for 1st octave of frequency range 27 dB including 1st octave of frequency range
-89 dBm to -79 dBm	20 dB

Noise Mode 124, 135, and 600 ohms

-89 dBm to +10 dBm	30 dB except for 1st octave of frequency range 27 dB including 1st octave of frequency range
-99 dBm to -89 dBm	20 dB

(c) Overload Indicator. The equipment shall have an overload indicator light. It shall light up when high level inputs as as follows are present

For 1800 channels

Tone Mode	25 dB overdrive typical
Noise Mode	45 dB overdrive typical

(d) Signal Indicator. The equipment shall have a signal indicator light that is activated when a signal equivalent to -8 dB is present.

(4) Bandwidth. The equipment shall provide three bandwidths. Filters shall have the following selectivity:

Filter		
75 Hz	75 Hz ± 10%	350 Hz
1.74 KHz (equivalent noise bandwidth)	2.3 KHz ± 10%	4.6 KHz
3.1 KHz	3.1 KHz ± 10%	6.2 KHz

(5) Internal Distortion

(a) Harmonic. The harmonic distortion for the equipment shall be 75 dB below the fundamental for a 60 dB sensitivity increase from 10 KHz to 25 MHz.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

(b) Noise power ratio (NPR). The NPR for the equipment shall be a minimum beof 54 dB for a 40 dB sensitivity increase for an equivalent noise power, equivalent to 1800 channels. For an equivalent 600 channels, NPR shall be a minimum of 56 dB for a 40 dB sensitivity increase.

(6) Intermediate Frequencies. The equipment shall provide three intermediate frequencies with an IF and Image rejection of 80 dB. These tnree frequencies shall be:

IF 1	50.7 MHz
IF 2	10.7 MHz
IF 3	80 KHz

(7) Audio Output. The equipment shall provide audio monitor jacks for external test equipment.

(a) Speaker. The equipment shall provide an internal speaker connected to the audio output.

(b) Control. Both the equipment's speaker and audio output monitor audio level shall be controlled by a volume adjust located on the front panel.

(c) Functions. The equipment shall provide USB, LSB and AM modulation functions. These functions shall be either pushbutton selectable or knob selectable. Control or controls shall be located on the front panel.

(8) Residual Phase Jitter. With a 20 dB overdrive to the meter, residual phase jitter for the equipment shall be 0.5° nominal for signals 20 Hz to 300 Hz from the carrier.

(9) Tracking Signal Output. The equipment shall provide a tracking signal output from rear panel at 0 dBm ± 0.25 dB into 75 ohms impedance from 1 KHz to 25 MHz.

(10) Recorder Output. The equipment shall provide 0.8 volts D.C., typical into a 1 K ohm impedance for 0 dB on meter for recorder operation.

(11) Time Base Output. The equipment shall provide a time base output from the rear panel. Frequency shall be 10 MHz with a level of -10 dBm ±4 dB into a 50 ohm impedance.

(12) Meter. The equipment shall have a meter located on the front panel to measure level.

(a) Scale. The meter shall have both a normal and an expanded scale. Normal scale shall be from +2 dBm to -20 dBm and the expanded scale shall be from -1.0 dBm to +1.0 dBm. Minimum resolution on the expanded band shall be .05 dB.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

(b) Linearity. The meter shall have a linearity error of less than ±.02 dB in the expanded scale and less than ±0.1 dB from +2 dBm to -10 dBm and less than ±0.7 dB from -30 dBm to -20 dBm in the normal scale.

(13) Calibration. Equipment shall have an internal calibration circuit. It shall provide its own calibration signals.

(a) Level source. A precision level source shall be provided at each discrete measurement frequency available from the instrument.

(b) Control. The equipment shall provide a front panel pushbutton and level control that shall make all the necessary connections for calibration. No external cabling shall be allowed.

(14) Probe power. The equipment shall provide a front panel output of 15 vdc, 50 ma maximum for a high impedance probe.

c. Electrical and functional requirements. This section shall reflect the minimum characteristics and accuracies which shall be acceptable for the spectrum display.

(1) Compatability. The equipment shall be compatible with the frequency selective levelmeter described in para 3.2.

(2) Frequency.

(a) Range. The equipment shall be capable of displaying frequencies with 1 KHz to 25 MHz.

(b) Sweep width. The equipment shall provide three sweep width capable of the following:

Continuous or recurring sweep.

Single Sweep. Single sweep shall be activated by a pushbutton on the front panel.

Manual Sweep. Manual sweep shall be provided by means of a tuning control.

(c) Sweep width calibration. Sweep width calibration shall be provided by the equipment using a nominal 4 KHz per channel, in channels and in KHz/division, for six horizontal sweep scales as follows:

1 channel
12 channel
60 channel
600 channel
1800 channel
3600 channel

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:46Z
Check the source to verify that this is the current version before use.

The sixth sweep mode shall sweep the entire frequency range of the instrument. The above scales represent deflection from the furthest left to right graticules of the cathode ray tube.

(d) Resolution. Resolution for the equipment shall equal or better the following:

Signal level separation	0 dB	20 dB	50 dB
25 MHz	250 KHz	350 KHz	500 KHz
10 MHz	100 KHz	150 KHz	200 KHz
3 MHz	30 KHz	40 KHz	50 KHz
300 KHz	5 KHz	10 KHz	15 KHz
60 KHz	1 KHz	2 KHz	4 KHz
4 KHz	100 Hz	200 Hz	500 Hz

(3) Level.

(a) Range. The equipment shall be capable of displaying levels in the following ranges:

Total -109 dBm to +10 dBm (for zero level on display)
Display 60 dB logarithmic (linear dB scale)
+40 dB to -20 dB (referred to 0 dB reading on levelmeter)

(b) Accuracy. The equipment shall have the same accuracy as reflected for the level meter in paragraph 3.2.2.5.

(c) Frequency response. The frequency response for the equipment shall be within ± 0.1 dB for full frequency span of 1 KHz to 25 MHz.

(d) Display linearity. The equipment shall have a vertical linearity of ± 0.5 dB in the 1 dB/cm mode and ±2 dB in the 10 dB/cm mode.

(4) IF Bandwidths. The equipment shall have the following sweep widths:

at 3 dB		
25 MHz	10 MHz and 3 MHz	5 KHz
300 KHz		1.5 KHz
60 KHz		400 Hz
4 KHz		45.0 Hz

(5) Sweep Rate. The equipment shall be capable of providing the following sweep rates:

sweep widths	sweep rate
25 MHz	0.9 Hz (2500 MS period)
10 MHz	1.0 Hz (1000 MS period)
3 MHz	
300 KHz	2.2 Hz (450 MS period)
60 KHz	
4 KHz	0.5 Hz (2000 MS period)

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:46Z
Check the source to verify that this is the current version before use.

(6) Display Linearity. The equipment shall have a display linearity equal to or better than the following:

	% of full scale	
sweep widths	25°C ± 5°C	0°C to 45°C
25 MHz	± 3%	± 4%
10 MHz	± 5%	± 10%
3 MHz	± 5%	± 10%
300 KHz	± 2%	± 2%
60 KHz	± 2%	± 2%
4 KHz	± 5%	± 8%

(7) Cathode Ray Tube (CRT). The equipment shall have either a 6 cm x 10 cm or a 8 cm x 10 cm CRT. It shall display level vs frequency.

(a) CRT controls. The equipment shall have an intensity, focus, astigmatism, and scale illumination control. The scale illumination shall be provided by means of an adjustable internal light.

(b) Bezel. A bezel shall be included for an oscilloscope camera attachment.

(c) Range. CRT shall have either 60 dB (10 dB/cm) and 6 dB (1 dB/cm) or 80 dB (10 dB/cm and 8 dB (1 dB/cm) depending on the size of the CRT.

(8) High Level Alarm Output. The equipment shall provide a high level alarm output. It shall be by means of a set of dry contacts on the rear panel rated at 1 amp at 125 vac.

(9) High Level Alarm Trigger Level. The high level alarm shall rigger at the presence of any of the following:

sweep widths	4 KHz to 10 MHz
10 dB/div mode	+1.0 dB graticule ± 2.5 dB
1 dB/div mode	+0.1 dB graticule ± 1 dB

14. Calibration interval. After calibration, the equipment shall meet each performance requirement within the tolerances specified for a period of at least 9 months, based on an operating time of 1500 hours.

Preservation, packaging, labeling, and marking. The preservation, packaging, packing, labeling, and marking shall be as specified in the contract or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:46Z
Check the source to verify that this is the current version before use.

Certification. The contractor shall certify that the product offered meets the salient characteristics of this description and that the product conforms to the producer's own drawings, specifications, standards and quality assurance practices and is the same product offered for sale in the commercial marketplace. Small businesses that cannot offer products with established commercial market acceptability, may compete by submission of bid samples. (The Government reserves the rights to require proof of such conformance before first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Known acceptable products: The following product(s) are known to have previously met the requirements of this CID and need not resubmit bid samples or certification. All other manufacturers must meet the requirements of this CID and submit to bid testing and certification as required.

Manufacturer	Model	NSN
Hewlett Packard	3586B-003-004-H18	
Anritsu	ML 422B	

Custodian:
Air Force - 99

Preparing activity:
Air Force - 82

(Project 6625-F892)

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:46Z
Check the source to verify that this is the current version before use.