
AMSC N/A
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

USER ACTIVITIES: ARMY - MD, EN
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

FED. SUP CLASS
6640



PART NUMBER	DIMENSIONS (cm) INSIDE HEIGHT	WIDTH	ALL H	DEPTH	TEMP RANGE ±1°C	POWER WATTS
MS36234-1				30	30 - 60	90
MS36234-2				23	25 - 40	60
MS36234-3				34	25 - 45	120

CANCELLED
CANCELLED AFTER 23 OCT 1991
NO SUPERSEDING STANDARD

M

DES

...ordance with ASTM A569.

...MS36234-3 shall have a door ... accordance with ASTM D702 or D1547. ... shelves shall permit free air circulation. ...ically controlled. Incubator shall have a thermome- ... plug for 115V, 50/60Hz.

...trical components shall conform to the standards of Underwriters Laborato- ...ies, Incorporated.

USE: MS36234-1 and -2 are bacteriological incubators. MS36234-3 may be used for growing cultures or egg incubation.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

Ⓐ CANCELLATION

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust. MS 99		INCUBATOR, BACTERIA AND EGG INCUBATION	MS36234
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1426, MAY 91 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE CANCELLATION

Project Number 6640-1736

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:36Z
Check the source to verify that this is the current version before use.