A-A-54458
31 July 1991

COMMERCIAL ITEM DESCRIPTION

BURNISHER, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-B-001294 (DSA-DM) which is cancelled.

This Commercial Item Description covers burnishers suitable for use in burnishing plastic materials, waxes and dental restorative materials.

Salient Characteristics:

Classification

Type I - Single - ended

No. 27
No. GF-33

Type II - Double - ended

No. 15 or "J"
No. 2 - 30
No. 26 - 27

Material. The burnishers shall be fabricated from Class 4, Type 440A corrosion-resistant steel in accordance with ASTM Designation F899. As an alternate, Type II handles shall be brass plated.

Hardness. The operative ends of the burnisher shall have a Rockwell hardness of C52 to C56 when tested in accordance with ASTM Designation E18 or ASTM E92 and E140.

Style, design and dimensions. The burnishers shall be designed for use in burnishing plastic materials, waxes and dental restorative materials within the mouth during dental restorative procedures.

The type I burnisher shall be single-ended, one or two piece construction, consisting of an operative end and a handle. The operative end shall be permanently secured to the handle.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

The type II burnisher shall be double-ended, one or three-piece construction, consisting of two operative ends and a handle. The operative ends shall be permanently secured to the handle.

The No. 27 burnisher shall be a small round burnisher with a spherical-shaped working end.

The No. GF-33 burnisher shall be of the spratley design with the operative end tapering to a point at the tip and the sides tapering to a sharp edge.

The No. 15 or "J" burnisher shall have a conical pointed tip at one end and a blunt rounded tip at the other end.

The No. 2-30 burnisher shall have a beavertail shape at one end and an ovoid shape at the other end.

The No. 26-27 burnisher shall have a large ball shape at one end and a small ball shape at the other end.

The style, design and dimensions of the burnishers shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 20088, No. 24418 and No. 24151 for the applicable type.

The burnishers shall be packaged in a container designed to prevent damage during shipping or handling.

Finish. The operative ends shall have a mirror or bright finish. The handle shall have a mirror or bright finish except for the serrated or knurled portion.

Corrosion resistance. The burnisher shall show no evidence of corrosion or plating of copper when tested in accordance with ASTM Designation F1089.

Identification marking. The burnishers shall be permanently and legibly marked with the name or registered trademark of the contractor, the instrument number and the word "stainless". Instruments of foreign manufacture shall bear the name of the country of origin.

Workmanship. The burnishers shall be free from defects which detract from their appearance or impairs their serviceability.

Unit. Each (EA). One burnisher of the type and dimensions specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality for raw material, supplied by the manufacturer of the steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 at inspection level II, with an AQL of 2.5 (percent defective) except for dimensions which shall be inspection level S-1, AQL of 1.0 (percent defective). The unit of product for examination purposes shall be one burnisher.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 at an inspection level of S-1 with an AQL of 1.0 (percent defective). The unit of product for test purposes shall be one burnisher.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by it subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statues, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Preservation, packaging packing, labeling, and marking. Unless other wise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Numbers	Item Identification	Type
6520-00-510-8150	Burnisher, Dental, Round, Small, No. 27	I
6520-00-510-9050	Burnisher, Dental, No. 15 or "J"	II
6520-01-055-0018	Burnisher, Dental, Beavertail and Ovoid, No. 2-30	II
6520-01-055-0853	Burnisher, Dental, Spratley, No. GF-33	I
6520-01-055-5086	Burnisher, Dental, Round, No. 26-27	II

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

A-A-54458

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6520-2420
Location: ML #1 /BURNISH

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54458	2. DOCUMENT DATE (YYMMDD): 31 July 1991

3. DOCUMENT TITLE: BURNISHER DENTAL

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) / e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME: Defense Personnel Support Center	b. TELEPHONE (Include Area Code) (1) Commercial: (215)737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code): ATTN: DPSC-MSE, 2800 S. 20th Street, Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office, 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466, Telephone (703) 756-2340, AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.