A-A-54463
30 August 1991

COMMERCIAL ITEM DESCRIPTION

PAPER, ELECTROCARDIOGRAPH RECORDING

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-P-37571 which is cancelled

This Commercial Item Description covers a heat sensitive, water resistant electrocardiograph recording paper

Salient Characteristics:

The electrocardiograph recording paper shall be heat sensitive, water resistant, with a black undercoat, white top coat, and an overprinted grid pattern It shall be furnished as a roll, spooled with printed side in

The finished paper shall be 0 0031 to 0 0037 inches thick, shall weigh 20 to 28 pounds per 500 - 20 by 24 inch sheets, shall have a minimum tensile strength of 4 5 kg per 15 mm width in the machine direction, shall have a minimum tensile strength of 2 0 kg per 15 mm width in the cross direction, and shall have an unprinted brightness of 62 to 74 percent, based on a magnesium oxide coating equalling 100 percent.

The paper shall be spooled on a core having an internal diameter of 0 620 to 0 634 inches The outside diameter of the roll shall not exceed 3 032 inches

The roll ends shall be smooth, straight, and parallel The electrocardiograph paper shall be 63 mm (nominal) in width by not less than 148 feet in continuous length with a printed grid 50 mm (nominal) in width Bottom grid line shall be located 0 266 inches (nominal) from the bottom edge to center of the line

The paper grid pattern shall extend horizontally for the length of the paper and consist of horizontal and vertical cross hatch lines spaced 1 mm (nominal) apart The first and every fifth horizontal and vertical line shall be of greater width and density than the intervening finer lines Grid

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

lines shall be clearly printed in a dark gray or black color of such density that traces can readily be distinguished from the grid background and trace measurement easily accomplished by counting grid lines Vertical timing lines, nominally spaced 75 millimeters apart, shall be printed in the top margin in line with the major vertical grid lines

Workmanship The electrocardiograph recording paper shall be free from defects which detract from its appearance or impair its serviceability

Unit Package (PG). One package containing two rolls of electrocardiograph recording paper, as specified, constitutes one unit

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of S-2 The unit of product for examination purposes shall be one roll of

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

electrocardiograph recording paper.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one roll of electrocardiograph recording paper.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, the following:

The paper shall be tested for thickness in accordance with TAPPI Standard T411.

The paper shall be tested for basis weight in accordance with TAPPI Standard T410

The paper shall be tested for tensile strength in accordance with TAPPI Standard T404

A drop of water shall be applied to the sensitized surface of the paper at room temperature (approximately 25°C). A visual check shall be made for discoloration after the drop of water has evaporated.

The unprinted paper shall be tested for brightness in accordance with TAPPI Standard T452

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTES: 1 The electrocardiograph recording paper is intended to be suitable for use with NSN 6515-01-115-0025, Electrocardiograph, Automatic.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

2 The following National Stock Number is covered by this document

6515-00-564 3076

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5070

Location PAPERCID 17D

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54463	2. DOCUMENT DATE (YYMMDD): 30 August 1991

3. DOCUMENT TITLE: PAPER, ELECTROCARDIOGRAPH RECORDING

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Street, City, State, ZIP Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING activity

a. NAME: DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870	(2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.