COMMERCIAL ITEM DESCRIPTION

RONGEUR, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-R-1549A which is cancelled.

This commercial item description covers rongeurs suitable for use in surgical procedures.

Salient characteristics:

The rongeurs shall be of the following types:

Mead No. 1A
No. 4A

Materials. The rongeurs shall be fabricated of class 4, type 420 or 410x martensitic stainless steel in accordance with ASTM Designation F899 for composition only.

Hardness. The jaws and handles of the rongeurs shall have a Rockwell hardness of C40 to C55. The springs shall have a Rockwell hardness of C42 to C52. The hardness shall be in accordance with ASTM Designations E18 (Rockwell) or E92 (Vickers) and E140.

Design and dimensions. The rongeurs shall be suitable for performing an alveolectomy, breaking down the alveolar process and other surgical procedures, and shall be in accordance with Defense Personnel Support Center (DPSC) Drawing Nos. 20508 and 20509 for design and dimensions.

Construction.

Cutting edges. The cutting edges shall be properly sharpened to cut cleanly throughout their entire length, when cutting 0.015 inch cardboard.

Locks. The locks of the rongeurs shall be either lap joint (scissors) type or box lock type. The edges of the joints shall be properly beveled to prevent pinching of the tissue. Pin in lock area shall be flush with the surface of the joint

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

in both open and closed positions. The rongeurs shall be free of excessive stiffness and shall not have excessive play when the rongeur is in the open or closed position.

Finish. The rongeurs shall have a bright finish, except the handles shall have a satin finish, and the concave surfaces of the beak shall have a finely honed surface.

Corrosion resistance. Corrosion resistance tests shall be in accordance with ASTM Designation F1089. The copper sulfate and boil tests shall apply to the entire rongeur.

Identification marking. Each rongeur shall be marked with the applicable instrument number. Each rongeur shall be marked with the country of origin, the name or registered trademark of the manufacturer (and contractor, if other than the manufacturer) and the word "stainless". All markings shall be permanent and legible.

Workmanship. Each rongeur shall be free from burrs, cracks, scratches, sharp edges and pits which detract from its appearance or impair its serviceability.

Unit. Each (EA). One rongeur of the design number, specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

(determination by technical means of physical and chemical properties) of the item.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination, and test characteristics. The unit of product for inspection, examination, and test purposes shall be one rongeur of each design number. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5 (percent defective) and sampling for tests (when applicable) shall be conducted at inspection level S-1. Acceptance number for tests shall be zero (0) for all samples.

Tests. Tests shall be conducted to determine compliance with specification requirements..

Chemical analysis. In the absence of contractor certification for the raw material, a chemical analysis shall be conducted. The chemical analysis shall be performed by reagent solutions, combustion methods, nonemission methods, or by spectrochemical analysis. Any ASTM method shall be acceptable.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification
6520-00-531-3150	RONGEUR, Mead, No. 1A
6520-00-531-3250	RONGEUR, No. 4A

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2480
LOCATION: KD1/Rongeur

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54464	2. DOCUMENT DATE (YYMMDD): 8-2-91

3. DOCUMENT TITLE: RONGEUR, DENTAL

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) e. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial / (2) AUTOVON
Defense Personnel Support Center	(215)737-2870 / 444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSTE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.