A-A-54465
15 August 1991

COMMERCIAL ITEM DESCRIPTION

SYRINGE KIT, DENTAL IMPRESSION AND IRRIGATING

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-S-42009 which is cancelled.

This commercial item description covers a disposable syringe kit suitable for use with elastomeric impression materials and irrigating solutions.

Salient characteristics.

Material. The syringe and tip shall be fabricated from inert polypropylene which meets the USP requirements for Class III plastics. The plunger tip shall be fabricated from inert natural rubber or neoprene. The finger grip extenders shall be fabricated of rigid plastic. The plastic sheets shall be fabricated from clear polystyrene.

Style, design and dimensions. The syringe kit shall consist of 50 disposable syringes, 50 sheets and 2 finger grip extenders. The plunger tip shall be lubricated with pharmaceutical grade silicone lubricant.

The plunger shall form a seal between the barrel and plunger such that there shall be no leakage past the plunger in either direction. The plunger tip shall be of a color which clearly contrasts with the syringe barrel for maximum visibility. The plunger shall operate smoothly in the syringe without binding or hanging up.

The reusable finger grip extenders shall be provided for use in delivering viscous materials from the syringe. The finger grip extender shall fit on the syringe and shall be readily removable.

The style, design and dimensions of the syringe and finger extender shall be in accordance with (DPSC) Defense Personnel Support Center Drawing No. 24501.

Plastic sheets. Fifty plastic sheets shall be provided for use in loading the syringe. Each sheet shall be 0.005 ± 0.0015 inch thick and shall measure 3 1/8 ± 1/8 inches by 2 ± 3/16 inches.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Sterilization. The syringe shall be capable of undergoing steam autoclaving without warping, deformation or binding when tested as specified herein.

Finish. The syringe barrel and tip shall have a smooth finish.

Marking. Each syringe shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

Instructions. Each unit package shall be supplied with accurate and adequate instructions detailing the proper use of the syringe kit.

Workmanship. The syringe shall be clean and free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One syringe kit consisting of fifty syringes enclosed in a clear plastic bag, fifty plastic sheets neatly inserted in an envelope, and two finger grip extenders enclosed in an envelope or bag. Complete and detailed instructions for use shall be supplied with each unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and Inspection level of II except for dimensions which shall have an inspection level of S-1 with an AQL of 1.0 (percent defective). The unit of product for examination purposes shall be one syringe.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as specified herein. The unit of product for test purposes shall be one syringe kit. AQL shall be 1.0 (percent defective) with an Inspection level of S-2.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sterilization. The syringe shall be autoclaved at 250°F ± 5° and 15 ± 1 pounds per square inch for 30 minutes. After the syringe is removed from the autoclave, the syringe shall be examined.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the handling activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-173-2081	SYRINGE KIT, Dental Impression and Irrigating, Disposable

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MD

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

Project No: 6520-2369
Location: Word/Syringe

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54465	15 August 1991

3. DOCUMENT TITLE: SYRINGE KIT, DENTAL IMPRESSION AND IRRIGATING

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.