A-A-54474
23 August 1991

COMMERCIAL ITEM DESCRIPTION

MICROSCOPE, SURGICAL (Co-observation, Binocular)

The General Services Administration has authorized the use of this Commercial Item Description in preference to Military Specification MIL-M-42053 which is cancelled

This Commercial Item Description covers a surgical operating microscope with interchangeable objectives

Salient Characteristics

Shall be a fiber optic illuminated and motorized surgical operating microscope with interchangeable objectives, for alternate use in ENT, Ophthalmic and Neurosurgery procedures

Shall be designed to provide a total magnification ratio of at least 1 4, with a maximum magnification of at least 16X, and with a field of view that ranges from at least 13 millimeters (mm) to 50 mm

Shall be suitable for operation from 50/60 Hz AC supply lines. The applicable line voltage shall be as indicated in the contract or purchase order (see 'Ordering data' paragraph)

Shall as a minimum include the following

A microscope body with motorized zoom magnification changer, focus adjustment, and body positioning, integrated coaxial fiber optic illumination with a lightguide of appropriate length, and interchangeable 175 mm, 250 mm and 300 mm objectives

A fiber optic coaxial illumination system which includes an illuminator housing with built-in power supply, lamp on/off and intensity control, enough light output ports to transmit light to the main microscope illumination system, slit illuminator and homogeneous illuminator, and any necessary column mounting brackets

An inclinable binocular tube equipped with removable high-eyepieces with locking diopter positions, and a continuously adjustable eyepiece interpupillary distance of approximately 50 mm to 75 mm

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

A mobile floorstand that contains base, column and horizontal arm assemblies, and any necessary counterweights and stabilizer weights to maintain safe balance during use The height of the column shall be at least six (6) feet (1 83 m)

A motorized control system consisting of a foot control panel, power control unit(s) and all necessary connectors to control, at a minimum, magnification changer zoom, focus adjustment, and microscope x-y axis positioning with automatic recentering

A motorized or manual body tilt mechanism that does not restrict the focusing range.

A motorized focusing range of at least 50 mm

External slit and homogeneous illuminators which are angle positioned (oblique) and prefocused for the 175 mm working distance provided by the main microscope objective lens, and allow a means for repositioning and refocusing should misalignment or unfocusing occur. Each illuminator shall include a manually operated light blocking design.

An assistant's inclined binocular stereo microscope to assist primary surgeon in surgical procedures, with either zoom or minimum 3-step magnification changer, an overall magnification range of at least 3 6X to 8 9X, and high eyepoint eyepieces with locking diopter positions

A beam splitter for connecting and providing the assistant microscope coaxial observation with the primary field of view (in order to assist primary surgeon in surgical procedures) Adding the beam splitter shall result in the following.

The primary and assistant's fields of view being at the same intensities

The light intensity to the point of focus does not change, although the amount of light at both pairs of eyepieces may be reduced.

There is no induced prism (i e the result of the two images not being in the same place)

Protective cover for the microscope

Four (4) sterilizable sleeves for the inclinable binoculars, or, as an alternate, a drape or sterilizable handles.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Two (2) sets of any necessary cable clamps to keep motor control cables out of a sterile field

One spare plug-in module with bulb for the illumination system

Workmanship. The surgical operating microscope shall be free from defects which detract from its appearance or impair its serviceability

Unit. Each (EA) One microscope system, as specified, constitutes one unit

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted on the surgical operation microscope in accordance with commercial practice to determine compliance with specification requirements. The sampling for inspections and tests shall be conducted in accordance with commercial practice

Performance testing. Each surgical operating microscope shall be completely assembled in its operational configuration and tested to demonstrate its capability for proper operation. The testing shall include checking for proper operation of the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Electrical controls

The motor control system with magnification changer zoom, focus adjustment, x-y axis positioning

System magnification adjustment

Illumination for coaxial, slit and homogenous illuminators

Positioning and focusing of the slit and homogenous illuminators

Any evidence of malfunction or improper operation by any component of the operating microscope system shall be cause for rejection of the microscope system.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Applicable line voltage, intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

NOTE. The following National Stock Number is covered by this document.

NATIONAL STOCK NO	ITEM IDENTIFICATION
6515-01-181-1405	MICROSCOPE, SURGICAL, Co-observation Binocular, 120/230 Volt, 50-60 Hertz, AC

MILITARY INTERESTS.

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
PHS
FDA-MPQAS

Project Number 6515-4938

Location ENABLE/MICSURCD/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMENDED / CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54474	23 August 1991

3. DOCUMENT TITLE

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME	b. ORGANIZATION
	(1) Commercial / (2) AUTOVON

7. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215) 737-2870	444-2870

c. ADDRESS (Include Zip Code)
ATTN: DPSC-MSE
2800 S. 20th Street
Philadelphia, PA 19101

IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT
Defense Quality and Standardization Office
5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466
Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.