A-A-54476
30 August 1991

COMMERCIAL ITEM DESCRIPTION

WATER QUALITY ANALYSIS SET (MBT)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a quality analysis test kit for determining the Mercaptobenzothiazole (MBT) concentration of a water sample by using a hand-held titration methodology.

Salient Characteristics:

Shall be a water quality analysis test kit monitoring the Mercaptobenzothiazole (MBT) concentration of water supplies in the range of 50 to 500 PPM (mg/L).

Shall determine the MBT concentration of a water sample by titrating with a premeasured liquid volume of oxidizing titrant agent whose color change signals the end of titration. The oxidizing agent shall be contained in an ampoule type titration cell. The titration cell shall be secured in a holding device for use in a hand-held titrimetric analysis testing procedure.

Shall include the following:

A sample cup with capacity to hold 25 milliliters (ml) of water sample. The sample cup shall include a suitable reference mark to indicate the 25 ml liquid level.

Thrity (30) vacuum sealed glass ampoules containing potassium permanganate in an acidic medium. Each ampoule shall include a pre-scored tapered tip. The side of each ampoule shall include scale unit markings that represent 50,55, 60, 65, 70, 75, 80, 90, 100, 125, 150, 175, 200, 250, 350 and 500 ppm levels of MBT concentration.

Thirty (30) tubing assemblies. One end of the tubing assembly shall be for inserting and providing a snug fit over the tapered tip of the ampoule. The other end of the tubing assembly shall be used for insertion in a water sample

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

A holding device to accommodate an ampoule with an attached tubing assembly. After proper insertion in the holding device the distal end of the tubing assembly shall project through an opening in the holding device for insertion in the water sample. The holding device shall include provisions for snapping off the pre-scored tip of an inserted ampoule. The holding device shall include a control bar or lever for manual suctioning of small amounts of water sample through the tube assembly and into the ampoule (after snapping off the pre-scored tip).

An instruction sheet that includes step-by-step directions for assembling components, testing procedures and conversion of test results to PPM (mg/L).

Shall provide testing results in accordance with the following procedure:

An ampoule, tubing assembly and holding device are assembled in accordance with directions in the instruction sheet.

The pre-scored tip of the ampoule is snapped off as indicated in the instruction sheet.

Small amounts of water sample are drawn through the assembly tube and into the ampoule to mix with the titrant contents, in accordance with directions in the instruction sheet.

Additional samples of water are drawn and mixed with the titrant contents, in accordance with directions in the instruction sheet, until a permanent color change of the titrant takes place.

After a permanent change to the titrant, the ampoule titrant cell is held in an upright position and scale reading for the resultant liquid level is converted to a PPM (mg/Liter) MBT concentration level, in accordance with directions in the instruction sheet.

Workmanship. The components of the water quality analysis kit shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Set (SE) One set consisting of a water quality analysis set with contents, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted on the water quality analysis set in accordance with commercial practice. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

NOTE: The following National Stock Number is covered by this document.

NATIONAL STOCK NUMBER	ITEM DESCRIPTION
6630-01-327-0617	WATER QUALITY ANALYSIS SET

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS

Project Number: 6630-0518

Location: ENABLE/WATERCD2/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.