A-A-54479
30 August 1991

COMMERCIAL ITEM DESCRIPTION

HEELS, ORTHOPEDIC, CAST

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-H-37577 which is cancelled.

This Commercial Item Description covers an adult size walking heel for incorporation into an orthopedic cast of the lower leg.

Salient characteristics:

General. The orthopedic heels covered herein shall be of the following types:

TYPE	DESCRIPTION
I	Firm walking surface; triangle design tread.
II	Medium firm walking surface; fine serrated tread.
III	Firm walking surface; coarse serrated tread.
IV	Resilient walking surface; diamond design tread.

Material. Heel types I, II, and IV shall be molded from either natural or synthetic rubber compound that is suitable in all respects for the purpose intended. Heel type III shall be molded from ethylene propylene terpolymer that is suitable in all respects for the purpose intended.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Physical properties. The physical properties of the material shall be shown in table 1 below:

Table 1. Physical Properties.

Type	Material	Durometer Hardness (shore A)	Tensile strength (min. psi)	Elongation minimum
I	Rubber	65-75	1000	350%
II	Rubber	50-65	1700	400%
III	Plastic	65-75	1500	700%
IV	Rubber	55-65	750	320%

Style, design, and dimensions. The heels shall be in accordance with the following Defense Personnel Support Center (DPSC) Drawing Numbers:

TYPE	DPSC Drawing No.
I	23184
II	23181
III	23182
IV	23183

Heels shall be molded of one piece construction.

Color and marking. The color of the heels shall be the manufacturer's commercial color except black and other floor marking colors.

Identification marking and labeling. The heels shall be legibly marked with the contractor's name or trademark. Foreign manufactured heels shall bear the name of the country of origin.

Workmanship. The heels shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One heel, of the type specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified herein. Unit of product for examination purposes shall be one heel. Sampling shall be Inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one heel. Sampling shall be Inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more text characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the currentversion before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	TYPE
6515-00-106-8475	HEEL ORTHOPEDIC CAST, Rubber, Orthopedic Equipment Design, Adult Size.	I
6515-00-106-8476	HEEL ORTHOPEDIC CAST. Rubber, Zimmer Design, Adult Size.	II
6515-00-106-8477	HEEL ORTHOPEDIC CAST, Plastic, Stryker Design, Adult Size.	III
6515-00-106-8478	HEEL ORTHOPEDIC CAST, Rubber, Wright Design, Adult Size.	IV

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DH

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-HFQAS

Project No. 6515-5081

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7.

3 The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE. This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

1 DOCUMENT NUMBER	2 DOCUMENT DATE (YYMMDD)
A-A-54479	30 August 1991

3 DOCUMENT TITLE: HEELS, ORTHOPEDIC, CAST

4 NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME

b ORGANIZATION

c ADDRESS (*Include Zip Code*)

d TELEPHONE (*Include Area Code*)
(1) Commercial
(2) AUTOVON

7 DATE SUBMITTED (*YYMMDD*)

8 PREPARING ACTIVITY

a NAME
fense Personnel Support Center

b TELEPHONE (*Include Area Code*)
(1) Commercial (215) 737-2870
(2) AUTOVON 444-2870

c ADDRESS (*Include Zip Code*)
ATTN: DSE
2800 S. 20th Street
Philadelphia, PA 19145

IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT
Defense Quality and Standardization Office
5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466
Telephone (703) 756 2340 AUTOVON 289 2340

DD Form 1426, OCT 89 *Previous editions are obsolete*

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.