A-A-54484
30 August 1991

COMMERCIAL ITEM DESCRIPTION

HUMIDIFIER, INHALATION THERAPY APPARATUS

(Plastic, Disposable, Heat & Moisture Exchanger Type)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a patient heat and moisture exchanger for use as a humidifier during patient mechanical ventilation

Salient Characteristics:

The humidifier shall be an adult, plastic, disposable, artificial nose type apparatus and shall be suitable for use in inhalation therapy It shall be suitable for use as a heat and moisture exchanger It shall be designed for use in intensive care and anesthesia settings during patient mechanical ventilation

The humidifier shall be a corrugated plastic flex-tube, approximately 20 cm long, which shall recycle patient heat and moisture The patient side of the tube shall be standard 15 mm inside diameter and the equipment side shall be standard 22 mm inside diameter in accordance with ASTM F 1054

The humidifier shall be designed to allow the corrugated plastic tube to fit any type ventilator meeting International Standards Organization (ISO) 5356-1, Anaesthetic and Respiratory/ICU Departments

The humidifier shall be in accordance with the appropriate International Standards Organization (ISO) standard for heat and moisture exchangers (see Note 1 at end of this document).

The humidifier shall be internally impregnated with an antibacterial agent and individually wrapped in a heat sealed package The package shall contain directions for use. Packaging and labeling shall be that usually furnished by the

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

supplier as commercial practice

Workmanship The humidifiers shall be free from defects which detract from their appearance or impair their serviceability

Unit. Package (PG) One package containing twenty-five individually packaged humidifiers, as specified, constitutes one unit

Quality Assurance Provisions.

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one individual humidifier, as specified

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of II The unit of product for test purposes shall be one individual

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:34Z
Check the source to verify that this is the currentversion before use.

humidifier, as specified.

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include, but shall not be limited to, applicable tests in the standards cited in this document and other tests normally conducted for the supplier's commercial market.

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U.S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTES: 1 At time of preparation of this document, International Standards Organization (ISO) Standard 9360, Part 2, Anaesthetic and Respiratory Equipment - Heat and Moisture Exchangers for Use in Humidifying Respired Gases in Humans, was in draft stage Until this ISO standard is finalized, reference to it within this document is negated

2 The following National Stock Number is covered by this document

6515-01-275-0093

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA MPQAS

Project Number: 6515-5036

Location: HUM2CID 12B

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:34Z
Check the source to verify that this is the currentversion before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54484	2. DOCUMENT DATE (YYMMDD) 30 August 1991

3. DOCUMENT TITLE: HUMIDIFIER, INHALATION THERAPY APPARATUS (Plastic, Disposable, Heat & Moisture Exchanger Type)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Street, City, State, Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial FC 215-737-2870	(2) AUTOVON AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.