A-A-54487
9 September 1991

COMMERCIAL ITEM DESCRIPTION

NEEDLE, HYPODERMIC, MALLEABLE, SPINAL PUNCTURE (LEMMON)

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-N-36220A which is cancelled.

This commercial item description covers malleable spinal puncture needles of the Lemmon design, which are interchangeable with, and which fit the appropriate connections of the Lemmon spinal anesthesia set used in continuous spinal anesthesia.

Salient Characteristics:

General. The needles shall be of the following sizes:

Size 1 - 18 Gage, 3-1/2 Inch
Size 2 - 18 Gage, 4 Inch
Size 3 - 19 Gage, 4 Inch

The spinal puncture needle consists of a needle with integral hub and a stylet with an integral knob.

Material. The needle cannula and stylet wire shall be malleable and fabricated from nickel silver or, as an alternate, Class 3 austenitic stainless steel in accordance with ASTM Designation F899. The needle cannula and stylet wire shall not be plated. The needle hub and stylet knob shall be fabricated from nickel-plated brass.

Style, design and dimensions. The hypodermic needle shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 20482 and as specified herein.

Cannula. The cannula shall be securely attached to and concentric with the hub. The needle cannula shall be attached so the cleaning wire can be readily inserted and so as to not cause accumulation.

Hub. The hub shall be Luer lock in accordance with ANSI/HIMA MD70.1.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Stylet. The wire shall be securely attached to the knob. The wire tip bevel shall blend with the needle tip bevel and fit flush when inserted into the needle. The knob shall be Luer slip in accordance with ANSI/HIMA MD70.1 and have a slot or indexing device to assure alignment with the needle bevel.

Security of assembly. The needle cannula shall not pull out of the needle hub when subjected to a 5 pound force for one minute.

Corrosion resistance. The plated parts shall show no signs of corrosion when boil tested in accordance with ASTM Designation F1089 and salt spray tested in accordance with ASTM Designation B117. The austenitic stainless steel parts shall have no signs of corrosion when boil tested in accordance with ASTM Designation F1089.

Identification marking and labeling. The needle hub shall be marked with the gage and the manufacturer's trademark. The stylet knob and needle shall have matching serial numbers.

Workmanship. Workmanship shall be first class throughout. Needles shall be free from defects which detract from their appearance or may impair their serviceability. The cannula shall show a smooth finished outer surface free from pits and tool marks, burrs, scale or corrosion, and inner surface free from scale, deep pits, deep furrows, or die marks. The lumen of the cannula and the hub cavity shall be free from burrs, chips, dirt, slivers or other foreign matter. Needles shall be sharp, not blunt or bent, when examined under 7 power magnification.

Unit. Each (EA). One needle, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one needle. Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one needle. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the currentversion before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE: The following National Stocks, their respective Item Identification, and Sizes are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6515-00-302-5020	Needle, Hypodermic, Malleable, Lemmon, Spinal Puncture, 18 Gage, 3-1/2 Inch	1
6515-00-302-5040	Needle, Hypodermic, Malleable, Lemmon, Spinal Puncture, 18 Gage, 4 Inch	2
6515-00-302-5060	Needle, Hypodermic, Malleable, Lemmon, Spinal Puncture, 19 Gage, 4 Inch	3

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
HHS
FDA-MPQAS

PGC 05253

Project No. 6515-5050
Location: Brown Disk #4
A-A-54487.doc

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54487	2 DOCUMENT DATE (YYMMDD) 9 September 1991

3 DOCUMENT TITLE

NEEDLE, HYPODERMIC, MALLEABLE, SPINAL PUNCTURE (LEMMON)

4 NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed*)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial (215)737-2870 (2) AUTOVON 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church, VA 22041 3466 Telephone (703) 756 2340 AUTOVON 289-2340

DD Form 1426, OCT 89 Previous editions are obsolete 8290

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.