A-A-54488
16 September 1991

COMMERCIAL ITEM DESCRIPTION

VIBRATOR, DENTAL MOLDING
(4 inch platform)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a dental molding vibrator suitable for vibrating dental stones, plasters and investments into impression trays and small size casting rings.

Salient characteristics:

General. The vibrator shall consist essentially of a top vibrating plate, activated by an electro-mechanical vibrating system contained in the base assembly, mounted on four rubber feet. When the coil is energized, the armature and table top shall be pulled down sharply. When the coil is de-energized, the springs, armature, and table top shall be returned for the next cycle.

Material. The deck shall be fabricated of cast aluminum covered with a smooth rubber snap-on protective cover. The springs shall be fabricated of a suitable carbon steel.

Power. The vibrator shall operate on 115 volts, 60 hertz, single phase, with a power consumption of 35 to 45 watts.

Design and dimensions. The vibrator shall weigh 3 ± 1/2 lbs. and shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 24571 for design and dimensions.

Vibration. The vibrator shall vibrate at a constant rate of 3600 vibrations per minute when set at the high position.

Electrical and control components. The control components shall consist of a line cord, a 4 position control switch with off, low, medium and high positions, and an electrical coil assembly to drive the vibrating intensity required. All electrical components shall be properly protected against moisture. The power cable shall be a minimum of 5-1/2 feet long and equipped with 125 volts, 15 ampere, 2-pole, 3-wire grounding plug in accordance with American National Standards Institute (ANSI) C73.11.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Operation. The vibrator shall be for intermittent use. It shall be capable of operating without failure when operated with a 1 lb. load at intervals of 5 minutes on and 5 minutes off for 100 cycles when set at the high position.

Fire and casualty. The vibrator shall conform to the requirements of Underwriter's Laboratories (UL) 1262 or an equivalent nationally recognized independent laboratory.

Finish. Metal components shall be corrosion-resistant or suitably protected against corrosion.

Identification marking. Each vibrator shall be permanently and legibly marked with the name or registered trademark of the manufacturer, model designation, serial numbers, and electrical characteristics, such as voltage, frequency and amperage.

Service data. Two copies each of a technical manual which contains complete instructions for installation, operation and maintenance, and list of component parts shall be furnished with each unit.

Workmanship. The vibrator shall be free from defects which detract from its appearance or impair its serviceability or intended use.

Unit. Each (EA). One vibrator, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be on vibrator. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 (percent defective), and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0 (percent defective).

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-01-261-5970	VIBRATOR, DENTAL MOLDING, 4 inch platform, 3-speed 110 volt, 50/60 Hz, AC

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6520-2060
Location: Word/VIBRATOR PB Floppy #2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54488	2. DOCUMENT DATE (YYMMDD): 9-16-91

3. DOCUMENT TITLE: VIBRATOR, DENTAL MOLDING (4 inch platform)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) / e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial / (2) AUTOVON
Defense Personnel Support Center	(215)737-2870 / 444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.