A-A-54489
30 September 1991

COMMERCIAL ITEM DESCRIPTION

ILLUMINATOR SET, VAGINAL SPECULUM

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a vaginal speculum illuminator set for use in vaginal examinations.

Salient Characteristics:

General. Shall be an illumination system suitable for use with a disposable plastic speculum. System shall be a permanent total system, including a light handle and transformer on-off switch and a halogen lamp in the system and a spare lamp.

Light handle and transformer. The light handle and transformer shall consist of the light handle containing the halogen lamp, a power cord connected to a plug for the light handle and to a transformer. The light handle shall provide sufficient light intensity to illuminate the vagina for medical examination. The light handle shall be approximately 4.125 inches long by .6 inches wide by 0.435 inches high and have a weight not to exceed 1.65 oz. The power cord shall be at least 5 foot long and shall be firmly attached to the transformer. The transformer shall be approximately 1.71 inches long by 2.06 inches wide by 2.18 inches high and shall weigh no more than 9.2 oz. Transformer input shall be 110-130 volts, 50/60 Hz, .1 amp and output shall be 6 volts at .7 amps. The transformer shall be doubly insulated and contain a 110-130 volts resetting thermal circuit breaker.

Material. Material of the illuminator set shall be medical grade plastic or other material suitable for medical use.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

On-off switch. Shall provide on-off power capability at the light handle and shall connect between the light handle and the transformer cord plug. The switch shall have 6 volt, .7 amp capability.

Instructions. Instructions for use of the system shall be provided with each set.

Sterilization. The light handle shall withstand repeated autoclaving at 270° F, 27 psi for 6 minutes and ethylene oxide at 130° F, 8 psi, for 4 hrs. The switch and transformer shall withstand sterilization with ethylene oxide as shown above.

U.L. Listing. The system must be listed to U.L. Standard No. 544, comply with U.L. 1310 and CSA Certified to Standard No. C22.2 No. 125, as patient-care equipment (Risk Class II).

Dielectric strength. 1500 V.A.C., 60 Hz shall be applied for 1.0 seconds, between the primary and secondary of the transformer. The plug prongs shall be connected together during this test. The pins at the end of the output cord shall be similarly connected. There shall be no evidence of breakdown or leakage in excess of 500 UA.

Leakage current. Leakage current shall be measured from each transformer output terminal to ground. During the test, the primary shall be connected to a 120V, 60 Hz power source and the secondary loaded at 710 mA. Leakage current shall not exceed 10 microamperes as measured through the standard load described in U.L. Standard No. 544.

Cord pull test. The output cord shall not break, separate from the transformer or sustain any damage when a longitudinal force of 20 lbs is applied to the cord plug with the transformer secured.

Workmanship. The illuminator set shall be free from defects which detract from its appearance or impair its serviceability. Component parts shall be free from burrs, sharp edges, deformation, and foreign material.

Unit. Set (SE). One set consisting of an illumination set, as specified, constitutes one unit. Complete and detailed instructions for use shall be supplied with each set.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one illuminator set. Sampling shall be inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one illuminator set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-335-9916	Illuminator Set, Vaginal Speculum, 110-130 volt, 60 Hz.

The illuminator set shall be compatible with NSN 6515-01-164-3073 Speculum, Vaginal, Plastic, Medium Size Adjustable 5 inch blade.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-5152
Location. Nish disk #1

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7.

3 The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54489	2. DOCUMENT DATE (YYMMDD) 30 September 1991

3. DOCUMENT TITLE
ILLUMINATOR SET, VAGINAL SPECULUM

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: (215) 737-2870 (2) AUTOVON: 444-2670
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 96/290

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:34Z
Check the source to verify that this is the current version before use.