A-A-54491
30 September 1991

COMMERCIAL ITEM DESCRIPTION

CANNULA, TRACHEOTOMY, UNIVERSITY OF KANSAS DESIGN

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-C-36672(1) which is cancelled.

This Commercial Item Description covers metal tracheotomy cannulae of the University of Kansas design for use in performing cricothyroiodomy.

Salient characteristics:

General. The tracheotomy cannulae shall be of the University of Kansas design, used in performing cricothyroidotomy. All inner tubes and outer sleeves shall be interchangeable with each other within the given size. The tracheotomy cannulae shall be of the following sizes:

Size I - Adult

Size II - Pediatric

Material. All components of the tracheotomy cannulae shall be fabricated from types 302 or 304, class 3 stainless steel in accordance with ASTM F899. As an alternate, inner tube and outer sleeve shall be fabricated from types 316, 317, class 3 stainless steel in accordance with ASTM F899.

Hardness. The tracheotomy cannulae shall have a Rockwell Hardness of B75 minimum when tested in accordance with ASTM E18 or E92 whichever is applicable. Conversion of hardness values shall be in accordance with ASTM E140.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Style, design and dimensions. Each tracheotomy unit shall consist of two components. A curved cannula member which fits within an outer curved bivalve sleeve member. The inner member shall be concentric and secured by either welding or silver-soldering to a metal plate at the proximal end designed to permit ease of insertion into and removable from the outer member. The outer bivalve member shall also be permanently secured by either welding or silver-soldering to a metal retaining plate at the proximal end. Distal end of inner cannula shall be open and smoothly finished. The curved cannula (part 1, inner tube shown on drawing) shall be interchangeable with outer sleeve (part 2, on drawing) with other tracheotomy units being of the same size. (See Note 1) Tracheotomy unit shall be in strict accordance with Defense Personnel Support Center (DPSC) Drawing No. 24120.

Finish. All components shall be polished to a smooth bright finish. Definitions of acceptable surface characteristics shall be as described in ASTM F86.

Passivation. Tracheotomy units that are of welded construction shall be properly passivated after completion of all final finishing operations. Passivation shall be accomplished by the methods described in ASTM F86 or equivalent. Tracheotomy units that are of silver-solder construction shall be passivated before they are silver-soldered.

Resistance to corrosion. The tracheotomy units shall not show evidence of corrosion when tested in accordance with ASTM F1089. As an alternate, the following saline test may be used: Scrub the tracheotomy units with soap and warm water and rinse in hot water followed by immersion in ethyl alcohol (95%) and then dry. Tracheotomy units shall be completely immersed into Normal Saline Solution (USP) and while immersed, placed into an oven at a temperature of 37± 1 degrees centigrade for a period of seven (7) days, removed from the oven and examined for corrosion while still immersed.

Identification marking. The tracheotomy units shall be legibly marked with the contractor's name or registered trademark, and the word "stainless". Foreign manufactured tracheotomy units shall bear the name of the country of origin.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:34Z
Check the source to verify that this is the current version before use.

Workmanship. The tracheotomy units shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One cannula of the size specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one cannula. Sampling shall be Inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one cannula. Sampling shall be Inspection level S-2 with an AQL of 1.0.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:34Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Note 1: The tracheotomy unit shall be suitable for use with National Stock Numbers (NSN) 6515-00-935-7091 and NSN 6515-00-935-7092.

NOTE 2: The following National Stock Numbers are covered by this document:

NSN	Item Identification
6515-00-935-7089	CANNULA, TRACHEOTOMY, UNIVERSITY OF KANSAS DESIGN, Metal, Pediatric.
6515-00-935-7090	CANNULA, TRACHEOTOMY, UNIVERSITY OF KANSAS DESIGN, Metal, Adult.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DH

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-4981

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.

2. The submitter of this form must complete blocks 4, 5, 6, and 7.

3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
A-A-54491	30 September 1991

3. DOCUMENT TITLE: CANNULA, TRACHEOTOMY, UNIVERSITY OF KANSAS DESIGN

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (*Last, First, Middle Initial*)

b. ORGANIZATION

c. ADDRESS (*Include Zip Code*)

d. TELEPHONE (*Include Area Code*)
(1) Commercial
(2) AUTOVON

e. MAILING ADDRESS (*Street, City, State, Zip Code*)

7. DATE SUBMITTED (*YYMMDD*)

8. PREPARING ACTIVITY

a. NAME
Defense Personnel Support Center

b. TELEPHONE (*Include Area Code*)
(1) Commercial: (215) 737-2870
(2) AUTOVON: 444-2870

c. ADDRESS (*Include Zip Code*)
ATTN: ISE
2800 S. 20th Street
Philadelphia, PA 19145

IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:
Defense Quality and Standardization Office
5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466
Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:34Z
Check the source to verify that this is the current version before use.