A-A-54496
25 September 1991

COMMERCIAL ITEM DESCRIPTION

PACEMAKER, CARDIAC

Non-Invasive, Temporary, Desk Top Type

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a non-invasive, temporary, desk top type, 120/230 volt, 50/60 Hz cardiac pacemaker

Salient Characteristics:

The cardiac pacemaker shall be of a temporary, non-invasive, external type It shall have either a demand or demand/fixed rate pacing mode The pacing rate, in pulses per minute (PPM) shall be operator selectable in increments of 10 PPM or less (including continuous), over a range of at least 40 to 170 PPM. The pacing threshold shall be adjustable in either 5 milliamperes (mA) incremental steps or continuously variable settings The output current shall be operator selectable from 0 to 140 mA The pacing pulse width shall be at least 20 milliseconds (ms), and not greater than 40 ms The pacing rate accuracy shall be plus or minus 2 percent

The pacemaker shall be protected from defibrillator shocks up to 400 joules The pacemaker electrodes shall have a high impedance, be disposable, and have a large pregelled surface. The electrodes shall be available in both adult and pediatric sizes and shall be clearly labelled for front and back placement A package of 50 adult size (at least 60 square centimeter conductive surface area) electrodes and 50 pediatric size (at least 30-60 square centimeter conductive surface area) electrodes shall be provided with the pacemaker

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

The display shall indicate heart rate, pacing rate, output current, low battery status and alarm, ECG lead and pacer electrode fault, and a "battery charging" indicator display to allow the operator to determine if the unit is operating from AC line power or battery. The battery shall be a rechargeable, sealed lead acid type with an operating time of at least two hours

The weight of the pacemaker shall be no more than 9 kilograms (20 pounds) The pacemaker shall have an alarm that sounds when an electrode becomes detached It shall include a recorder with a frequency response of 0 5 to 30 Hz, plus or minus 3dB The pacemaker shall have a 1 mv calibration signal The recorder shall be a single channel hot stylus type using 10 mm grid with thermal paper The recorder shall operate with a 4 to 8 second delay

The pacemaker shall have the capability to recharge its battery while in operation via line power

The pacemaker shall be supplied with a case.

The pacemaker specifications shall meet all pertinent ANSI, AAMI, CSA, and ANSI/NFPA99 standards for patient and operator safety

The pacemaker shall include all necessary auxiliary power cables and user operating directions It shall be shielded against radio frequency interference (RFI) The pacemaker shall have an operating adjustable OPS detection sensitivity

The electrical characteristics of the pacemaker shall be 120/230 volt, 50/60 Hz

Workmanship The cardiac pacemaker shall be free from defects which detract from its appearance or impair its serviceability

Unit Each (EA) One cardiac pacemaker, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of S-2 The unit of product for examination purposes shall be one cardiac pacemaker

Sampling for tests Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1 0 (percent defective) and an inspection level of S-2 The unit of product for test purposes shall be one cardiac pacemaker.

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics Tests shall include, but shall not be limited to, the following:

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:34Z
Check the source to verify that this is the currentversion before use.

The cardiac pacemaker source patient risk current shall not exceed 10 uA, sink patient risk current shall not exceed 20 uA, ground open chassis source current shall not exceed 100 uA, and ground intact chassis source current shall not exceed 100 uA, when tested per ANSI/AAMI ES1, Safe Current Limits for Electromedical Apparatus

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Regulatory requirements

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U.S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order

NOTES: 1 The following National Stock Number is covered by this document

6515-01-321-8813

2 The pacemaker shall be compatible and interface with NSN: 6515 01-173-2053, Defibrillator, Monitor/Recorder

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5098

Location PACECID 17D

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54496	2. DOCUMENT DATE (YYMMDD): 25 September 1991

3. DOCUMENT TITLE: PACEMAKER, CARDIAC (Non-Invasive, Temporary, Desk Top Type)

4. NATURE OF CHANGE (Identify paragraph number and include proposed wording if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TYPE OF ORGANIZATION (Mark one): (1) Commercial (2) AUTOVON (If applicable) / e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME: DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code): (1) Commercial: 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code): 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:34Z
Check the source to verify that this is the current version before use.