A-A-54503
22 October 1991

COMMERCIAL ITEM DESCRIPTION

SCISSORS, COLLAR AND CROWN

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Federal Specification GG-S-1616A which is cancelled.

This commercial item description covers scissors suitable for use in cutting bands, preformed crowns, and matrices in dental operative procedures.

Salient Characteristics:

The scissors shall be as specified herein:

Curved, Beebee
Straight, Universal
Curved, Quimby

Materials. The scissors shall be fabricated of Class 4, Type 420B, 420X, 420C or 420F martensitic stainless steel. Except, the screw shall be fabricated of Class 4, Type 416 martensitic stainless steel or Class 3, Type 303 austenitic stainless steel. All materials shall be in accordance with ASTM Designation F899, for composition only.

Hardness. After proper stress relieving the scissors shall have a Rockwell hardness of C50 to C55 (with the exception of the screw), in accordance with ASTM Designation E18 or E92 and E140.

Design and dimensions. The scissors shall be suitable to contour preformed crowns, and to cut dental wires and matrices.

Defense Personnel Support Center (DPSC) Drawing No. 20178 shall apply for the Beebee design which shall have a curved blade and the universal design which shall have a sturdy, straight blade.

Defense Personnel Support Center (DPSC) Drawing No. 23201 shall apply for the curved Quimby design with doubled pointed ends.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Construction. The scissors shall be of three piece construction consisting of two scissor sections locked together by a screw, which is permanently set in a fixed position. The screw lock and scissor sections shall be properly assembled to provide a cutting portion which shall be free of both excessive stiffness and looseness. The inside of the cutting portion of each scissors shall be ground to properly mate with the opposing scissor during cutting procedures. The cutting edges of the blades shall be properly sharpened.

Finish. The scissors shall have a bright finish, except for the ground inside portion of the blades which shall have a crocus finish.

Cutting performance. The Beebee and Quimby design scissors shall cut a continuous length, along the blade to the tip. The Universal scissors shall cut stainless steel wire as specified. In addition, the Quimby design shall also properly cut a copper band as specified.

Corrosion resistance. All scissors shall meet the boil and copper sulfate tests in accordance with ASTM Designation F1089.

Identification markings. The scissors shall be marked with the word "stainless", manufacturer, and the country of origin. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless shall be acceptable, provided prior approval is received from the cognizant contracting officer. All markings shall be permanent and legible.

Individual package. Each scissors shall be wrapped in neutral tissue or other suitable anticorrosive material. Each scissors shall be packaged, in a sturdy close fitting box or in a rigid plastic container having a suitable closure, to prevent opening during shipping and handling.

Workmanship. Each pair of scissors shall be free from pits, cracks, burrs, coarse grind marks, nicks, feathers or any other defects which detract from their appearance or impairs their serviceability.

Unit. Each (EA). One pair of scissors of each design specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection, examination, and test characteristics. The unit of product for inspection, examination, and test purposes shall be one scissor of each type. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5 (percent defective) and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0 (percent defective), except for material and cutting performance acceptance number shall be zero for all sample sizes.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Chemical analysis. In the absence of contractor certification for the raw material, a chemical analysis shall be conducted. The chemical analysis shall be performed by reagent solutions, combustion methods, nonemission methods, or by spectrochemical analysis. Any ASTM method shall be acceptable.

Cutting performance. Test materials shall be a strip of austenitic stainless steel 0.002 ± 0.0005 inch thick, with a Rockwell hardness of C35 to C40, and a surgical stainless steel wire, 0.020 inch nominal. Scissors to be tested shall be thoroughly cleaned prior to testing. Three separate, consecutive tests shall be performed on each material for the respective scissors. Each test shall consist of making a 1/2 inch (minimum length) cut for the Beebee and Quimby design scissors, and a cut through the stainless steel wire for the Straight, Universal scissors. The cuts shall be made using the scissors in the normal manner, with the fingers in the finger rings of the scissors, except that no lateral pressure shall be used. In addition, the Quimby design scissors shall be provided with blades of a width, thickness, and curvature that permit precise gingival festooning on an annealed No. 1 copper band, 0.189 ± 0.005 inch inside diameter and 0.005 ± 0.0015 inch wall thickness.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-00-118-6250	SCISSORS, COLLAR and CROWN, Curved, Quimby, 4-3/4 inch
6520-00-554-8140	SCISSORS, COLLAR and CROWN, Curved, Beebee, 4-1/4 inch
6520-00-554-8150	SCISSORS, COLLAR and CROWN, Straight, Universal, 4-1/2 inch

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2419
LOCATION: KD#2/SCISSORS

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54503	2. DOCUMENT DATE (YYMMDD) 22 October 1991

3. DOCUMENT TITLE
SCISSORS, COLLAR AND CROWN

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — *89/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:34Z
Check the source to verify that this is the current version before use.