A-A-54509
1 November 1991

COMMERCIAL ITEM DESCRIPTION

TESTER, FLASH POINT, CLOSED-TYPE (120 Volt, 50/60 Hz, AC)

The General Services Administration has authorized the use of this Commercial Item Description in preference to Military Specification MIL-C-36385C which is cancelled.

This Commercial Item Description covers a Pensky-Martens closed-type flash point tester with an electric resistance type heater and a stirrer motor.

Salient Characteristics:

Shall be an electrically operated Pensky-Martens closed-type tester for determining flash points in accordance with the methodology in ASTM D93 (Standard Test Methods for Flash Point by Pensky-Martens Closed Tester).

Shall be suitable for operation from 120 Volt, 50/60 Hz AC supply lines.

Shall be capable of testing for the flash points of oils and suspensions of solids that are described in ASTM D93, Method A and Method B.

Shall consist of the following:

A test cup, cover and stove assembly with a stirring device.

A metal base housing.

An electric resistance type heater assembly.

A controller assembly for the heater.

An electric motor assembly with suitable coupling for driving the stirring device.

A flexible shaft for alternate manual rotation (hand stirring) of the stirring device.

A support stand with a mounting rod for the electric motor and test cup holder, if they require attachment to a mounting rod for proper operation.

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

An ASTM 9F Pensky-Martens low range thermometer conforming to the requirements of ASTM E1 (Standard Specifications for ASTM Thermometers).

An ASTM 10F Pensky-Martens high range thermometer conforming to the requirements of ASTM E1.

An electric power cord with plug.

Shall be supplied with a test cup, cover and stove assembly, and a stirring device conforming to the requirements in Annex A1 of ASTM D93. The assembly shall be top-mounted on the base housing.

Shall be supplied with a rigid metal base housing that is suitably protected from corrosion. The housing shall include suitable cutouts for ventilation and mounting of electric controls.

Shall be supplied with an electric resistance type heater assembly that is rated for 750 watts consumption and is designed to conform to the requirements in Annex A1 of ASTM D93. The heater assembly shall include a controller with a reference dial to vary the input heater current in a range that includes 0 to 750 watts. The reference dial shall be mounted on the front of the metal base housing.

Shall be supplied with an electric motor assembly that includes suitable coupling to drive the stirring device at the rate specified in ASTM D93 for Method A and Method B.

Shall be supplied with two separate on-off switches for controlling the electric input power to the electric heater and motor assemblies. An integral control with "off" position shall be acceptable for use with either assembly as an alternate to an on-off switch.

Shall be supplied with minimum five (5) foot exposed length three-wire flexible electric power cord that terminates with a non-detachable three-blade male grounding plug. The male grounding plug shall be suitable for use with receptacles shown in ANSI Standard C73.11 or shown in Configuration No. 5-15 of NEMA Standard WD6-88.

Shall be supplied with service data that includes a parts listing and instructions for the propper maintenance of the flash point tester.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Workmanship. The flash point tester shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One tester flash point, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted in accordance with commercial practice. Sampling for the inspections and testing shall be in accordance with commercial practice. In addition, each flash point tester shall be tested to demonstrate its capability for proper electrical and mechanical operation.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the extent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposal shall apply.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

C73.11 Plugs and Receptacles - General Purpose 125 Volts, 15 Amperes, 2 Pole, 3 Wire Grounding Type, Dimensions of.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018).

ASTM

ASTM D93 Standard Test Methods for Flash Point by Pensky-Martens Closed Tester

ASTM E1 Standard Specification for ASTM Thermometers.

NATIONAL ELECTRICAL MANUFACTURERS ASSOCIATION (NEMA).

WD 6-88 Wiring Devices - Dimensional Requirements (Revision and Redesignation of ANSI C73-73).

(Application for copies should be addressed to the National Electrical Manufacturers Association, 2101 L Street, NW Suite 300, Washington, DC 20037).

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

NOTE: The following National Stock Number is covered by this document.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6630-00-530-0987	TESTER, FLASH POINT, CLOSED TYPE (120 VOLT, 50/60 HZ, AC)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DLA-DM

Project Number: 6630-0520

Location: ENABLE/TESTERCD/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-54509	2. DOCUMENT DATE (YYMMDD) 1 November 1991

3. DOCUMENT TITLE
TESTER, FLASH POINT, CLOSED-TYPE (120 Volt, 50/60 Hz, AC)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center	b. TELEPHONE (Include Area Code) (1) Commercial: (215)737-2870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSTE 2800 S. 20th Street Philadelphia, PA 19101	d. IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:35Z
Check the source to verify that this is the current version before use.