GG-B-426b
AUGUST 30, 1960
SUPERSEDING
Fed. Spec. GG-B-426a
April 2, 1953

FEDERAL SPECIFICATION

BLOCK, BITE

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 **Scope.**—This specification covers bite blocks suitable for facilitating the passage of bronchoscopes.

1.2 **Classification.**

1.2.1 *Sizes.* — Bite blocks shall be of the following sizes, as specified (see 6.1):

Size 1. — Adult.
Size 2. — Adolescent.
Size 3. — Child.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

PPP-B-566 — Boxes, Folding, Paperboard.
PPP-B-585 — Boxes, Wood, Wirebound.
PPP-B-601 — Boxes, Wood, Cleated-Plywood.
PPP-B-621 — Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 — Box, Fiberboard.
PPP-B-676 — Boxes, Set-Up, Paperboard.

Federal Standards:

Fed. Std. No. 102 — Preservation, Packaging, and Packing Levels.
Fed. Std. No. 123 — Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

Military Specifications:

MIL-L-10547 — Liners, Case, Waterproof.

Military Standards:

MIL-STD-105 — Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 — Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions, should be obtained from the procuring activity or as directed by the contracting officer.)


570307/60/5


FSC 6515


Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

NOTE

1. UNLESS OTHERWISE SPECIFIED TOLERANCES = ± 1/32
2. ALL EDGES SHALL BE SMOOTHLY ROUNDED.







SIZE	A	B	C	D	E	F +1/16 −1/16	G
1	1 1/4	27/32	23/32	1 29/32	1 25/32	1 1/8	17/32
2	31/32	27/32	23/32	1 25/32	1 5/8	15/16	17/32
3	13/16	3/4	5/8	1 17/32	1 3/8	7/8	15/32

FIGURE 1.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:50Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 **Material.**—The bite block shall be fabricated of hard rubber or nylon. Material shall be such as to be nonirritating to the tissues of the mouth of the patient.

3.2 **Style, design, and dimensions.** — The bite block shall be of the Jackson design, wedge-shaped, and shall be suitable for facilitating the passage of bronchoscopes. Style, design, and dimensions shall be in accordance with figure 1. External surfaces of the sides of the block shall be corrugated as shown in figure 1. The block shall have a finger hole located at the posterior end for holding the block in position.

3.3 **Finish.**—The block shall be free from sharp corners or edges, and shall be polished to a high gloss.

3.4 **Workmanship.**—Workmanship shall be first class throughout. Bite blocks shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 **Supplier responsibilities for inspection.** —The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. Inspection records of the examinations and tests shall be kept complete and available to the Government as specified in the contract or order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 *Certificates of quality.* — When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 **Sampling.**—Sampling for examination shall be conducted in accordance with Military Standard MIL–STD–105:

	Inspection level	AQL (percent defective)
For visual examination:		
Major defects (A and B)	II	1.0
Minor defects	II	2.5
For dimensional examination	L2	2.5

4.3 **Examination.**—The bite blocks shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 *Classification of defects.* — Examination shall be conducted in accordance with the following classification of defects:

*Categories and defects**

Major A: Unit of product shall be one block bite.

101. Block not free of sharp edges and corners.
102. Block not free of cracks.
103. Block not free of deformed or chipped surfaces.

Major B: Unit of product shall be one intermediate package and contents.

151. Marking on package not correct (particularly stock number and nomenclature).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

Minor: Unit of product shall be one block bite.

201. Block not polished to a high gloss.
202. Finish not free of excessively scratched or abraded surfaces.
203. Label or printing on unit package not legible and free of defacing marks, scratches, and tears.

* Examination is not restricted to the classified possible defects listed above.

4.3.2 *Dimensional examination.* — The bite blocks shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests.—Tests shall be conducted, as necessary, to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

For civil agency procurement the levels of preservation, packaging and packing used are defined in Federal Standard No. 102.

5.1 Packaging.

5.1.1 *Unit of issue.* — One block, as specified, constitutes one unit.

5.1.2 *Packaging quantities.* — The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container as applicable, for the required level of protection specified in the procurement document.

TABLE I.—*Packaging quantities*

Unit package	Intermediate package	Exterior container
1 unit ----------	12 units	288 units

5.1.2.1 *Packing variation permitted.* —If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.*

5.1.3.1 *Unit package.* — Each unit shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566 or PPP–B–676. Each carton shall be adequately secured to prevent accidental opening.

5.1.3.2 *Intermediate package.* — Intermediate package shall be a carton of appropriate size and design, or a complete wrapping of kraft paper of suitable weight. Carton or wrapping shall be adequately secured to prevent accidental opening.

5.1.4 *Level C.* — Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 Packing.

5.2.1 *Level A.*

5.2.1.1 *Exterior container.* — Exterior container shall be designed for type 2 load, and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 3 use; Federal Specification PPP–B–601, table III, using type III, class 1 plywood; or Federal Specification PPP–B–621, table III.

5.2.1.2 *Case liner.* — Each level A box shall be lined with a waterproof case liner conforming to Military Specification MIL–L–10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 *Strapping.* — Strapping, when required, shall be in accordance with applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

Note.—Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 *Level B.*

5.2.2.1 *Exterior container.* — Exterior container shall be constructed in accordance with Federal Specification PPP–B–585, style 3, for class 2 use and type 2 load; Federal Specification PPP–B–621, table I; or Federal Specification PPP–B–636, type I, class 1, fabricated from not less than 275-pound test fiberboard.

5.2.3 *Level C.* — The subject commodity shall be packed in substantial commercial containers of the type, size, and kind, commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 **Marking—nonmilitary.**

5.3.1 *Interior packages and exterior containers.* — In addition to any special marking specified in the contract or order, marking shall be in accordance with Federal Standard No. 123.

5.4 **Marking—military.**

5.4.1 *Interior packages.* — Each unit package and intermediate package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

5.4.2 *Exterior container.* — Exterior container shall be marked in accordance with Military Standard MIL–STD–129.

6. NOTES

6.1 **Ordering data.** — Purchasers should exercise any desired options offered herein, and procurement documents should specify the following:

(*a*) Title, number, and date of this specification.

(*b*) Sizes required (see 1.2).

(*c*) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Federal Government requirements.

6.3 **Transportation description.**—Transportation description applicable to these items is:

Rubber goods, not otherwise indexed by name.

Carload weight 20,000 pounds subject to Rule 34, Uniform Freight Classification.

Truckload volume weight 20,000 pounds.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded, by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

CUSTODIANS:
Army—Army Medical Service
Navy—Bureau of Medicine and Surgery
Air Force.

Copies of this specification may be purchased for 5 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity *(as indicated on reverse hereof)*.

SPECIFICATION

ORGANIZATION *(Of submitter)* | CITY AND STATE

CONTRACT NO. | QUANTITY OF ITEMS PROCURED | DOLLAR AMOUNT $

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES.

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO IF "YES", IN WHAT WAY?

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity)* | DATE

DD FORM 1 APR 63 1426 REPLACES NAVSHIPS FORM 4863, WHICH IS OBSOLETE C-8279

Source: https://assist.dla.mil -- Downloaded: 20261O-26T11:50Z
Check the source to verify that this is the current version before use.

FOLD

DEPARTMENT OF THE NAVY

POSTAGE AND FEES PAID
NAVY DEPARTMENT

OFFICIAL BUSINESS

COMMANDER
DEFENSE PERSONNEL SUPPORT CENTER (MEDICAL)
2800 SOUTH 20TH ST.
PHILADELPHIA, PA. 19101

FOLD

PLATE NO. 15419 (BACK)

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:50Z
Check the source to verify that this is the currentversion before use.