GG-D-00426b(DSA-DM)
February 17, 1969
INTERIM REVISION OF
Fed. Spec. GG-D-426a
January 9, 1957

INTERIM FEDERAL SPECIFICATION

DISKS; ABRASIVE, SILICON-CARBIDE, DENTAL

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal Specification GG-D-426a.

1. SCOPE AND CLASSIFICATION

* 1.1 Scope. This specification covers flat, unmounted silicon-carbide abrasive disks suitable for attachment to standard dental screw mandrels used in dental handpieces.

1.2 Classification.

* 1.2.1 Classes. The disks shall be of the following classes, as specified (see 6.1):

Class A. Plain, double cutting.
Class B. Plain, safe side.

* 1.2.2 Sizes. The disks shall be of the following sizes, as specified (see 6.1):

Size 1. 7/8 inch in diameter
Size 2. 1/2 inch in diameter

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following specifications and standards, of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Setup.

STANDARD

Federal

FED-STD-123	Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Fort Worth, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

SPECIFICATION

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of Military specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

U.S. DEPARTMENT OF COMMERCE

National Bureau of Standards Commercial Standard (CS)

CS 271-65 Grading of Abrasive Grain for Grinding Wheels

(Application for copies should be addressed to the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402).

3. REQUIREMENTS

3.1 Material.

* 3.1.1 Abrasive. The abrasive shall be high quality silicon carbide crystals and shall be free from extraneous material and suitable for the purpose intended. Abrasive size shall conform to Grit number 90 to 100 of U.S. Department of Commerce Commercial Standard (CS) 271-65.

* 3.1.2 Binder. Disks shall have a vulcanized binder to allow flexibility of disks when in operation. The binder shall be a suitable compound of synthetic rubber.

* 3.2 Description. Disks shall be suitable for use in the mouth as separating disks for cutting metallic dental appliances. All classes and sizes of disks shall be flat, shall have a uniform thickness of 0.023 ± 0.003 inch, and shall be provided with a centered mandrel hole 0.076 ± 0.003 inch in diameter which shall enable disks to be attached to standard dental screw mandrels used in dental appliances.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

* 3.2.1 Class A disks. Class A disks shall be plain flat, double cutting disks. Class A disks shall be furnished in either size 1 or size 2, as specified (see 1.2.2 and 6.1). Size 1 disks shall be 7/8 ± 1/64 inch in diameter; size 2 disks shall be 1/2 inch ± 1/64 inch in diameter.

* 3.2.2 Class B disks. Class B disks shall be plain, flat, safe-sided disks. Class B disks shall be furnished in size 2, 1/2 inch ± 1/64 inch in diameter, (see 1.2.2 and 6.1).

* 3.3 Surfaces. Class A shall have a uniform coating of abrasive which shall cover the entire surface of the disk on both sides. Class B shall have one smooth surface and one uniformly and entirely coated abrasive surface.

3.4 Workmanship. Workmanship shall be first class throughout. Disks shall be symmetrically made, and holes shall be centered. Disks shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

* 4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

* 4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

* 4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

* 4.2 Quality conformance inspection. The examinations required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I and II (see 4.2.1 and 4.3). Sampling for tests shall be as specified in 4.2.2.

* 4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in table I. Unit of product for examination purposes shall be 1 package (box) of disks (100 disks).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.