FEDERAL SPECIFICATION

MIRRORS, HAND

This specification was approved on the above date by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. — This specification covers a round hand mirror suitable for use by dentists to aid patients in the selection of shades and molds of teeth while the teeth are being tried in the mouth. It shall also be suitable for use in demonstrating other types of treatment, including brushing of teeth.

1.2 Classification.

1.2.1 *Size and classes.*—The hand mirrors shall be of one size and of the following classes, as specified:

Class A.—Corrosion-resisting steel frame.
Class B.—Chromium-plated brass frame.

2. APPLICABLE SPECIFICATIONS

2.1 The following Federal Specifications, of the issues in effect on the date of invitation for bids, form a part of this specification:

DD–G–451—Glass, Flat and Corrugated, for Glazing, Mirrors, and Other Uses.
DD–M–411 — Mirrors; Plate Glass, Framed.
GG–I–526—Instruments, Dental and Surgical; General Specifications.

(Copies of Federal specifications and the Federal Specifications Index may be obtained upon application, accompanied by money order, coupon, or cash, to the Superintendent of Documents, Government Printing Office, Washington 25, D. C. Prices may be obtained from the Superintendent of Documents.)

2.2 Specifications applicable only to individual departments are listed in section 7.

3. REQUIREMENTS

3.1 Materials.

3.1.1 *Glass.* — The glass shall be type I polished plate glass of mirror glazing quality, $\frac{1}{8}$ inch $\pm$ $\frac{1}{32}$ inch thick, conforming to the requirements of Federal Specification DD–G–451.

3.1.2 *Steel.*—The steel shall be corrosion resisting and of suitable thickness, conforming to the requirements of Federal Specification GG–I–526, for class 2, composition G.

3.1.3 *Brass.*—The brass shall be of a thickness suitable for the purpose, conforming to the requirements of Federal Specification GG–I–526, for class 4, composition N.

3.2 Components.—The hand mirrors shall consist essentially of the mirror glass, frame and back, and the handle.

3.2.1 *Mirror glass.*—The mirror glass shall be circular with beveled edge. The back of the glass shall be covered with not less than two coats of silver over which shall be protective electroplating of copper conforming to the applicable requirements of Specification DD–M–411, and finally there shall be a heavy coating of water-resistant, baked-on black enamel, which shall also cover the beveled edge of the glass. Layers of water-resistant paper or other suitable cushioning material, adequate to protect the back of the mirror from injury, shall be between the frame and the mirror. The overall diameter of the mirror, including the frame, shall be a minimum of 5 inches and a maximum of 6 inches. The mirror shall be attached to the handle in the same plane as the axis of the handle.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

3.2.2 *Frame and back.*—The frame and back shall be of one-piece construction, fabricated of corrosion-resisting steel as specified in 3.1.2, for class A, or of chromium-plated brass as specified in 3.1.3, for class B. It shall cover the back of the mirror glass and shall be rolled over onto the beveled edge of the mirror so that the mirror is held securely within the frame. The rolled-over edge shall be formed into a bead of sufficient height above the plane of the glass to prevent its coming into contact with any plane surface while in the face-down position and to protect it from being damaged.

3.2.3 *Handle.*—The handle shall be fabricated of either durable plastic, chromium-plated zinc die-casting, or metal of same grade and composition as the frame and back. The handle shall measure approximately 4 inches long and ½ inch in diameter and shall be of a shape suitable for efficient and comfortable use. The handle shall be securely and firmly attached to the frame and back by welding or threaded means; or the handle, frame, and back shall be of one-piece material.

3.3 Finish. — All metal parts shall be smoothly and properly finished and polished to a satin finish.

3.4 Identification marking.—Each mirror shall be legibly and permanently marked with the name or registered trade-mark of the manufacturer.

3.5 Workmanship. — The workmanship shall be first class in every respect. The finished product shall be free from defects which affect its appearance or which may adversely affect its serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Sampling.—Unless otherwise specified, not less than two hand mirrors of each class shall be taken at random from each lot delivered for inspection and tests.

4.2 Inspection. — Samples shall be inspected for material, requirements of component parts, finish, imperfections in glass, workmanship, packaging, packing, and marking.

4.3 Tests.

4.3.1 *Metal and plating.*—Metal and plated parts shall be tested as described in Specification GG–I–526 for the applicable class and composition.

4.3.2 *Mirror backing.*—The mirror backing shall be tested in accordance with the applicable requirements of Specification DD–M–411.

5. PREPARATION FOR DELIVERY

5.1 Packaging. — Unless otherwise specified, each mirror shall be packaged in accordance with accepted commercial practice.

5.2 Packing.—The mirrors shall be prepared for shipment to permit acceptance by carrier for transportation at the lowest applicable rate and to afford maximum protection from normal hazards of transportation.

5.3 Marking.

5.3.1 *Packages.* — Unless otherwise specified, packages shall be marked with the name of the commodity, class, quantity contained therein, and the names of the contractor and the manufacturer.

5.3.2 *Shipping containers.*—Unless otherwise specified, shipping containers shall be marked with the name and class of the commodity, as defined by the contract or order under which the shipment is made, quantity contained therein, name of the contractor, number of the contract or order, name and address of the consignee, and the gross weight.

6. NOTES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

6.1 Ordering data. — Purchasers should specify the class of hand mirror required and should exercise any desired options offered herein. (See 1.2.1, 4.1, 5.1, 5.3.1, 5.3.2, 6.3, and 7.2.6.)

6.2 Stock number. — This specification covers the following item appearing in the Armed Services Catalog of Medical Matériel:

Stock No. 5–416–000, Mirror, Hand, 5-Inch.

6.3 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.4 This specification does not include all types, classes, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., which are normally procured to meet Federal Government requirements.

Patent notice.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

7. DEPARTMENTAL REQUIREMENTS

7.1 The following specifications, of the issues in effect on date of invitation for bids, and special requirements, form a part of this specification for purchases made under this specification by the respective departments.

7.2 Army, Navy, and Air Force.

7.2.1 *Applicable specifications.*

Federal Specifications:

NN–B–601 — Boxes, Wood-Cleated-Plywood, for Domestic Shipment.

NN–B–621—Boxes, Wood, Nailed and Lock-Corner.

Military Specifications:

JAN–P–105—Packaging and Packing for Overseas Shipment—Boxes, Wood, Cleated, Plywood.

JAN–P–106—Packaging and Packing for Overseas Shipment—Boxes; Wood, Nailed.

JAN–P–125—Packaging and Packing for Overseas Shipment—Barrier-Materials, Waterproof, Flexible.

JAN–P–133—Packaging and Packing for Overseas Shipment—Boxes, Set-Up, Paperboard.

JAN–P–140—Packaging and Packing for Overseas Shipment — Adhesives, Water-Resistant, Case-Liner.

MIL–M–3157 — Marking of Domestic Shipments by Contractors for Supplies and Equipment Consigned to the Medical Departments of the Armed Services.

Navy Department Specification:

General Specifications for Inspection of Material.[1]

(Copies of specifications required for Army and Air force purchases may be obtained upon application to the procuring agency. Both the title and identifying symbol or number should be stipulated when requesting copies.

(Copies of Federal and Military specifications required for Navy purchases and Navy Department specifications may be obtained without cost upon application to the Bureau of Supplies and Accounts, Department of the Navy, Washington 25, D. C., ex-

[1] Applicable to Navy purchases only.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

cept that Naval activities should make application to the Commanding Officer, Naval Supply Depot, Scotia 2, New York. Both the title and identifying symbol or number should be stipulated when requesting copies of specifications.)

7.2.2 *Inspection procedures (Navy only).*—For Naval purchases, the general inspection procedures shall be in accordance with Navy Department Specification, General Specifications for Inspection of Material.

7.2.3 *Packaging.*—One mirror constitutes one unit. Each unit shall be wrapped in tissue and packaged in a carton of appropriate size, conforming to the requirements of Specification JAN–P–133. Each carton shall be adequately secured to prevent accidental opening. Carton design shall include suitable interior fittings or sufficient cushioning material to insure against shifting or breakage.

7.2.4 *Packing.*

7.2.4.1 *Domestic.* — Forty-eight packages (48 units) shall be packed in a shipping container conforming to the requirements of Specification NN–B–601 or NN–B–621.

7.2.4.2 *Overseas.* — Forty-eight packages (48 units) shall be packed in a shipping container designed for type-3 load and conforming to the requirements of Specification JAN–P–105 or JAN–P–106. Closure and strapping shall conform to the requirements of the appendix to the applicable box specification. Each shipping container shall be lined with a sealed waterproof bag, or its equivalent, made from material conforming to the requirements of Specification JAN–P–125, for case liners; seams and closures shall be completely sealed with a continuous seam at least ¾ inch wide of water-resistant adhesive conforming to the requirements of Specification JAN–P–140.

7.2.4.3 *Packing variation permitted.* — If the number of units in any shipment is not the exact multiple of the number of units specified to be contained in each shipping container, the units not constituting the nearest such multiple shall be packed in a shipping container of suitable size and design, so constructed as to insure safe delivery to specified destination.

7.2.5 *Marking.*

7.2.5.1 *Packages.*—Each package shall be marked with the stock number, nomenclature, class, unit, quantity, and the name of the contractor. When the contractor is not the manufacturer, the name of the manufacturer shall also appear, as the manufacturer.

7.2.5.2 *Shipping containers.*

7.2.5.2.1 *For domestic shipment.* — Each shipping container shall be marked in accordance with Specification MIL–M–3157.

7.2.5.2.2 *For overseas shipment.* — Each shipping container shall be marked as specified by the procuring agency.

7.2.5.2.3 *Precautionary markings.*—In addition to the markings specified in 7.2.5.2.1 or 7.2.5.2.2, each shipping container shall have a precautionary label permanently affixed to three of its surfaces with waterproof adhesive. The labels shall be RED with white print or WHITE with red print and shall bear a suitable legend indicating that the container contents are fragile.

7.2.6 *Ordering data.* — Procurement documents should specify whether the subject commodity is to be packed and shipping containers marked for domestic or overseas shipment. (See 7.2.4 and 7.2.5.2.)

NAVY INTEREST: Md

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.