FEDERAL SPECIFICATION

NIPPER, INGROWN TOENAIL

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal Agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers two types of ingrown toenail nippers, suitable for use in cutting and in raising and splitting ingrown toenails.

1.2 Classification.

1.2.1 *Types.* Nippers shall be of the following types, as specified (see 6.1):

Type I.—Concave cutting edge.

Type II.—Straight, narrow, tapered jaws.

2. APPLICABLE SPECIFICATIONS AND STANDARDS.

2.1 The following specifications and standards of the issues in effect on date of invitation for bids form a part of this specification to the extent specified herein:

Federal Specifications:

GG–I–526—Instruments, Dental and Surgical; General Specifications.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging, and Packing Levels.

Fed. Std. No. 123—Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington 25, D.C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

Military Specifications:

MIL–P–12902—Packaging of Surgical, Dental and Veterinary Instruments, Diagnostic Equipment, and Related Items.

Military Standards:

MIL–STD–105—Sampling Procedures and Tables for Inspection by Attributes.

MIL–STD–129—Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The nippers shall be fabricated of class 1, composition B material in accordance with Federal Specification GG–I–526.

3.2 Style, design, and dimensions.

3.2.1 *Type I.* Nail nipper with concave cutting edge and volute spring. Type I nail nipper shall be in accordance with figure 1.

3.2.2 *Type II.* Nail nipper with straight, narrow, tapered jaws. Type II nail nippers shall be in accordance with figure 2.

3.3 Finish. Finish shall be as specified in Federal Specification GG–I–526, except that the handle may be satin finish.


3.4 Additional requirements. Except as specified herein, the nippers shall comply with the applicable requirements and tests of Federal Specification GG–I–526.

3.5 Workmanship. Workmanship shall be first class throughout. Nippers shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Supplier responsibilities for inspection. The supplier is responsible for the performance

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.




NOTES:

1. CUTTING EDGES SHALL EXTEND TO TIPS OF BLADES.
2. POINTS OF BLADES SHALL BE SHARP.
3. TOLERANCES INDICATED DELINEATE ACCEPTABLE EXTREME DIMENSIONS, AND ARE NOT INTENDED TO PERMIT VARIATIONS WITHIN A SINGLE INSTRUMENT WHICH WILL SACRIFICE UNIFORMITY, SYMMETRY AND FIRST QUALITY WORKMANSHIP.

FIG. 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.



NOTES: 1. CUTTING EDGES SHALL EXTEND TO TIPS OF BLADES.
2. POINTS OF BLADES SHALL BE SHARP.
3. UNLESS OTHERWISE SPECIFIED TOLERANCES ON FRACTIONS ± 1/32.
4. DIMENSIONS ARE IN INCHES.

FIG. 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 *Certificates of quality.* When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling for lot acceptance.

4.2.1 *Inspection lot.* The inspection lot shall be determined by the cognizant Government inspector and shall be based on the unit of product being one nipper (end item).

4.2.2 *Sampling for examination.* A random sample of nippers shall be selected from each lot offered to the Government in accordance with Military Standard MIL–STD–105. Sampling for examination of dimensions shall be in accordance with Military Standard MIL–STD–105, and as indicated in table 1.

4.2.3 *For tests.* Sampling for tests shall be conducted in accordance with Military Standard MIL–STD–105, and as indicated in table 1.

4.2.4 *Raw materials.* Four ounces of each lot of each metal used in the construction of the nippers shall be selected for chemical analysis.

4.2.5 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The nippers shall be examined to determine compliance with all requirements contained in this specification.

5. PREPARATION FOR DELIVERY

For civil agency procurement, Federal Standard No. 102 should be referred to for definitions and applications of the various levels of packaging and packing protection for supplies and equipment.

5.1 Packaging and packing.

5.1.1. *Unit of issue.* One nipper, as specified, constitutes one unit of issue.

5.1.2 *Packaging quantities.* The number of units of issue indicated in table II shall be packaged in each unit, intermediate and exterior container, as applicable, for the required level of protection specified in the procurement document.

5.1.2.1 *Packing variation permitted.* If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

TABLE I.—*Sampling criteria*

Component	Characteristic	Requirement	Test procedure	Requirement applicable		Unit of product	Inspection level	AQL (percent defective)
				Individual unit	Lot average			
Nipper	Hardness	GG–I–526	GG–I–526	X	----	1-nipper	L3*	(**)
	Salt spray	GG–I–526	GG–I–526	X	----	1-nipper	L3*	(**)
	Boil test	GG–I–526	GG–I–526	X	----	1-nipper	L3*	(**)
	Dimensions	(3.2.1)	----	X	----	1-nipper	L3	4.0
		(3.2.2)	----	X	----	1-nipper	L3	4.0

*A minimum of three units shall be tested regardless of lot size.

**Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

TABLE II.—*Packaging quantites*

Unit package	Intermediate package	Exterior container
1 unit	12 units	144 units

5.1.3 *Level A.* Packaging and packing shall conform to the requirements of Military Specification MIL–P–12902, Procedure Code No. 30

5.1.4 *Level B.* Packaging and packing shall conform to the requirements of Military Specification MIL–P–12902, Procedure Code No. 20.

5.1.5 *Level C.* The subject commodity shall be packaged and packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2 Marking. Nonmilitary.

5.2.1 Unless otherwise specified in the contract or order, unit packages and shipping containers shall be marked in accordance with Federal Standard No. 123.

5.3 Marking. Military.

5.3.1 *Interior packages.* Marking shall conform to the requirements of Military Specification MIL–P–12902, except that the ASMPA stock number will not be required.

5.3.2 *Exterior container.* Exterior container shall be marked in accordance with Military Standard MIL–STD–129.

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein, and procurement documents should specify the following:

(*a*) Title, number, and date of this specification.

(*b*) Type required (see 1.2).

(*c*) Selection of applicable levels of packaging and packing (see 5.1).

(1) Level B preservation and packaging is intended to provide economical but limited protection, and should be specified only when it is determined the nipper will be held in covered storage not more than 12 months from date of initial packaging.

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Federal Government requirements.

6.3 Transportation description. Transportation description and minimum weights applicable to this commodity are:

Rail:

Scissors, not otherwise indexed by name.
Carload minimum weight 20,000 pounds, subject to Rule 34, Uniform Freight Classification.

Motor:

Toenail clippers, not otherwise indexed.
Motor volume minimum weight 20,000 pounds.

Notice.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no repsonsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY CUSTODIANS:
Army—M
Navy—Md
Air Force—AFCSG

Copies of this specification may be purchased for 5 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.