**GG-S-328b**
**FEBRUARY 26, 1958**
**SUPERSEDING**
**Int. Fed. Spec. GG-S-00328a (GSA-FSS)**
**April 13, 1956, and**
**Fed. Spec. GG-S-328**
**December 23, 1943**

**FEDERAL SPECIFICATION**

# SHIELDS, EYE SURGERY

*This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.*

## 1. SCOPE AND CLASSIFICATION

**1.1 Scope.**—This specification covers eye shields designed to eliminate infiltration of light to the eye and to protect an injured or traumatized condition of the eye.

**1.2 Classification.**

1.2.1 *Types and classes.*—Eye shields shall be of the following types and classes, as specified (see 6.2).

Type I.—Shield, eye, single.
Class 1.—Canvas or heavy pressed cardboard.
Class 2.—Plastic.

Type II.—Shield, eye, double.
Class 2.—Plastic.

## 2. APPLICABLE SPECIFICATIONS AND STANDARDS

**2.1 Specifications and standards.**—The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

*Federal Specifications:*

QQ-S-781—Strapping, Flat; Steel.
PPP-B-566 — Boxes, Folding, Paperboard.
PPP-B-585—Boxes; Wood, Wirebound.
PPP-B-601—Boxes, Wood, Cleated-Plywood.
PPP-B-621—Boxes, Wood, Nailed and Lock-Corner.
PPP-B-676 — Boxes, Set-Up, Paperboard.

*Federal Standards:*

Fed. Std. No. 102—Preservation, Packaging, and Packing Levels.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

*Military Specifications:*

MIL-B-10377 — Box, Wood, Cleated, Veneer, Paper Overlaid.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

MIL–L–10547 — Liners, Case, Waterproof.

*Military Standards:*

MIL–STD–105 — Sampling Procedures and Tables for Inspection by Attributes.

MIL–STD–129—Marking for Shipment and Storage.

(Copies of Military specifications and standards required by the contractor in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

## 3. REQUIREMENTS

**3.1 Material.**—Material shall be of either class 1 or class 2, as specified (see 1.2.1 and 6.2).

3.1.1 *Class 1.*—Class 1 shall be canvas or heavy pressed cardboard.

3.1.2 *Class 2.*—Class 2 shall be cellulose acetate plastic or similar nonflammable material (nonflammable material being interpreted as slow-burning as distinguished from fast-burning material, such as cellulose nitrate plastic).

3.1.3 *Tape.*—Tape shall be real or artificial silk.

3.1.4 *Cord.*—Cord shall be rayon covered elastic.

3.1.5 *Covering.*—Covering shall be real or artificial silk.

**3.2 Description.**

3.2.1 *Type 1, shield, eye, single.*—Type I eye shield shall be dome shaped for secure and gentle retention of dressing and shall be designed to fit well over either eye of the average normally developed adult face with adequate clearance for eyelashes.

3.2.1.1 *Class 1, canvas or heavy pressed cardboard.*—Eye shields, fabricated of class 1 material as specified in 3.1.1, shall be covered with black silk (see 3.1.5), on the outer surface of the shield, and unless otherwise specified, shall be covered with black or green silk on the inner surface (see 6.2). Class 1 shields shall be provided with two tapes for tying around the head in order that shield may be held securely in place. The tapes, as specified in 3.1.3, shall be a minimum of 15 inches in length and shall be properly attached to shield for accomplishing satisfactorily the purpose intended. Unless otherwise specified (see 6.2), tapes shall be black. The edges of the shield and covering shall be securely bound or turned in and properly sewn.

3.2.1.2 *Class 2, plastic.* — Eye shields, fabricated of class 2 material as specified in 3.1.2, shall be flesh colored and opaque, and both the inner and outer surfaces shall have a dull frosted finish. All edges shall be smooth. Shields shall be designed with two pierced holes, one near each end of the major axis. Each hole shall be of proper diameter to permit the passage of the cord through it. Shield shall be provided with cord to fit around the head of the patient. The cord, as specified in 3.1.4, shall be a minimum of 18 inches in length (unstretched), and shall be attached to the shield by inserting the cord through the holes in the shield and knotting securely. Cord shall be flesh colored. Material used shall be nonirritating to the skin.

3.2.2. *Type II (class 2 only), shield, eye, double.*—Type II eye shield shall consist of two single eye shields, as specified in 3.2.1 and 3.2.1.2, connected by a cord, as specified in 3.1.4, approximately 1 inch long (unstretched), one end of cord knotted to one shield (through one of the holes in the shield) and the other end of the cord knotted to the other shield. Another piece of cord, minimum length 16 inches (unstretched) to fit around the head of a patient, shall be attached to the shield by knotting one end of the cord to one shield (through its other, or free hole) and



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

knotting the other end of the cord to the other shield.

3.2.3 *Dimensions.* — Type I and type II shields shall be of the following over-all dimensions: 2⅝ inches plus 0, minus ⅛ inch in the horizontal axis, and 2 inches plus 0, minus ⅛ inch in the vertical axis (both measurements taken at the widest points in the horizontal and vertical directions). The thickness of the plastic shall be approximately 0.025 inch.

3.3 **Workmanship.**—Workmanship shall be first class throughout. Eye shields shall be free from defects which detract from their appearance or may impair their serviceability.

## 4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 **Sampling.**—Sampling procedures shall be conducted in accordance with the procedures set forth in Military Standard MIL–STD–105.

4.2 **Inspection.**—The inspector shall ascertain that eye shields have been inspected, tested, and accepted in accordance with specification requirements.

4.3 **Tests.**—Tests shall be conducted in accordance with the applicable specifications, as necessary, to determine compliance with this specification.

## 5. PREPARATION FOR DELIVERY

5.1 **Packaging.**

5.1.1 *Unit of issue.*—One shield, as specified, constitutes one unit.

5.1.2 **Packaging quantities.**—The number of units of issue indicated in table I shall be packaged in each intermediate and exterior container, as specified, for the required level of protection specified in procurement document.

TABLE I.—*Packaging quantities*

| Federal Stock Number | Unit package | Intermediate package | Shipping container |
|---|---|---|---|
| 6515–367–0400 | Not required | 12 | 600 |
| 6515–367–0420 | Not required | 24 | 1,152 |

5.1.2.1 *Packing variation permitted.* — If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.*

5.1.3.1 *Intermediate package.*—The number of units specified in table I shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP–B–566 or PPP–B–676. Each carton shall be adequately secured to prevent accidental opening.

5.1.4 *Level C.*—Unless otherwise specified, shields shall be safely packaged in a suitable cardboard box; each box shall contain one dozen shields.

5.2 **Packing.**

5.2.1 *Level A.*

5.2.1.1 *Exterior container.*—Exterior container shall be designed for type 2 load and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 3 use; PPP–B–601, table II, using type III, class 1 plywood; PPP–B–621, table III; or Military Specification MIL–B–10377, table II.

5.2.1.2 *Case liner.*—Each level A box shall be lined with a waterproof case liner conforming to Military Specification MIL–L–10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.



Source: https://assist.dla.mil -- Downloaded: 202610-26T11:50Z
Check the source to verify that this is the currentversion before use.

5.2.1.3 *Strapping.* — Strapping, when required, shall be in accordance with applicable box specification.

*Note.*—Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 *Level B.*

5.2.2.1 *Exterior container.*—Exterior container shall conform to Federal Specification PPP–B–601, table I, using type III, class 1 plywood, or shall be designed for type 2 load and constructed in accordance with Federal Specification PPP–B–585, style 3, for class 2 use; PPP–B–621, table I; or Military Specification MIL–B–10377, table I.

5.2.3 *Level C.*—Unless otherwise specified, 100 cardboard boxes (1,200 shields) shall be packed in a wood shipping case. Cases shall be made throughout of sound, well seasoned lumber, S1S (the outside) or S2S, finished to not less than 5/8-inch thickness. Except for one piece in each end, side, top, and bottom, none of the boards in any part of the case shall be less than 4½ inches wide. The grain in all boards shall run parallel with the length. Cases shall be of nailed construction with double cleated ends. Nails shall be of proper size and sufficient in number and properly spaced for strong construction. Each case shall be strapped approximately four inches from each end with a painted steel strap in accordance with Federal Specification QQ–S–781. Straps shall be drawn taut, cut into edges, and set tight about all sides of the case.

**5.3 Marking.**

5.3.1 *Intermediate package.* — Each intermediate package shall be marked in accordance with section 5 of Military Standard MIL–STD–129.

5.3.2 *Shipping container.*—Shipping container shall be marked in accordance with Military Standard MIL–STD–129. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

## 6. NOTES

6.1 Federal Standard No. 102 should be referred to for definitions and applications of the various levels of packaging protection for supplies and equipment.

6.2 **Ordering data.** — Purchasers should exercise any desired options offered herein, and procurement documents should specify the following (see 3.2.1.1, 5.1.4, and 5.2.3):

(*a*) Title, number and date of this specification.

(*b*) Type required (see 1.2.1).

(*c*) Class required for type I (see 1.2.1).

(*d*) Color of covering for class 1 (see 3.2.1.1).

(*e*) Color of tapes for class 1 (see 3.2.1.1).

(*f*) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.4 Federal specifications do not include all types, classes, grades, sizes, etc., of the commodities indicated by the title of the specification, or which are commercially available, but are intended to cover the types, etc., which are suitable for Federal Government requirements.

6.5 This specification covers the following items appearing in the Armed Services Medical Stock List:



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

6515–367–0400 — SHIELD, EYE SURGERY, Double.

6515–367–0420 — SHIELD, EYE SURGERY, Single.

6.6 **Transportation description.**—Transportation description applicable to this item is:

Bandages, physicians. (by analogy)

Carload minimum weight 24,000 pounds, subject to Rule 34, Uniform Freight Classification.

Motor volume minimum weight 30,000 pounds.

**Notice.** — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

**MILITARY INTERESTS:**
**Army—M**
**Navy—Md.**

**Copies of this specification may be purchased for 5 cents each.**



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.