GG-T-00311b(DSA-DM)
12 October 1967
Interim Revision of
Fed. Spec. GG-T-311a
2 September 1953

INTERIM FEDERAL SPECIFICATION

# THERMOMETERS, SELF-INDICATING, MERCURY IN GLASS (FEVER)

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal Specification GG-T-311a.

## 1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers fever self-indicating, mercury-in-glass thermometers (see 6.1).

1.2 Classification.

1.2.1 Types. Fever thermometers covered by this specification shall be of the following types, as specified (see 6.2):

Type I. Flat magnifying lens
Type II. Triangular magnifying lens

1.2.2 Classes. Fever thermometers shall be furnished in the following classes, as specified (see 6.2):

Class A. Oral bulb, Fahrenheit scale (cylindrical).
Class B. Oral bulb, Centigrade (Celsius) scale (cylindrical).
Class C. Rectal bulb, Fahrenheit scale, (pear shaped).
Class D. Rectal bulb, Centigrade (Celsius) scale (pear shaped).
Class E. Oral or rectal bulb, Fahrenheit scale (stubby).
Class F. Oral or rectal bulb, Centigrade (Celsius) scale (stubby).

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 The following specifications and standards, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

| | |
|---|---|
| PPP-B-566 | Boxes, Folding, Paperboard. |
| PPP-B-585 | Boxes, Wood, Wirebound. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | Box, Fiberboard. |
| PPP-B-676 | Boxes, Setup. |

STANDARDS

Federal

| | |
|---|---|
| FED-STD-123 | Marking for Domestic Shipment (Civilian Agencies). |

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D. C. 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Military

MIL-L-10547 Liners, Case, and Sheet, Overwrap, Water-Vaporproof, or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage.

(Copies of Military specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Materials.

3.1.1 Stems. The stems shall be of lens front thermometer tubing having white enameled backing. The tubing shall be Corning Standard Thermometer Lead Glass No. 0041 (having a coefficient of expansion of 85 ± 2 x $10^{-7}$ centimeters per centimeter per degree Centigrade (cm/cm/°C.), strain point of 421° C. and annealing point of 463° C.) or equal.

3.1.2 Bulbs. The bulbs shall be made of Corning Normal Glass No. 7560 (having a coefficient of expansion of 86 ± 2 x $10^{-7}$ cm/cm/°C., strain point of 500° C. and annealing point of 539° C.) or equal, or Kimble R-6 (having a coefficient of expansion of 93 ± 2 x $10^{-7}$ cm/cm/°C., strain point of 486° C. and annealing point of 525° C.) or equal. Colored bulbs or bulbs blown from stem glass shall not be used.

3.1.3 Mercury. The mercury shall be pure, clean, dry, and free from foreign material.

3.1.4 Markings. All markings shall be permanent and legible and shall be applied without etching, by application of a stain, enamel or other suitable material fired into the glass. When tested as specified in 4.4.1, the markings shall not fade or discolor in such a manner as to impair their legibility. The loss of one complete graduation or numeral shall be cause for rejecting the thermometer.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

3.1.5 Cases. Unless otherwise specified, cases shall be made of aluminum, nickel- or chrominum-plated metal, hard rubber, or plastic material acceptable to the purchaser. When specified (see 6.2), a case shall be furnished with each thermometer delivered. Cases shall be of suitable size to contain any of the herein specified thermometers and the proper cushioning material. The length of the cases shall not exceed 4-7/8 inches overall. The barrel and cap of case fabricated from hard rubber or plastic material shall be of the screw-on type. All cases shall be well constructed.

3.1.6 Magnifying lens tubing.

3.1.6.1 Flat magnifying lens. Flat magnifying lens tubing for type I thermometers shall have the numerals and graduations on a flat glass surface or surfaces adjacent to the magnifying lens and perpendicular to the line of sight.

3.1.6.2 Triangular magnifying lens. Triangular magnifying lens tubing for type II thermometers shall be triangular in cross section with the magnifying lens at the junction of two sides with the numerals and graduations placed on the sides adjacent to the magnifying lens.

3.2 Dimensions, style, and design.

3.2.1 Length of thermometers. The total length of any thermometer shall not be more than 4-1/4 inches or less than 3-7/8 inches.

3.2.2 Thickness of stem.

3.2.2.1 Type I, flat magnifying stem. Unless otherwise specified, the cross section of the stem shall be such that no dimension is greater than 0.300 inch and no dimension is less than 0.145 inch, when tested as specified in 4.4.2.

3.2.2.2 Type II, triangular magnifying stem. The cross section of the stem shall be such that no dimension is greater than 0.235 inch and no dimension is less than 0.145 inch, when tested as specified in 4.4.2.

3.2.3 Bulb dimensions.

3.2.3.1 Oral (cylindrical) bulbs, classes A and B. The diameter of the bulb shall not be less than 0.078 inch or more than 0.150 inch. The length shall not be less than 7/16 inch or more than 3/4 inch, when measured from the end of the white-enamel backing.

3.2.3.2 Rectal (pear shaped) bulbs, classes C and D. The smaller end of the bulb shall be joined to the stem. The greatest diameter shall be no less than 0.15 inch or more than 0.24 inch. The length shall be no less than 1/4 inch or more than 1/2 inch, when measured from the end of the white-enamel backing.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

3.2.3.3 Oral or rectal (stubby) bulbs, classes E and F. The diameter of the bulb shall be no less than 0.146 inch or more than 0.230 inch. The length shall be no less than 0.125 inch or more than 0.281 inch, when measured from the end of the white-enamel backing.

3.2.4 Scale and numerals. The scale and numerals shall be clearly marked and fired into the glass. The graduation marks shall be substantially straight, regularly spaced, of uniform width, and perpendicular to the axis of the stem. The width of the graduation marks shall be no greater than 0.018 inch or less than 0.004 inch, and the width shall in no case be more than the distance between graduation marks. Subgraduations marks shall be at least 0.050 inch long and main graduation marks shall be at least 25 percent longer. Numerals shall be so placed that the marks to which they pertain can be readily identified.

3.2.5 Fahrenheit thermometers, classes A, C, and E.

3.2.5.1 Range of scale. The scale shall cover a range of at least 96 degrees Fahrenheit (°F.) to 106° F., but may extend below 96° F. or above 106° F., or both. The 96° F. mark shall be no less than 3/8 inch from the top of the constriction, that is, from the point at which the capillary resumes its normal shape above the constriction. The 106° F. mark shall be at least 3/8 inch from the end of the thermometer or 1/8 inch below any part of any cap attached to the thermometer, and at least 1/8 inch from the end of the bore.

3.2.5.2 Graduations. Thermometers shall be graduated in intervals of 0.2 degrees Fahrenheit, each degree mark and the mark at the normal point (98.6° F.) shall be longer than the intervening marks.

3.2.5.3 Numerals. Numerals shall be marked on the scale at even-numbered degree marks.

3.2.5.4 Normal point. The graduation at 98.6° F. shall be designated by an arrow or other suitable mark within a tolerance of ± 1/2 scale division.

3.2.5.5 Degrees per inch of scale. There shall be no more than 9.0 degrees Fahrenheit per inch of scale.

3.2.6 Centigrade (Celsius) thermometers, classes B, D, and F.

3.2.6.1 Range of scale. The scale shall cover a range at least from 35.5° C. to 41.0° C., but may extend below 35.5° C. or above 41.0° C., or both. The 35.5° C. mark shall be no less than 3/8 inch from the top of the constriction, that is, from the point at which the capillary resumes its normal shape above the constriction. The 41.0° C. mark shall be at least 3/8 inch from the end of the thermometer or 1/8 inch below any part of any cap attached to the thermometer, or at least 1/8 inch from the end of the bore.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

3.2.6.2 Graduations. Thermometers shall be graduated in intervals of 0.19 degree Centigrade, each degree and half-degree mark shall be longer than the intervening marks.

3.2.6.3 Numerals. Numerals shall be marked on the scale at each degree mark, except at 37° C.

3.2.6.4 Normal point. The graduation at 37° C. shall be designated by an arrow or other suitable mark within a tolerance of ± 1/2 scale division.

3.2.6.5 Degree per inch of scale. There shall be no more than 5.0 degrees Centigrade per inch of scale.

3.3 Mercury column.

3.3.1 Entrapped gas. All thermometers in which there is sufficient gas or other foreign material to cause separation of the mercury column above the constriction during shipment, test, or use, shall be classified as defective. There shall be no gas or other foreign material in the mercury column or bulb of the thermometer when tested as specified in 4.4.4.

3.3.2 Hard shakers. All thermometers shall be so made that the column of mercury can be shaken down by hand without undue effort, as determined by the test specified in 4.4.5.

3.4 Accuracy.

3.4.1 Fahrenheit thermometers. At the 98° F. and 102° F. points, the thermometers shall be accurate within 0.20 degree Fahrenheit, and at the 106° F. point, within 0.30 degree Fahrenheit, when tested as specified in 4.4.6.

3.4.2 Centigrade thermometers. At the 37° C. and 39° C. points, the thermometers shall be accurate within 0.11 degree Centigrade, and at the 41° C. point, within 0.17 degree Centigrade, when tested as specified in 4.4.6.

3.5 Consistency of reading.

3.5.1 Fahrenheit thermometers. The two readings of the thermometer at 98° F. and 102° F. shall not differ from each other by more than 0.20 degree Fahrenheit, and the two readings at 106° F. shall not differ from each other by more than 0.30 degree Fahrenheit, when tested as specified in 4.4.6.

3.5.2 Centigrade thermometers. The two readings of a thermometer at 37° C. and 39° C. test points shall not differ from each other by more than 0.11 degree Centigrade, and the two readings at 41° C. shall not differ from each other by more than 0.17 degree Centigrade, when tested as specified in 4.4.6.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

3.6 Aging. Thermometers shall be aged so that the performance requirements involved will be met after they have been issued for use. Thermometers shall be aged after sealing of bulb and stem. All thermometers shall be artificially aged by holding each thermometer at a temperature of 830° F. ± 10° F. for at least two hours, followed by cooling to 500° F. at a rate not to exceed 100° F. per hour. As an alternate, thermometers having stems fabricated from material specified in 3.1.1 and bulbs fabricated from Corning Normal Glass No. 7560, or equal, may be naturally aged by storing them at room temperature for at least four months.

3.6.1 Certificate of aging. A certificate listing all lot or control numbers shall state that each thermometer has been aged as specified above. The certificate shall include the method used.

3.6.1.1 For Military procurement. When thermometers are purchased by the Defense Supply Agency, certificates shall be furnished to the procuring activity via the government inspector, when source inspection is required. For destination inspection, certificates shall be submitted by the contractor directly to Technical Operations Division, Directorate of Medical Materiel, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa. 19101, attention: Quality Control Branch, and one copy of each shall be forwarded with each shipment to the consignee.

3.7 Identification marking. Each thermometer shall bear in permanent, legible characters the name or registered trademark of the manufacturer and either a serial number or code designation indicating date of calibration within 30 days.

3.8 Workmanship. Workmanship shall be of the best commercial grade. All thermometers shall be free from constructional defects which prevent observations of temperatures within a tolerance of ± 0.2 degree Fahrenheit or ± 0.1 degree Centigrade in the range of 96° F. to 104° F. or 35° C. to 40° C., and within ± 0.3 degree Fahrenheit or ± 0.15 degree Centigrade in the range of 104° C. to 106° F., or 40° C. to 41° C. Unhealed fire cracks or fractures shall be considered evidence of discreditable workmanship.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

4.1.1 Inspection. Inspection, as used in this specification is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties).

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I.

Table I. Sampling for Examination

| Type Defect | Inspection level | AQL (percent defective) |
|---|---|---|
| Major | II | 1.5 |
| Minor | II | 2.5 |

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and with table II.

Table II. Sampling for Tests

| Characteristic | Req. | Test Procedure | Unit of Product | Insp. Level | AQL (percent defective) |
|---|---|---|---|---|---|
| Permanency of Marking | 3.1.4 | 4.4.1 | 1 Thermometer | S-1 | * |
| Entrapped Gas | 3.3.1 | 4.4.4 | 1 Thermometer | S-3 | * |
| Hard Shaker | 3.3.2 | 4.4.5 | 1 Thermometer | S-3 | * |
| Accuracy | 3.4 | 4.4.6 | 1 Thermometer | S-3 | * |
| Consistency | 3.5 | 4.4.6 | 1 Thermometer | S-3 | * |

* Acceptance number shall be 1.5 total for all five tests.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

4.3 Examination. The thermometers shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with tables I and III. Inspection, however, shall not be restricted to the defects listed in table III. Any thermometers in the sample containing one or more defects shall be rejected, and if the number of defective thermometers in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Table III. Classification of Defects in Accordance with MIL-STD-105

| Categories | Defects |
|---|---|
| Critical: | None Defined |
| Major: | |
| 101 | Crack or chipped out area in stem or bulb. |
| 102 | Case cracked or broken. |
| 103 | White enameled backing missing or not complete. |
| 104 | Magnifying lens front does not function properly, or is omitted. |
| 105 | Graduation markings missing. |
| 106 | Graduation markings not in intervals of 0.2 degree Fahrenheit or 0.1 degree Centigrade. |
| 107 | Numerals not marked at every 2 degrees Fahrenheit beginning at lowest even degree, or at each degree Centigrade except 37° C. |
| 108 | Scale range not at least 96 to 106 degrees Fahrenheit or at least 35.5 to 41.0 degrees Centigrade. |
| 109 | Thermometer not type and class specified. |
| 110 | Thermometer not free from sharp edges or projections. |
| 111 | Thermometer not properly protected by cushioning material in case. |
| 112 | Thermometer not free from breaks or foreign matter in column. |
| 113 | Barrel and cap of hard rubber or plastic case not "screw-on" type. |
| 114 | Identification marking interferes with reading of the thermometer. |
| 115 | Case and cushioning material not of size suitable for the thermometer. |
| Minor: | |
| 201 | Normal point not marked with arrow. |
| 202 | Degree lines not longer than the intervening lines. |
| 203 | Graduation marks not straight, uniform and regularly spaced. |
| 204 | Numerals so placed that marks to which they pertain are not readily identified. |
| 205 | Identification markings missing or not legible. |
| 206 | Poor workmanship (not resulting in a major defect). |
| 207 | Dimensions of thermometer and of scale not as specified. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements. In the absence of detailed descriptions of test methods in these specifications, compliance with the requirements shall be determined by methods appropriate to the several requirements.

4.4.1 Permanency of marking test. Thermometers shall be completely immersed in concentrated hydrochloric acid (specific gravity 1.18) at room temperature, covered, and allowed to stand in the acid for 48 hours. The thermometer shall then be removed and dried by rubbing vigorously with a cloth. The thermometer shall then be completely immersed in concentrated sulfuric acid (specific gravity 1.84) at room temperature and allowed to stand for one hour. The thermometer shall then be removed and dried by rubbing vigorously with a cloth.

4.4.2 Tests for thickness of stem. Stem dimensions shall be tested with ring gages of the maximum size specified and a slot gage of the minimum size specified.

4.4.3 Tests for diameter of bulb. The diameter of the bulb shall be tested with maximum and minimum ring gages of the sizes specified.

4.4.4 Test for entrapped gas.

4.4.4.1 Visual inspection. Thermometers in which visual inspection shows the presence of gas in the bulb shall be classified as defective.

4.4.4.2 Centrifuge test. The thermometer shall be heated so that the reading is 102° F. or 39° C. The mercury shall be shaken away from the bulb in the centrifuge at 200 revolutions per minute (r.p.m.) with the top of the thermometer 32 centimeters (cm.) from the axis of rotation. A broken mercury column above the constriction shall be an indication that there is gas or other foreign material present.

4.4.5 Test for hard shakers. All thermometers, after having been heated to 106° F. or 41° C., shall be mounted in a centrifuge, and whirled at the specified speed. For a centrifuge in which the ends of the bulbs are 17 cm. from the axis of rotation, the specified speed is 580 r.p.m. For a centrifuge in which the ends of the bulbs are 32 cm. from the axis of rotation, the specified speed is 400 r.p.m. The larger type is preferred. When thermometer is whirled at the specified speed for one minute, the index shall fall below 96° F. or 35° C. Thermometers which do not comply shall be classified as defective.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

4.4.6 Tests for accuracy and consistency of reading. Thermometers shall be compared at 98°, 102°, and 106° F., and 37°, 39°, and 41° C., with a thermometer of appropriate scale certified by the National Bureau of Standards for this purpose, by being heated in a well-stirred water bath, removed from bath, and read. The temperature of the bath shall be raised at a uniform rate during the last two minutes before the test temperature is reached, and the rise during this two-minute period shall be no more than 0.2 degree Fahrenheit. The thermometer shall be removed from the bath as soon as the temperature of test is reached, except in the second test at 106° F. where the procedure specified in paragraph 4.4.6.2 shall be followed. A suitable magnifying glass shall be used and the readings estimated to the nearest one-tenth division (graduation interval), or better. Two independent comparisons shall be made and recorded for each thermometer at each test temperature. In the first test, at 106° F., or 41° C., the thermometers shall be removed from the bath as soon as the test temperature is reached. In the second test at 106° F. or 41° C., the bath shall be brought up to the test temperature and then allowed to cool slowly to 105° F. or 40.4° C., or below, at a uniform rate not exceeding 1° F. or 0.55° C. in three minutes. The thermometers shall then be removed from the bath and read, and the readings recorded as before.

4.4.6.1 Accuracy. No individual reading, in comparison with the certified standard, shall be in error by more than the amount permitted in 3.4.

4.4.6.2 Consistency. The two readings at each test point shall not differ from each other by more than the amount permitted in 3.5, except as provided in 4.4.6.3.

4.4.6.3 Retest. If either reading at one and only one test temperature is in error by more than the tolerance specified in 3.4, or if the readings at one and only one test temperature differ from each other by more than the tolerance specified in 3.5, the thermometer shall be completely retested for accuracy or consistency of reading. If, on retest, the thermometer fails to meet the requirements of 4.4.6.1 or 4.4.6.2, it shall be classified as defective.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

## 5. PREPARATION FOR DELIVERY

### 5.1 Packaging.

5.1.1 Unit of issue. One each thermometer with case for federal stock number (FSN) 6515-793-2000 (type II, class A) or 6515-793-2075 (type II, class C); or one package (6 thermometers) for FSN 6515-299-8258 (type II, class A) or FSN 6515-299-8263 (type II, class C), as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

| Type | Class | Federal Stock Number | Unit package | Intermediate package | Exterior container |
|---|---|---|---|---|---|
| II | A | 6515-793-2000 | NOT REQUIRED | 144 Units | 1728 Units |
| II | C | 6515-793-2075 | NOT REQUIRED | 144 Units | 1728 Units |
| II | A | 6515-299-8258 | 1 Unit | 24 Units | 288 Units |
| II | C | 6515-299-8263 | 1 Unit | 24 Units | 288 Units |

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package, FSN 6515-793-2000 (type II, class A), and 6515-793-2075 (type II, class C). Unit packaging not required.

5.1.3.2 Unit package, FSN 6515-299-8258 (type II, class A), and 6515-299-8263 (type II, class C). Each unit (6 thermometers) shall be packaged in a carton of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676. Carton shall be designed with suitable interior fittings or shall contain sufficient cushioning material to individually protect each thermometer from damage. Closure shall be adequate to prevent accidental opening.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

5.1.3.3 Intermediate package. Intermediate package shall be a carton of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic and method I closure shall be utilized.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Civil agencies. Shipments shall be marked in accordance with FED-STD-123.

5.3.2 Military agencies.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The thermometers covered by this specification are of the types commonly used for measuring body temperatures, except that special thermometers having different ranges and degrees of subdivision designed for a specific use, such as for tubercular cases and for ovulation temperature measurements are not included.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Type and class of thermometer required.
(c) Whether or not case is required (see 3.1.5)
(d) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Federal Government requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

*U.S. GOVERNMENT PRINTING OFFICE 1967-701-513 2266



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

# SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION: CO-T-00311b(DSA-DM)

ORGANIZATION:

CITY AND STATE:

CONTRACT NUMBER:

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO (If "yes", in what way?)

4. REMARKS (Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)

SUBMITTED BY (Printed or typed name and activity - Optional)

DATE

DD FORM 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

**CAUTION**

**NOTICE TO BIDDERS/OFFERORS**

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

---

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

---

Fold



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.