GG-T-0031lb
Amendment - 1
29 July 1970

INTERIM FEDERAL SPECIFICATION

THERMOMETERS, SELF-INDICATING, MERCURY IN GLASS (FEVER)

This amendment was developed by Defense Personnel Support Center, Directorate of Medical Materiel, 2800 South 20th Street, Philadelphia, Pennsylvania, 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

The General Services Administration has authorized the use of this amendment as a valid exception to Interim Federal Specification GG-T-0031lb, dated 12 October 1967.

Add the following paragraph:

"2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

American Society for Testing Materials Standards (ASTM)

ASTM-C-336 Test for Annealing Point and Strain Point of Glass

(Application for copies should be addressed to A. S. T. M., 1916 Race Street, Philadelphia, Pa. 19103)."

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

3.1.1 Stems. Line 4, after "421°C" add "±10°C"; line 5, after "463°C" add "±10°C".

At end of paragraph add: "Strain and annealing points shall be tested as specified in 4.4.7."

3.1.2 Bulbs. Line 3, after "500°C" add "±10°C"; line 3, after 539°C" add "±10°C"; line 4, after "486°C" add "±10°C"; line 5, after "525°C" ± 10°C".

At end of paragraph add: "Strain and annealing points shall be tested as specified in 4.4.7."

Delete 4.1 thru 4.1.2 and substitute:

"4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

"4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

"4.1.2 Inspection. Inspection, as used in this specification is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:50Z
Check the source to verify that this is the currentversion before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the glass and mercury, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements."

4.2.2 For tests. In Table II, under column "AQL (percent defective)", delete the five asterisks ("*"), and substitute "1.5" in the five places thereof. Delete following under Table II:··"*Acceptance number shall be 1.5 total for all five tests."

Add the following paragraph:

"4.4.7 Test for strain and annealing points. Strain and annealing points shall be tested in accordance with ASTM-C-336 - Test Method."

5.2 Packing. Delete entirely and substitute:

"5.2 Packing.

"5.2.1 Level A.

"5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

"5.2.2 Level B.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.

"5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

"5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order."

Add the following paragraph:

"6.4 This specification covers the following items appearing the Federal Supply Catalog:

Stock No.	Item Identification
6515-299-8258	THERMOMETER, CLINICAL, HUMAN, Oral, 6s
6515-299-8263	THERMOMETER, CLINICAL, HUMAN, Rectal, 6s
6515-793-2000	THERMOMETER, CLINICAL, HUMAN, Oral
6515-793-2075	THERMOMETER, CLINICAL, HUMAN, Rectal

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:

Defense Supply Agency - DM

Project No. 6515-1486

Review activities:

Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.