INTERIM AMENDMENT

TO

FEDERAL SPECIFICATION

THERMOMETER, SELF-INDICATING, LIQUID-IN-GLASS, MAXIMUM, (METEOROLOGICAL)

This Interim Amendment was developed by the General Services Administration, Federal Supply Service, Washington, D C 20406, based upon currently available technical information It is recommended that Federal agencies use it in procurment and forward recommendations for changes to the preparing activity at the address shown above

The General Services Administration has authorized the use of this interim amendment as a valid exception to Federal Specification GG-T-355B, dated September 25, 1967

PAGE 3

Paragraph 3 5 3, line 4 Delete "and the range scale length shall be not less than 7-1/2 inches" and substitute "and the range scale length shall be not less than 7 inches"

PAGE 10

Figure 2, from the top of the page, line 10 Delete "NOT LESS THAN 7-1/2" and substitute "NOT LESS THAN 7"

FSC 6660

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:50Z
Check the source to verify that this is the current version before use.