A-A-50994B
10 July 1990
SUPERSEDING
A-A-0050994A(DM)
25 July 1986

COMMERCIAL ITEM DESCRIPTION

CATHETER, URETHRAL, (FOLEY-TYPE)

The General Services Administration has authorized the use of commercial item description as a replacement for Military Specification MIL-C-36891(1) which is cancelled.

This commercial item description covers a Foley-Type urethral catheter suitable for use in urinary drainage during surgical procedures.

Salient characteristics:

Except as specified below, the catheter shall be in accordance with ASTM Designation F623 Foley Catheters, for requirements, test methods, and style and design.

The catheter shall be fabricated from soft, durable natural latex rubber. The entire lumen and external surface of the catheter shall be bonded with a layer of silicone rubber. The valve shall be fabricated from a rigid plastic compound.

The catheter shall be constructed with an integral self-retaining balloon. The capacity of the balloon, inside diameter and the breaking strength shall be as specified in Table I for each respective French size.

Table I. Dimensions and Physical Properties.

French Size (FR)	Capacity of Balloon (ml)	Inside Diameter Minimum (Inches)	Breaking Strength (Minimum Pounds)
16	5	0.084	58
18	5	0.098	70
22	30	0.140	94
24	30	0.171	106

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Sterility. The catheter shall be sterile when tested in accordance with USP or AAMI guidelines. Sterility assurance level shall be 0.000001.

Workmanship. The catheter shall be processed in such a manner as to be uniform in quality and free from defects that will affect life, serviceability, or appearance.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). Each catheter shall be individually packaged, uncoiled, in a peel-open type container capable of maintaining sterility of contents unless package is damaged or opened. One package containing 10 or 12 individually packaged catheters as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Sampling for tests. Sampling shall be conducted in accordance with MIL-STD-105, inspection level S-1, AQL 1.0. The unit of product for tests purposes shall be one catheter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

Physical properties. Two straight specimens shall be cut from each test unit (catheter). One specimen shall be tested as cut. The other specimen shall be subjected to accelerated aging in accordance with Federal Method Standard 601.

The specimens shall be tested for breaking strength and ultimate elongation in accordance with Method 4111 and 4121 of Federal Tests Methods 601 respectively, except paragraphs 5.2 and 5.3 of each Method shall be disregarded. The valve of breaking strength shall be the valve of the breaking force measured by Method 4111.

Unaged specimens shall have a minimum breaking strength as shown in Table I and an average ultimate elongation (arithmetic mean) of 750 percent. No unaged specimen shall have an ultimate elongation less than 680 %. The breaking strength and ultimate elongation of the aged specimen shall not be less than 75 percent of the values obtained for the unaged specimen cut from the catheter,

Any break which can be shown to have been caused by the test machine grips shall be classified as an invalid test result, and a complete entire catheter shall be re-tested.

Performance. Catheters shall be tested for performance in accordance with the performance tests ASTM Designation F623. As an alternate test method for the Balloon Integrity and Balloon Volume Maintenance Tests, catheters may be overinflated with 50cc of air, submerged in water, and the entire inflation system checked for leakage. Sampling for performance tests shall be S-1 (minimum) with an AQL of 1.0.

As an alternate test method for shaft size, the shaft may be measured by any acceptable method without cutting shaft.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the lastest revision of Federal Standard No. 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the products is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Spubpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-104-8694	CATHETER, URETHRAL, SILICONIZED RUBBER, DISPOSABLE, ROUND TIP, FOLEY, With 5 ml Balloon, 16 Fr, 10s or 12s Sterile

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

6515-00-104-8695	CATHETER, URETHRAL, SILICONIZED RUBBER, DISPOSABLE, ROUND TIP, FOLEY, With 5 ml Balloon, 18 Fr, 10s or 12s Sterile
6515-00-104-8696	CATHETER, URETHRAL, SILICONIZED RUBBER, DISPOSABLE, ROUND TIP, FOLEY, With 30 ml Balloon, 22 Fr, 10s or 12s Sterile
6515-00-104-9006	CATHETER, URETHRAL, SILICONIZED RUBBER, DISPOSABLE, ROUND TIP, FOLEY, With 30 ml Balloon, 24 Fr, 10s or 12s Sterile

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 05116

Project No. 6515-4779

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-50994B	CATHETERS, URETHRAL, (FOLEY-TYPE)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:49Z
Check the source to verify that this is the current version before use.