A-A-51941
29 May 1987

COMMERCIAL ITEM DESCRIPTION

POSITIONING SET, SURGICAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a vacuum moldable pad set suitable for holding a patient stably and securely on an operating table.

Salient characteristics:

The surgical positioning set shall consist of five moldable pads with conductive covers which, when molded, hold the patient stably and securely on an operating table.

The positioning set shall consist of one each of the following nominally sized pads which shall be constructed of rubberized plastic skin filled with plastic beads, approximately 2 mm to 6 mm in diameter.

a. 16 inches (40.6 cm) by 12 inches (30.5 cm) with an 8 inch (20.3 cm) U-shaped neck cutout in the center of one 16 inch side.

b. 20-3/4 inches (52.7 cm) by 11 inches (27.9 cm)

c. 18-1/2 inches (47.0 cm) wide at wide end by 36-1/2 inches (92.7 cm) long by 14-3/4 inches (37.5 cm) wide at narrow end.

d. 35-1/2 inches (90.2 cm) by 28-3/4 inches (73.0 cm) with an 8 inch (20.3 cm) U-shaped cutout in the center of one 28-3/4 inches side.

e. 34-1/2 inches (87.6cm) by 28 inches (71.1 cm) with two 12 inch by 6 inch corners removed on one of the 34-1/2 inch sides.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

The pads shall be soft and moldable when the inside pad pressure is atmospheric. Each pad shall be capable of becoming a solid molded mass by creating a vacuum within the pad. A vacuum of up to -8 mm Hg from atmospheric pressure shall be created through the attachment of a vacuum line. All pads, except size "2", shall have internal partitions to help assure even distribution of the beads.

Each pad shall be equipped with a valve which closes automatically following the application of a vacuum. Once closed the automatic valve shall seal the vacuum created in the pad. The automatic valve shall be capable of being manually opened. Once the automatic valve is opened and the pressure inside the pad returns to atmospheric, the pad shall regain a soft moldable condition. The automatic valve shall also have an attachable cap to further assure no air leakage after vacuum application.

Each pad shall be furnished with a conductive cover.

Pads shall be able to be disinfected or sterilized with liquid disinfectants or gas sterilized.

Pads shall have x-ray transparency with only minimal and clearly identifiable artifacts.

All pads shall be moldable. Pads shall not be ripped, torn, or improperly sealed. All pads shall be clean.

Workmanship. The positioning set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). One set consisting of 5 moldable pads, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:03Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling.

For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one complete surgical positioning set as specified.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defect	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional defects	S-1	2.5

Examination. The positioning set shall be examined to determine compliance with the requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

TABLE II. Categories and Defects *

Categories	Defects
Major	
101	Any Component of the set missing, broken, or otherwise unusable.
102	Any pad not soft and moldable when the inside pad pressure is atmospheric.
Minor	
201	Pad covers are not smooth.
202	Pad covers are not clean and free of foreign material.

Dimensional Examination. The surgical positioning set shall be examined for defects in dimensions. Any dimension not as specified herein shall be classified as a defect.

Tests. The surgical positioning set shall be tested to determine compliance with this specification. Testing shall include the following:

Each pad will be attached to a vacuum line. Failure to achieve a vacuum of -8mm from atmosperic pressure shall constitute a major defect.

Each pad shall be drawn to a vacuum of -8mm from atmospheric pressure and vacuum line then removed. Failure to seal the vacuum for a minimum period of one hour shall constitue a major defect.

Each pad will be gas sterilized for a period of eight hours. Any damage to the pad at the completion of gas sterilization shall constitute a major defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4. to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-166-5093

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1894

PGC 51356

Location: CID/POSITIONING/H13-12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.