A-A-51953
29 May 1987

COMMERCIAL ITEM DESCRIPTION

APRON, X-RAY PROTECTIVE
(Dental Patient)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a lead-impregnated vinyl plastic apron suitable for protection of patients during dental x-ray examinations.

Salient characteristics:

The x-ray protective apron shall be constructed of three layers of lead impregnated vinyl sheeting. Each layer shall be a minimum of 0.015 inch thick. The lead impregnated sheets shall be sandwiched between a top sheet of any solid color vinyl sheet, except black or white, a minimum of 0.009 inch thick, and a flexible urethane foam laminated vinyl sheet, a minimum of 0.005 inch thick. The urethane foam shall provide a clinging action to the patient's clothing, eliminating the need for ties or fasteners on the apron.

The five sheets shall be bound together with a sewn cotton binding or other suitable edge binding. All edges shall be tightly secured, with no openings. All stitching shall be strong but shall not damage the material. The stitching shall be at least 6 stitches per inch for each line of stitching used.

The apron shall be formed to provide protection to the chest, abdomen, neck, and groin area of a sitting patient. The apron shall provide a minimum lead protection equivalency of 0.30 mm up to 85 KVP.

The apron shall be approximately 35 inches long and 23 inches wide. The apron shall have a tapered neck cut-out so that the distance from the bottom of the apron to the bottom (narrow portion) of the neck cut-out is approximately 27 inches. The distance across the neck cut-out from one shoulder piece to the other shall be approximately 8 inches. The height of the neck cut-out shall be approximately 8 inches.

AMSC N/A FSC 6532

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

The apron shall have a neck collar with a velcro closure. The collar shall be fabricated of the same materials as the body and shall be three inches wide by twenty-four inches long prior to sewing. A three inch long by one inch wide velcro tab shall be attached to the collar to facilitate securing of the apron to the patient with adjustment to the patient's neck size.

The finished apron shall be supple, washable, resistant to cracking after repeated flexing, shall not soil user's clothing when worn, and shall not exceed 7 pounds (3.15 Kg) in total weight.

The apron shall have no cuts., holes, blisters, thin or hard spots, or abraded areas. The apron shall have no repairs or patches.

The apron shall be clean and free of grease or oil stains, or embedded foreign material.

Workmanship. The x-ray protective apron shall be free of defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One x-ray protective apron, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one x-ray protective apron as specified in this description.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defect	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional examination	S-1	2.5

Examination. The x-ray protective apron shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

TABLE II. Categories and Defects *

Categories	Defects
Major	
101	Apron does not have a flexible urethane foam laminated vinyl sheet.
102	Apron does not have velcro closure on neck collar.
103	Apron does not have edge binding.
104	Apron has a cut, hole, or thin spot.
Minor	
201	Apron stitching is not at least 6 stitches per inch for each line of stitching used.
202	Apron has a repair or patch.
203	Apron is not clean and free of grease or oil stains, or embedded foreign material.

Dimensional Examination. The x-ray protective apron shall be examined for defects in dimensions. Any dimension not as specified herein shall be classified as a defect.

Tests

Lead Equivalent test.

Contractor shall radiograph two apron samples at 85 KVP with Intensifying screens at an appropriate milliampere-seconds and target-film distance using a lead step wedge. The lead equivalent of the two samples shall then be determined. Failure to achieve an equivalency of 0.30 mm on either of the two samples shall constitute a major defect.

Weight test

The apron shall be weighed. Failure to weigh 7 pounds or less shall constitute a major defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6532-01-019-6933

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

A-A-51953

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6532-1168

PGC 50812

Location: CID/XR-APRON/H13-12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4066 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTC
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51953	APRON, X-RAY PROTECTIVE(Dental Patient)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.