COMMERCIAL ITEM DESCRIPTION

BULB, PIPET, BACTERIOLOGICAL (Rubber)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers an amber color rubber bulb fo use with disposable glass Pasteur pipets.

Salient characteristics:

Shall be an amber color 1 milliliter (ml) capacity rubber bulb for use with glass Pasteur pipets conforming to ASTM E 732 (Standard Specification for Disposable Pasteur-type Pipet).

Shall be fabricated from a highly flexible amber latex that allows for easy insertion into and removal from a glass Pasteur pipet. In addition, the open end fabrication of the bulb shall include a rolled rim in the form of a doughnut shaped ring.

Workmanship. The rubber bulb shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing twelve bulbs, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exter. package quantities and labeling and marking shall be as specified i the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-01-252-0078

The following National Stock Numbers covers the Pasteur pipets intended for use with the rubber bulb.

6640-00-053-9835
6640-00-782-6008
6640-01-033-3137

Ordering data. (Intermediate/exterior package quantities and label and marking must be specified in the contract and/or order.)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
FDA-MPQAS
PHS

Project No. 6640-1451

Location: Disk #3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for us ubmitting comments and suggestions for improvements All users of military standardization documents are invited to pro suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, o rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorizatio to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POST
NECESS
IF MAIL
IN THE
UNITED ST

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-51968A	BULB, PIPET, BACTERIOLOGICAL (RUBBER)

3a NAME OF SUBMITTING ORGANIZATION

b ADDRESS *(Street City State ZIP Code)*

4 TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *Last First MI* – Optional

b WORK TELEPHONE NUMBER *(Include Code)* – Optional

c MAILING ADDRESS *(Street City State ZIP Code)* – Optional

8 DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.