A-A-51969
2 June 1987

COMMERCIAL ITEM DESCRIPTION

PRESSURE INDICATING COMPOUND and BRUSH, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a pressure indicating compound and a brush (replacement).

Type I - Pressure Indicating Compound
Type II - Brush (replacement)

Salient characteristics:

Type I pressure indicating compound shall be suitable for use in accurately locating areas of pressure on the internal surfaces of dentures, inlays and crowns. Type II, brush (replacement) shall be a disposable, dental, pressure indicating compound brush suitable for use with NSN 6520-00-890-2185, Pressure Indicating Compound, Dental.

Pressure Indicating Compound. The pressure indicating compound shall be furnished in a kit consisting of the following:

1/2 ounce, minimum, white silicone paste in a suitable collapsible metal tube or as an alternate, in a glass jar with a screw cap.

2 ounces, minimum, silicon emulsion in a plastic bottle with screw cap.

1 suitable plastic atomizer with screw cap for spraying the emulsion.

1 brush suitable for spreading the paste on denture surfaces.

1 complete, accurate manufacturer's instructions.

The paste shall adhere to dry surfaces. The consistency shall be free-flowing and shall be such that the paste can be lightly brushed on the inner denture surfaces being tested. When sprayed with silicone emulsion (wetting agent), the treated area shall not adhere to the moist oral tissues.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

When used in accordance with manufacturer's instruction, moderate hand or closing pressure shall displace the pressure indicating paste, clearly outlining excessive pressure areas to permit precise adjustments. The paste and emulsion shall be suitable for use in contact with the oral mucosa. The compound shall clean readily from dentures, crowns, and inlays after normal use. The materials shall exhibit characteristics of long shelf life.

The materials shall not have either an unpleasant odor or flavor.

The paste shall have a smooth, creamy white appearance.

Consistency. The silicone paste shall have penetration values between 261 and 315 when tested in accordance with American Society for Testing and Materials (ASTM) Designation D 217. The consistency shall remain constant under ambient temperature conditions.

Material. The paste shall be a homogeneous, silicone mixture colored white by a suitable coloring agent. The mixture shall be free-flowing under slight pressure. The silicone emulsion shall be suitable for use as a spray which acts as a wetting agent following application of the paste. The pressure indicating compound shall contain no ingredients which are toxic or irritating to the oral tissues.

Brush. (replacement). The brush shall consist of fibers held by a tubular metal handle. The handle shall be crimped on one end to hold the fibers securely. The brush shall contain adequate fibers to spread the compound. The fiber bundle shall be 3/8 $\pm$ 1/16 inch wide and 3/4 $\pm$ 1/16 inch long. The tube shall be 1/4 $\pm$ 1/32 inch in diameter, with an overall length of 3-1/4 $\pm$ 1/8 inches. The fibers shall be made from natural animal bristles.

Workmanship. Pressure indicating compound shall be free of lumps or granules or from any defect which detracts from its appearance or impairs its serviceability. The brush shall be free of stiff, inflexible or loose fibers or any defect which detracts from its appearance or impairs its serviceability.

Unit. Package (PG). One package containing 1/2 oz of silicone, 2 oz silicone emulsion, atomizer, brush, and instructions, or one package containing 12 brushes, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time, to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used in this specification is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one package of pressure indicating compound or one package of brushes, as specified. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:03Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

Stock Number	Item Identification	Designation
6520-00-890-2185	PRESSURE INDICATING COMPOUND, DENTAL Paste, 1/2 oz.	Type I
6520-01-125-6596	BRUSH, DENTAL PRESSURE INDICATING COMPOUND, Disposable, 12s	Type II

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVIES:
VA-OSS
PHS
FDA-MPQAS

06004

PROJECT NO. 6520-1994

Location: NCR Disk #2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-04-26T09:03Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
AA-51969	PRESS'R Indicating Compound & Brush, Dental

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.