A-A-0051974(DM)
29 May 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

TRANSFER SET, INTRAVENOUS SOLUTION
(Non-Vented, 27 Inches)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a sterile, non-vented, disposable intravenous solution transfer set.

Salient characteristics:

Materials. The transfer set shall be sterile, non-vented, disposable and suitable for transfering medication from one bag to another or from one bottle to another when the bottles are equipped with airways.

All parts of the transfer set in contact with the intravenous solution being transferred shall comply with USP Class IV plastic requirements. Also the same surfaces shall be in conformance with USP requirements for non-pyrogenicity and non- toxicity. The clear tubing shall be made of polyvinyl chloride having a Shore Durometer A hardness of 71 ± 5. The spike, roller clamp, needle adapter and protective caps shall be made of suitable rigid plastic.

The spike must be able to be immersed in isopropanol for 5 minutes, dried on paper toweling for an hour and then examined for crazing or cracking. There shall be no such signs when examined at 2X magnification.

Construction. The transfer set shall consist of a spike with flange to facilitate handling, polyvinyl chloride tubing, roller clamp and needle adapter. A protective cap shall be present on the spike and on the needle adapter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

The overall length shall be 27 inches +5 inches, -1 inch long. The spike shall be 3/4 to 13/16 inch long, have a single lumen and have a sharp, durable conical point. There shall be a slight taper to the first shoulder. The diameter just above the shoulder shall be 0.210 +.003, -.002 inch.

The tubing shall have an ID of 0.110 ± .005 inch and a wall thickness of 0.022 ± .003 inch. The male Luer needle adapter shall be of slip tip design and conform to ANSI/HIMA MD70.1. The roller clamp shall control flow from closed to fully open positions and at any intervening adjustment between these extremes. It shall be able to hold the adjustment constant. The roller in the fully open position shall be able to withstand a 3.65 kg load directed toward the open end of the clamp without the roller being removed from the clamp body. The roller clamp shall be easily operated with one hand.

The complete assembly shall withstand 45 psig water pressure for 30 seconds with no more than one clinging drop leakage and no separation. The spike and needle adapter shall support a 2.3 kg tensile load for 1 minute without any separation from the tubing.

Sterility The transfer set shall comply with the sterility requirements shown below under Sterility test.

A. Sterility test. The transfer set shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it is clearly differentiated from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation medium for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30^0 to 35^0C for one half of the specimens and soybean-casein digest at 20^0 to 25^0C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Medium	Paper spore strip	Inoc. Prod. or Inoc. Simulated Prod.
Soybean-casein digest	S-3*	S-3*
Fluid thioglycollate or Alternative thioglycollate		S-3*

*Use a minimum of 20 specimens. Acceptance number is zero.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

B. Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism
Steam	B. stearothermophilus
Dry Heat	B. stearothermophilus
	B. subtilis
Ethylene oxide	B. subtilis
Gamma radiation	B. pumilus

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.

C. Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586A, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in A. Sterility test and B. Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in B. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs A and B.

D. Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

E. Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP. The inspection level shall be as stated in paragraph A above.

Workmanship. The transfer set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing forty-eight transfer sets, each set individually packaged, sterile, in a sealed peel open or tear open waterproof package, with detailed instructions for use, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:03Z
Check the source to verify that this is the currentversion before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified in the table below. The unit of product for test purposes shall be one transfer set.

Sampling criteria for tests.

Characteristic	Inspection level	AQL
Hardness of tubing	S-1	*
Class IV USP plastic	S-1	*
Security of assembly and leakage	S-3	*
Security of roller in clamp body	S-2	*
Non-pyrogenicity, USP	S-1	*
Safety, USP	S-1	*
Sterility	(See context above)	

* Acceptance number of zero.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-551-2331	TRANSFER SET, INTRAVENOUS SOLUTION, Non-Vented, Disposable, 27 Inches (68.6cm) Long, 48s. Sterile.

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4120

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.