A-A-51974A
5 May 1989
SUPERSEDING
A-A-0051974(DM)
29 May 1987

COMMERCIAL ITEM DESCRIPTION

TRANSFER SET, INTRAVENOUS SOLUTION
(Non-Vented, 27 Inches)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile, non-vented, disposable intravenous solution transfer set.

Salient characteristics:

Materials. The transfer set shall be sterile, non-vented, disposable and suitable for transfering medication from one bag to another or from one bottle to another when the bottles are equipped with airways.

All parts of the transfer set in contact with the intravenous solution being transferred shall comply with USP Class IV plastic requirements. Also the same surfaces shall be in conformance with USP requirements for non-pyrogenicity and non-toxicity. The clear tubing shall be made of polyvinyl chloride having a Shore Durometer A hardness of 71 + 5. The spike, roller clamp, needle adapter and protective caps shall be made of suitable rigid plastic.

Style and design. The transfer set shall consist of a spike with flange to facilitate handling, polyvinyl chloride tubing, roller clamp and needle adapter. A protective cap shall be present on the spike and on the needle adapter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

The overall length shall be 27 inches +/- 2.7 inches long. The spike shall be 3/4 to 13/16 inch long, have a single lumen and have a sharp, durable conical point. There shall be a slight taper to the first shoulder. The diameter just above the shoulder shall be 0.213 +/- 0.003 inch.

The tubing shall have an ID of 0.110 ± .005 inch and a wall thickness of 0.022 ± .003 inch. The male Luer needle adapter shall be of slip tip design and conform to ANSI/HIMA MD70.1. The roller clamp shall control flow from closed to fully open positions and at any intervening adjustment between these extremes. It shall be able to hold the adjustment constant.

Performance. The roller in the fully open position shall be able to withstand a 3.65 kg load directed toward the open end of the clamp without the roller being removed from the clamp body. The roller clamp shall be exactly operated with one hand.

The complete assembly shall withstand 45 psig water pressure for 30 seconds with no more than one clinging drop leakage and no separation. The spike and needle adapter shall support a 2.3 kg tensile load for 1 minute without any separation from the tubing.

The spike must be able to be immersed in isopropanol for 5 minutes, dried on paper toweling for an hour and then examined for crazing or cracking. There shall be no signs of crazing or cracking when examined at 2x magnification.

Sterility The transfer set shall be sterile in accordance with USP or AAMI requirements. Sterility assurance level shall be $1x10^{-6}$.

Heavy metals extrability. Plastic components coming in contact with intravenous solutions or patient's blood shall be tested. The quantity of lead or cadmium shall not exceed 5 parts per million (PPM) when tested in accordance with USP.

Workmanship. The transfer set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing forty-eight transfer sets, each set individually packaged, sterile, in a sealed peel open or tear open, or equivalent waterproof package, with detailed instructions for use, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one transfer set. Inspection level shall be S-1 and the acceptance number shall be zero.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580,as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-551-2331	TRANSFER SET, INTRAVENOUS SOLUTION, Non-Vented, Disposable, 27 Inches (68.6cm) Long, 48s. Sterile.

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4589

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51974A	TRANSFER SET, INTRAVENOUS SOLUTION (Non Vented, 27 Inches)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.