A-A-51978
20 May 1987

COMMERCIAL ITEM DESCRIPTION

CATHETER AND NEEDLE UNIT, INTRAVENOUS (20 Gage)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a catheter-over-needle unit suitable for administering intravenous medication from a single venipuncture.

Salient characteristics:

This catheter-over-needle unit consists of a 23 gage by 1-3/4-inch hypodermic needle, a flash chamber and plug.

The 20 gage by 1-1/4 inch catheter shall be made of radiopaque fluorinated ethylene propylene plastic and it and all parts of the fluid path shall comply with USP Class IV plastic requirements for biocompatibility. The catheter hub shall be made of polypropylene, nylon or polybutylterephalate. The hypodermic needle shall be made of Type 304, stainless steel in accordance with ASTM F899 and have a transparent flash chamber made of Tenite propionate or other clear suitable plastic.

The catheter hub shall conform to Luer requirements of ANSI/HIMA MD70.1 for inside dimensions and the locking lugs. The catheter at its leading edge shall be smooth and precisely fitted to the OD of the needle and tapered to minimize patient trauma. The exterior of the cannula of the needle shall be coated with a silicone lubricant which is inert, non-toxic and suitable for intravenous use.

The bevel of the needle shall be ground flat or concave and have a secondary grind near the tip to enhance cutting action and create a reinforcing ridge at the extreme tip of the bevel. All edges of the bevel shall be sharp except the heel which shall be dull to prevent coring. The bevel shall be 0.078 inch ± 10% long.

There shall be no burrs, chips or foreign matter present in or on the needle or catheter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

The catheters shall not exhibit leakage exceeding one clinging drop when tested in accordance with 5.2.1 Test for Water Leakage of ANSI/HIMA MD70.1.

All components shall comply with the sterility requirements shown below under Sterility test:

A. Sterility test. The catheter and needle unit shall be tested for sterility in accordance with the method specified in the USP using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it is clearly differentiated from product intended for distribution. The incubation medium for the paper spore strip (carrier) shall be soybean-casein digest. The incubation medium for the inoculated product, or inoculated simulated product shall be fluid thioglycollate or alternative thioglycollate at 30° to 35° C for one half of the specimens and soybean-casein digest at 20° to 25° C for the remaining half of the specimens. The minimum incubation time shall be ten days. Sampling shall be as follows:

Medium	Paper spore strip	Inoc. Prod. or Inoc. Simulated Prod.
Soybean-casein digest	S-3*	S-3*
Fluid thioglycollate or Alternative thioglycollate		S-3*

*Use a minimum of 20 specimens. Acceptance number is zero.

B. Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism
Steam	B. stearothermophilus
Dry Heat	B. stearothermophilus
	B. subtilis
Ethylene oxide	B. subtilis
Gamma radiation	B. pumilus

The number of viable spores shall be sufficient to provide assurance that the product has been sterilized.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

C. Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586A, Sterilization Test Strip Set, Bacterial Spore. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in A. Sterility test and B. Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in B. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP and paragraphs A and B.

D. Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

E. Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP. The inspection level shall be as stated in paragraph A above.

Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection level and acceptable quality level shall be as shown in Table I.

Table I. Sampling for examination.

Characteristic	Inspection level	AQL (% defective)
Dimensions & Gaging	S-2	*
Visual defects	S-4	1.0

* Acceptance number shall be zero for all sample sizes.

Sampling for tests shall be in accordance with MIL-STD-105. The inspection level and acceptable quality limits shall be as shown in Table II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Table II. Sampling for tests.

Characteristic	Inspection level	AQL (% defective)
USP Class IV	S-1	*
Leakage	S-2	*
Sterility	See sterility clauses above	*

* Acceptance number shall be zero for all sample sizes.

Instructions. Instructions for use shall be present with the container for the catheter and needle unit of issue.

Workmanship. The catheter and needle unit shall be free from defects which detract from its appearance or impair its serviceability.

Unit of issue. Package (PG). One package containing 50 individually packaged catheter and needle units, as specified constitutes one unit (of issue). The individual catheter and needle unit shall be contained in a rigid tray or container having a sealed, peel-open type lid. A sealed pouch of high density polyethylene (4 mil minimum) is also acceptable. The package for the individual catheter and needle unit shall be capable of maintaining sterility of contents unless it is damaged or opened. The package shall also provide mechanical protection against handling and transportation exigencies that may cause damage or kinking of the catheter and needle unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one catheter and needle unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order. Words such as "STORE IN A COOL DRY PLACE" shall be printed on the shipper.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-405-9243	CATHETER AND NEEDLE UNIT, INTRAVENOUS, 20 Gage by 1-1/4 Inches, FEP, Radiopaque, Disposable, 50s

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-CLS
PHS
FDA-MPQAS

Project No. 6515-3965

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
AA-51978	CATHETER, AND NEEDLE UNIT, INTRAVENOUS (16 GAGE)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
b. ADDRESS (Street, City, State, ZIP Code)	☐ USER
	☐ MANUFACTURER
	☐ OTHER (Specify) _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.