A-A-51983
22 May 1987

COMMERCIAL ITEM DESCRIPTION

SPECULUM, EAR (Boucheron)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers ear speculum of the Boucheron design.

Salient characteristics.

Sizes. Speculum shall be supplied in the following sizes as specified (see Figure 1 for tolerances):

Specification Size	Description
4	5.5 mm i.d.
5	6.5 mm i.d.
6	7.5 mm i.d.
7	9.0 mm i.d.
set	one each, size 4-7

Material. Speculum shall be manufactured from brass. Brass speculum shall be plated with chromium on the exterior surface. As an alternate, the speculum shall be manufactured from austenitic corrosion-resistant steel of composition in accordance with American Society for Testing and Materials (ASTM) Standard F899.

Style and Design. The speculum shall be of the style and dimensions shown in Figure 1. The ends of the speculum shall be well rounded and free from any sharp edges. The tapered portion of the speculum shall be smooth on both the inside and outside surfaces, with no raised seams visible.

Finish. Steel speculum shall have a matte or satin finish. The interior surfaces of brass speculums shall be smooth. The speculum shall be free from burrs, pits, scale, rough areas, and tool marks.

Plating. Plating (brass speculum) shall be uniform and free from blistering, chips, and cracks.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Corrosion resistance. Speculum shall be tested for corrosion-resistance as specified under Quality Assurance Provisions.

Markings. Speculum shall be permanently marked with the manufacturer's name or registered trademark and the size (specification size or nominal i.d.). Speculum of foreign manufacture shall be marked with the name of the country of manufacture. Speculum manufactured from corrosion-resistant steel shall be marked with the word "stainless."

Workmanship. The speculum shall be processed in such a manner as to be uniform and free from defects that will affect life, utility, or appearance.

Product Quality Level

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Set (SE) or Each (EA). One speculum set (containing sizes 4, 5, 6 and 7 for NSN 6515-01-164-8869) or one speculum of the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Speculum shall be examined to determine compliance with specification requirments.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as specified herein. The unit of product for test purposes shall be one speculum. Sampling shall be level S-1 with an AQL of 1.0.

Corrosion-test. Speculum shall be tested for corrosion resistance in accordance with the Resistance to Boil test of Federal Specification GG-I-526.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-164-8869	SPECULUM SET, EAR, Round, Boucheron, Set of four
6515-01-165-6702	SPECULUM, EAR EXAMINING, Boucheron, Round, Size 7
6515-01-165-6703	SPECULUM, EAR EXAMINING, Boucheron, Round, Size 4
6515-01-165-6704	SPECULUM, EAR EXAMINING, Boucheron. Round, Size 6
6515-01-167-3785	SPECULUM, EAR EXAMINING, Boucheron, Round, Size 5

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM


CIVIL AGENCY COORDINATING ACTIVITIES:


VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4008

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

FIGURE 1. BOUCHERON SPECULUM





NSN	SIZE	A	B	C
6515-01-164-8869	SET OF (4) NSN'S BELOW			
6515-01-165-6703	4	32	5.5	34
6515-01-167-3785	5	33	6.5	35
6515-01-165-6704	6	34	7.5	35
6515-01-165-6702	7	35	9.0	36

UNLESS OTHERWISE SPECIFIED
DIMENSIONS ARE IN MILLIMETERS
TOLERANCES ON

FRACTIONS	DECIMALS	ANGLES
±	± .5	±

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4956 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-LS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.