A-A-51987
26 June 1987

COMMERCIAL ITEM DESCRIPTION

TABLE, OPERATING, ORTHOPEDIC

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers an orthopedic, operating table with a formica top, pull-out instrument board, and adjustable legs, suitable for use with standard operating table in surgical procedures of the hand.

Salient Characteristics:

The table shall be Boyes-Parker style orthopedic operating table for hand surgery that shall be compatible with any standard operating table.

The table shall have a formica top with a frame-work made of stainless steel. The table shall have collapsible legs to provide for ease in moving and storing. The legs shall be vertically adjustable to allow fitting to any operating table height.

The formica table top shall have the following approximate dimensions:

Length:	46-1/4 inches (117.5 cm)
Width:	18 inches (45.7 cm)

The table shall have a pull-out instrument board to allow easy access to and positioning of instruments during surgery.

The underside of the top of the table shall have two tracks for the placement of a stainless steel pan to catch fluids from handwashings. The pan shall have the following approximate dimensions:

Length:	16-5/8 inches (42.2 cm)
Width:	9-1/2 inches (24.1 cm)
Depth:	4 inches (10.2 cm)

AMSC/NA FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

The stainless steel used for the framework and the drain pan shall be American Iron and Steel Institute (AISI) 300 series.

The formica table top finish shall be uniform and free of scratches, discoloration, and burrs.

All welds shall be complete with no breaks or cracks. The welds shall be ground smooth to blend with adjacent surfaces, and clean with no pitting, spatter, or discoloration.

The table corners shall be rounded and all edges shall be smooth.

Workmanship. The orthopedic operating table shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One orthopedic operating table, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T09:03Z
Check the source to verify that this is the currentversion before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one orthopedic operating table as specified in this description.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional examination	S-1	2.5

Examination. The operating table shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

TABLE II. Categories and Defects.

Categories	Defects
Major	
101	Any component part of table missing or unusable.
102	Table top not formica.
103	Table frame and drain pan not stainless steel.
104	Legs not vertically adjustable.
105	Welds have breaks or cracks.
Minor	
201	Table legs not collapsible.
202	Drain pan not easily installed.
203	Table top finish not uniform and free of scratches, discoloration, and burrs.
204	Welds not ground smooth and clean with no pitting, spatter or discoloration.
205	Table corners not rounded and edges not smooth.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurdisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-166-5096

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:03Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51296

Project No. 6530-1890

Location: CID/TABLE/H13-24

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.