A-A-51988
26 June 1987

COMMERCIAL ITEM DESCRIPTION

STOOL, REVOLVING

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification AA-S-696.

This commercial item description covers a surgeon's adjustable revolving stool.

Salient characteristics:

The surgeon's stool shall be corrosion-resisting steel with four legs and a revolving, adjustable seat.

All parts of the stool except the seat screw and adapter flange shall be made of corrosion-resisting steel, American Iron and Steel Institute (AISI) 300 series. As an alternate, the screw socket may be carbon steel with a corrosion-resisting steel housing.

The seat screw and adapter flange shall be AISI 1010 or 1215 carbon steel.

The seat height shall be adjustable within a minimum range of 19 to 25 inches. It shall not be possible to inadvertently remove the seat.

The seat shall have a diameter between 12-1/2 and 15-1/2 inches with a slightly concave upper surface and a well-rounded apron. Seat thickness shall be 0.078 inch minimum.

Top of seat shall be parallel to the floor within 2 degrees when measured throughout one full revolution of the seat.

The four legs shall be constructed of tubing with a minimum wall thickness of 0.049 inch. The end of each leg shall be closed off flush with a metal cap or plug. A securely fitting electrically conductive rubber tip shall be on the bottom of each leg.

The stool shall have a leg brace parallel to the floor approximately 7 inches above the lower tips of the legs and welded thereto. The brace metal shall be 3/8 inch minimum diameter rod or 3/4" minimum O.D. tubing.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

The threaded portion of the screw and screw socket shall have acme machined threads which shall permit easy height adjustment of the seat without noise, chatter, or play.

Welding shall be durable, sound in every detail, and in accordance with the latest practices recommended by the American Welding Society. All welds shall be continuous with no visible voids, pin holes, or cracks.

All exposed surfaces of the stool shall have a No. 4 commercial finish or better.

Stool shall be free of rocking when the four legs are placed on a smooth flat surface.

Seat shall maintain height either free standing or under normal load.

Stool shall be free of sharp edges.

The surgeon's stool shall meet the standards of NFPA 99.

Workmanship. The revolving stool shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One revolving stool, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one complete revolving stool as specified in this description.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	S-4	1.0
Total defects (Major and minor combined)	S-4	4.0
For dimensional examination	S-1	2.5

Examination. The revolving stool shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

TABLE II. Categories and Defects.

Categories	Defects
Major	
101	Stool not free of sharp edges.
102	Stool not free of rocking when the four legs are placed on a smooth flat surface.
103	Seat does not maintain height either free standing or under normal load.
104	Vertical position of the seat cannot be changed; seat cannot be revolved without binding.
105	Steel caps or plugs not present on floor end of legs.
106	Steel caps or plugs not permanently and firmly secured.
107	Rubber crutch tips not present.
108	Rubber crutch tips do not fit snugly.
109	Type, number, and length of welds not as specified.
110	Welds not free of cracks.
Minor	
201	Stool not free of grind marks, trim marks, bulges, and scratches.
202	Upper surface of seat not concave.
203	Unexposed welds not properly cleaned
204	Undersurface of seat not painted.
205	Painted surface not free of blisters.
206	Rubber crutch tips not free of tears.
207	Unpainted welds not ground and polished smooth.

Dimensional Examination. The stools shall be examined for defects in dimensions. Any dimension not within the specifications herein shall be classified as a defect.

Tests. Tests shall be conducted, as necessary, to determine compliance with specification requirements.

Electrical conductivity. The stool with rubber tips in place shall be tested in accordance with NFPA 56A.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 2026-04-26T09:03Z
Check the source to verify that this is the current version before use.

f a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-00-708-8000

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:03Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6530-1838

PGC 50296

Location: CID/STOOL/H13-3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides [illegible] submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTC
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51988	Stool, Revolving

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*. _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.