METRIC
A-A-50223A
17 July 2001
SUPERSEDING
A-A-50223
23 September 1986

COMMERCIAL ITEM DESCRIPTION

SCALE, BEAM INDICATING: WEIGHING, SPECIFIC GRAVITY; CAPACITY 0.01 NOT LESS THAN 200 GRAMS, THREE OR FOUR WEIGHT BEAMS: 0.01 GRAM SENSITIVITY

The General Services Administration has authorized the use of this commercial item description for all federal agencies.

1. SCOPE. This commercial item description (CID) establishes the government acquisition requirements for specific gravity weighing beam indicating scales to be used as accurate weighing instruments.

2. CLASSIFICATION. The scales shall be classified by the types listed below. The selected type to be supplied shall be specified in the acquisition order (see 7.5(b)).

Type I - Three weight beams
Type II - Four weight beams

3. SALIENT CHARACTERISTICS. The scale shall be of a three or four beam type with an adjustable platform and have a capacity of 0.01 gram to not less than 200 grams with a sensitivity of 0.01 gram at the capacity load.

3.1 Metals. Metals shall be corrosion-resistant or metal made corrosion-resistant by a suitable treatment.

Beneficial comments, recommendations, additions, deletions, clarifications, etc. and any data that may improve this document should be sent to: Defense Supply Center Richmond (DSCR), ATTN: DSCR-VBD, 8000 Jefferson Davis Highway, Richmond, VA 23297-5610.

AMSC N/A FSC 6670

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:41Z
Check the source to verify that this is the current version before use.

3.2 Components. The scales shall consist of the following components:

3.2.1 Weight beams. The weight beams shall be equipped as follows:

a. Two or three beams with hanging sliding poises or riders.
b. A fractional bar having a capacity of 1 gram, graduated in 0.01 gram increments, and provided with a sliding poise or rider.
c. An adjustment for setting zero point.
d. An adjustment means for leveling the scale when out of level condition affects accuracy.

3.2.2 Stirrup. The stirrup shall be corrosion-resistant steel, approximately 203.2 mm (8 inches) high and approximately 101.6 mm (4 inches) wide.

3.2.3 Pan. The pan shall be made of aluminum, polypropylene, or corrosion-resistant steel. It shall be removable from the stirrup or be movable to a vertical position within the stirrup.

3.2.4 Platform. The platform for specific gravity determinations shall be adjustable.

3.2.5 Damping. Damping shall be incorporated into the scale and shall be achieved by one of the following methods: magnetic, air, or oil.

3.2.6 Accuracy. Accuracy shall not be affected by the scale being out of level.

3.2.7 Cover. A plastic cover designed to fit the scale shall be provided.

4. REGULATORY REQUIREMENTS

4.1 Recovered materials. The offeror/contractor is encouraged to use recovered materials to the maximum extent practicable, in accordance with paragraph 23.403 of the Federal Acquisition Regulation (FAR).

5. PRODUCT CONFORMANCE PROVISIONS

5.1 Product conformance. The products provided shall meet the salient characteristics of this commercial item description, conform to the producer's own drawings, specifications, standards, and quality assurance practices, and be the same product offered for sale in the commercial marketplace. The government reserves the right to require proof of such conformance.

5.2 Market acceptability. The product offered must have been previously sold either to the government or on the commercial market.

6. PACKAGING

6.1 Preservation, packing, and marking. For acquisition purposes, the products shall be preserved, packed, and marked as specified in the acquisition order (see7.5(c)).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:41Z
Check the source to verify that this is the current version before use.

6.2 Overseas packaging. Movable components of the balance shall be secured in a manner that will prevent movement or damage during shipment and storage. The balances shall be packed in a fiberboard box conforming to ASTM D 5118/D 5118M, "Standard Practice for Fabrication of Fiberboard Shipping Boxes" (DoD adopted). The box shall be closed, waterproofed, and reinforced in accordance with ASTM D 1974, "Methods of Closing, Sealing, and Reinforcing Fiberboard Boxes" (DoD adopted).

6.3 Overseas marking. Marking shall be in accordance with MIL-STD-129, "Standard Practice for Military Marking", or FED-STD-123, "Marking for Shipment (Civil Agencies)", as applicable.

6.4 Domestic packaging, packing, palletization, and marking. Packaging, packing, palletization and marking shall be in accordance with ASTM D 3951, "Standard Practice for Commercial Packaging" (DoD adopted).

7. NOTES

7.1 Part or identification number (PIN). The following part or identification numbering procedure is for government purposes and does not constitute a requirement for the contractor.



7.2 Sources of documents.

7.2.1 Government documents. Copies of FAR, military specifications and federal standards may be obtained from the U.S. Government Printing Office, Superintendent of Documents, Mail Stop: SSOP, Washington, DC 20402-9328. Electronic copies of military and federal standards may be obtained from http://astimage.daps.dla.mil/quicksearch/.

7.2.2 ASTM standards. Copies of ASTM standards may be obtained from the American Society for Testing and Materials, 100 Barr Harbor Drive, West Conshohocken, PA 19428-2959.

7.3 National stock number (NSN). The NSN listed below is assigned for the listed product types (see table I). Other NSNs may also correspond with this document.

NSN	Type
6670-00-494-3845	II

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:41Z
Check the source to verify that this is the current version before use.

7.4 Sources of supply. The manufacturer and/or suppliers listed below are known to supply products that meet the salient characteristics requirements of this document. Competition is not limited to the listed firms.

Ohaus Corporation
Florham Park, NJ 07932

A. Daigger and Company
Vernon Hills, Il 60061

Alliance Scale Incorporated
Canton, MA 02021-0509

Apex Scales
Kent, WA 98042

Phillips Morris Scale Company
Seattle, WA 98119

7.5 Ordering data. Acquisition documents must specify the following information:

a. CID document number, revision, and CID PIN
b. Product classification (type) (see 2)
c. Preservation, packing, and marking requirements (see 6.1)

7.6 Subject term (key word) listing.

damping
pan
platform
poise
stirrup

MILITARY INTERESTS:

Custodians:
Army - CE

Reviewers:
Army - MD

CIVIL AGENCY
COORDINATING ACTIVITY:

GSA - 7FXE

Preparing activity:
DLA - GS1

(Project 6670-0213)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:41Z
Check the source to verify that this is the current version before use.