INCH-POUND
A-A-50226B
30 June 2003
A-A-50226A
31 October 1998

COMMERCIAL ITEM DESCRIPTION
SHAKING MACHINE, TESTING SIEVE, LABORATORY, PORTABLE, HAND-OPERATED, CAPACITY 6 FULL-HEIGHT OR 13 HALF-HEIGHT, 8-INCH STANDARD SIEVES

The General Services Administration has authorized the use of this Commercial Item Description (CID) for all federal agencies. The original issue of this CID replaced MIL-S-11568.

1. SCOPE.

1.1 Abstract. This CID covers a hand-operated, portable shaking machine intended for use in determining grain-size distribution of soil.

2. SALIENT CHARACTERISTICS.

2.1 Description. The shaking machine shall consist of a sieve carriage and support (without sieves), oscillating mechanism, and a base.

2.2 Materials. The shaking machine shall be fabricated from compatible materials, inherently corrosion resistant, or treated to provide protection against corrosion and deterioration to which the shaking machine may be exposed.

2.2.1 Treatment and painting. Unless otherwise specified (see 6.2), the shaking machine shall be treated and painted in accordance with the manufacturer's normal method for retail distribution.

2.3 Physical characteristics.

2.3.1 Sieve carriage and support. The sieve carriage and support shall be removable for disassembly, cleaning, and maintenance. The carriage shall hold one through six (minimum) full-height, or from one through 13 (minimum) half-height, 8-inch sieves conforming to ASTM E 11.

Beneficial comments, recommendations, additions, deletions, clarifications etc., and any other data which may improve this document should be sent to: ENGINEER RESEARCH AND DEVELOPMENT CENTER, CORPS OF ENGINEERS, U.S. ARMY TOPOGRAPHIC ENGINEERING CENTER, ATTN CEERD-TS, 7701 TELEGRAPH ROAD, ALEXANDRIA, VA 22315-3864.

AMSC N/A FSC 6635

Distribution Statement A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:41Z
Check the source to verify that this is the current version before use.

2.3.2 Oscillating mechanism. The oscillating mechanism shall move the sieve platform in an eccentric orbit and receive a minimum of two through a maximum of eight impacts during each orbit. The oscillating mechanism shall operate through the means of a hand-cranking device, or similar mechanism. The hand operation shall not exceed 10 pounds of applied force.

2.3.3 Base. The base shall be of the same material as the carriage and provide rigid support for the oscillating mechanism and the sieve support.

2.3.4 Lubrication. All surfaces requiring lubrication shall be provided with a means of performing maintenance. Access for routine maintenance shall be provided in a manner maximizing user safety and ease of repair.

2.3.5 Dimensions and weight. The shaking machine shall not exceed 30.00 inches in length, width, or height and 42 pounds in weight.

2.4 Identification marking. Each shaking machine shall have a durable data plate identifying, as a minimum, the manufacturer's name and part number. The data plate shall be located in a place where the data will be visible and legible, but not interfere with the safe operation and maintenance of the shaking machine.

3. REGULATORY REQUIREMENTS

3.1 Recycled, recovered, or environmentally preferable materials. Recycled, recovered, or environmentally preferable materials shall be used to the maximum extent possible provided the material meets or exceeds the operational and maintenance requirements, and promotes economically advantageous life cycle costs. Used, rebuilt, or remanufactured components, parts, and pieces shall not be incorporated into the shaking machine.

4. QUALITY ASSURANCE PROVISIONS.

4.1 Product conformance. The shaking machine shall meet the salient characteristics of this CID, conform to the manufacturer's own drawings, specifications, standards, and quality assurance practices, and be the same product offered for sale in the commercial marketplace. The government reserves the right to require proof of such conformance.

4.2 Market acceptability. The shaking machine offered shall be the latest model of the standard product of the supplier and shall have been used for at least one year by commercial organizations.

4.3 Examinations. Each shaking machine shall be examined for compliance with the requirements specified in 2.2 through 2.4. Any modification necessary following failure to meet the specified requirements shall receive particular attention for adequacy and suitability. This element of inspection shall encompass all examinations of the performance and dimensional requirements. Non-compliance to any specified requirement shall constitute cause for rejection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:41Z
Check the source to verify that this is the current version before use.

5. PACKAGING.

5.1 Preservation, packing, and marking Preservation, packing, and marking shall be as specified in the contract, or order.

6. NOTES.

6.1 Intended use. The shaking machine covered by this specification is intended to determine grain-size distribution of soil.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number, and date of this CID.
b. If special treatment or paint is required (see 2.2.1).
c. Packing requirements (see 5.1).

6.3 Data requirements. The contracting officer should include in the contract or order, any data requirements, such as maintenance and operation manuals, instructional materials, and technical publications to be furnished with each shaking machine.

6.4 Provisioning. The contracting officer should include in the contract or order, provisioning requirements for repair parts, maintenance tools (including special tools), as necessary; and instructions regarding shipment of the shaking machine.

6.5 Source of documents.

6.5.1 Copies of ASTM specifications are available from the American Society for Testing and Materials, 100 Barr Harbor Drive, West Conshohocken, PA 19428-2959.

6.6 Subject term (key words) listing.

Soil distribution
Grain distribution

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:41Z
Check the source to verify that this is the current version before use.

Custodian:
Army - CE3
DLA - GS

Preparing Activity:
Army - CE3
(Project 6635-0203)

Review Activity:
Army - GL

Civil Coordinating Activity:
GSA - FCGC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:41Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter must be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-50226B	2. DOCUMENT DATE (YYYYMMDD) 20030630

3. DOCUMENT TITLE: Shaking Machines, Testing Sieve, Laboratory, Portable, Hand-operated, Capacity 6 Full-Height or 13 Half-Height, 8-Inch Standard Sieves

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) DSN (if applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME TOPOGRAPHIC ENGINEERING CENTER	b. TELEPHONE (Including Area Code) (1) Commercial (703) 428-6862 (2) DSN 328-6862
c. ADDRESS (Include Zip Code) Topographic Engineering Center ATTN: CEERD-TS 7701 Telegraph Road Alexandria, VA 22315-3864	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 DSN 761-9340

DD Form 1426, Oct 89 *Previous editions obsolete*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:41Z
Check the source to verify that this is the current version before use.